



Aalborg Universitet

AALBORG UNIVERSITY
DENMARK

Compensatory Movements Involved During Simulated Upper Limb Prosthetic Usage:
A Musculoskeletal Model-based Evaluation and Validation Study

Nagaraja, Vikranth; Bergmann, Jeroen; Andersen, Michael Skipper; Thompson, Mark

Publication date:
2018

Document Version
Accepted author manuscript, peer reviewed version

[Link to publication from Aalborg University](#)

Citation for published version (APA):

Nagaraja, V., Bergmann, J., Andersen, M. S., & Thompson, M. (2018). *Compensatory Movements Involved During Simulated Upper Limb Prosthetic Usage: A Musculoskeletal Model-based Evaluation and Validation Study*. Abstract from 8th World Congress of Biomechanics, Dublin, Ireland.
<https://app.oxfordabstracts.com/stages/123/programme-builder/submission/21310?backHref=/events/123/programme-builder/view/sort/author&view=published>

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal -

Take down policy

If you believe that this document breaches copyright please contact us at vbn@aub.aau.dk providing details, and we will remove access to the work immediately and investigate your claim.

Compensatory Movements Involved During Simulated Upper Limb Prosthetic Usage: A Musculoskeletal Model-based Evaluation and Validation Study

Vikranth H. Nagaraja^{1*}, Jeroen H.M. Bergmann¹, Michael S. Andersen², and Mark S. Thompson¹

¹*Institute of Biomedical Engineering, University of Oxford, Oxford - OX3 7LD, UK;* ²*Department of Mechanical and Manufacturing Engineering, Aalborg University, Fibigerstræde 16, DK-9220 Aalborg Ø*

*E-mail: vikranth.harthikotenagaraja@eng.ox.ac.uk

Introduction

Prosthetic arms have generally seen poor outcomes as the perceived functional gain is limited by the disadvantages [1]-[2]. Most upper limb (UL) prostheses, especially affordable prostheses, lack controllable distal joint(s) which necessitates compensatory movements at the proximal joint(s) during functional use [3], and these movements are linked to prosthesis rejection [4]. Studies characterising such movements are mostly limited to kinematic analyses [3],[5]-[7] and few studies have adopted a multi-body musculoskeletal (MS) model-based approach in this field. MS models facilitate estimation of *in vivo* loading and enable 'what-if' analyses; although, it should be noted that MS model Verification & Validation is a non-trivial endeavour and remains a vital topic of ongoing research [8]-[10]. Hence, the objective of this study is to perform (i) MS model validation and (ii) MS model-based evaluation of the compensatory movements adopted during simulated prosthesis usage.

Methods

This study was approved by the local Research Ethics Committee (Ref no. 16/SC/0051). Eleven adult able-bodied participants consented to participate in this study. Three trials of a *Reach-to-grasp* task in a 3D marker-based optical motion capture laboratory setting were performed in a seated position using a custom-built apparatus. Surface EMG (sEMG) data pertinent to maximum voluntary contractions of select muscle groups were recorded for model validation purposes. Movement data and sEMG data were collected during task execution. Prosthesis usage in the participants was simulated using a commercially-available wrist brace that mimics lack of a controllable wrist in a typical prosthetic arm. The *Standard MoCap* model in the AnyBody Managed Model Repository v.1.6.5 distributed with the AnyBody Modelling System™ was used to estimate *in vivo* loading. Study validation was performed by indirect, qualitative, and quantitative comparison [11].

Results and Discussion

Study validation reveals that the magnitudes of shoulder joint loading were generally in agreement with those reported in relevant published studies [12]-[13]. For qualitative comparison, the shapes of calculated and recorded muscle activity curves matched well. In general, joint loading values were lower for the braced condition. Trends such as smaller joint reaction forces at the shoulder and higher joint moments for the trunk and elbow were seen for the braced condition.

Conclusion

The first-ever study involving MS model-based characterisation of compensatory movements adopted during simulated prosthesis use and model validation has been carried out. This study is expected to improve the objectivity in measuring prosthetic function and personalising device.

References

1. Østlie et al., 2012.
2. Biddiss et al., 2007.
3. Metzger et al., 2012.
4. Silcox, et al., 1993.
5. Hussaini et al., 2016.
6. Carey, 2008.
7. Major et al., 2014.
8. Hicks et al., 2015.
9. Erdemir et al., 2007
10. Lund et al., 2012.
11. Nigg & Herzog, 1999.
12. Bergmann et al., 2007.
13. Westerhoff et al., 2009.