



AALBORG UNIVERSITY
DENMARK

Aalborg Universitet

Long-Term Stability and Calibration of the Reference Thermometer ASL F200

Johra, Hicham

DOI (link to publication from Publisher):
[10.54337/aau328894425](https://doi.org/10.54337/aau328894425)

Publication date:
2019

Document Version
Publisher's PDF, also known as Version of record

[Link to publication from Aalborg University](#)

Citation for published version (APA):
Johra, H. (2019). *Long-Term Stability and Calibration of the Reference Thermometer ASL F200*. Department of Civil Engineering, Aalborg University. DCE Technical Reports No. 266 <https://doi.org/10.54337/aau328894425>

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal -

Take down policy

If you believe that this document breaches copyright please contact us at vbn@aub.aau.dk providing details, and we will remove access to the work immediately and investigate your claim.



DEPARTMENT OF CIVIL ENGINEERING
AALBORG UNIVERSITY

Long-Term Stability and Calibration of the Reference Thermometer ASL F200

Hicham Johra



Aalborg University
Department of Civil Engineering
Architectural Engineering

DCE Technical Report No. 266

**Long-Term Stability and Calibration of the
Reference Thermometer ASL F200**

by

Hicham Johra

December 2019

© Aalborg University

Scientific Publications at the Department of Civil Engineering

Technical Reports are published for timely dissemination of research results and scientific work carried out at the Department of Civil Engineering (DCE) at Aalborg University. This medium allows publication of more detailed explanations and results than typically allowed in scientific journals.

Technical Memoranda are produced to enable the preliminary dissemination of scientific work by the personnel of the DCE where such release is deemed to be appropriate. Documents of this kind may be incomplete or temporary versions of papers—or part of continuing work. This should be kept in mind when references are given to publications of this kind.

Contract Reports are produced to report scientific work carried out under contract. Publications of this kind contain confidential matter and are reserved for the sponsors and the DCE. Therefore, Contract Reports are generally not available for public circulation.

Lecture Notes contain material produced by the lecturers at the DCE for educational purposes. This may be scientific notes, lecture books, example problems or manuals for laboratory work, or computer programs developed at the DCE.

Theses are monographs or collections of papers published to report the scientific work carried out at the DCE to obtain a degree as either PhD or Doctor of Technology. The thesis is publicly available after the defence of the degree.

Latest News is published to enable rapid communication of information about scientific work carried out at the DCE. This includes the status of research projects, developments in the laboratories, information about collaborative work and recent research results.

Published 2019 by
Aalborg University
Department of Civil Engineering
Thomas Manns Vej 23
DK-9220 Aalborg Ø, Denmark

Printed in Aalborg at Aalborg University

ISSN 1901-726X
DCE Technical Report No. 266

Contents

1. Foreword	6
2. ASL F200 Thermometer Specifications.....	7
3. Long-Term Stability.....	8
4. Current Calibration Correction	11
5. Conclusion	12
6. References	13
7. Appendix.....	14

1. Foreword

The aim of this technical report is to provide detailed information about the long-term stability of the thermometer ASL F200 (WIKA Instruments Limited [1]) that is used as temperature reference to calibrate other temperature sensors in the Laboratory of Building Energy and Indoor Environment at Aalborg University – Department of the Built Environment [2]. This ASL F200 thermometer is regularly sent for recalibration at the “Temperature Laboratory” of the Danish Technological Institute, which is a National Reference Laboratory [3].

In this report, the stability of the thermometer is assessed as the difference in the temperature reading of the instrument at a specific temperature over time. The latter is calculated as the yearly deviation (or stability) in between consecutive recalibrations, which is equivalent to the difference in the calibration correction term in between two consecutive recalibrations divided by the elapsed time in between these two consecutive recalibrations.

The long-term stability of the ASL F200 thermometer is only assessed here for the first channel “Chan 1” of the instrument.

All calculations of this technical report are based on calibrations reports from the National Reference Laboratory of the Danish Technological Institute [3]. The main results of those calibration reports can be found in the Appendix at the end of this document.

2. ASL F200 Thermometer Specifications

The precision thermometer ASL F200 has 2 measurement channels equipped with 2 RTD temperature probes Pt100. This report is only concerning the calibration and stability of the first channel denominated "Chan 1". In addition, the thermometer is recalibrated and therefore used as a reference, in the temperature range from -20 °C to 100 °C. According to the manufacturer's specifications, the yearly deviation (stability) of the ASL F200 precision thermometer should be better (lower) than 0.005 °C per year.

3. Long-Term Stability

One can see in *Figure 1* the different calibration correction curves of the ASL F200 reference thermometer that have been collected over the last years. One can observe that the calibration correction curves do not change much from one calibration to the other.




Figure 1: The different calibration correction curves of the ASL F200 reference thermometer.

One can see in *Figure 2* the yearly deviation of the ASL F200 reference thermometer for different temperatures at the different recalibration times. One can observe that, apart from the negative temperatures in 2009 and 2012, and the 100 °C temperature in 2009, the yearly deviation of the ASL F200 thermometer is fairly stable and within the conformity limit of the manufacturer’s documentation (± 0.005 °C per year.), for the whole range of temperature measurement.




Figure 2: Yearly deviation of the ASL F200 reference thermometer at the different recalibration times.

One can see in *Figure 3* the evolution of the thermometer's yearly deviation (average of the absolute deviation and maximum of the absolute deviation over the full range of temperature measurement) as a function of time. One can observe that the average of the absolute deviation of the thermometer's measurement is always below the manufacturer's conformity limit. However, for the years 2009 and 2012, the maximum deviation over the full range of measurement is larger than the conformity limit. As mentioned before, this is due to some extreme deviations restricted to the far ends of the range of measurement below 0 °C and at 100 °C. Fortunately, since 2018, the yearly deviation of the ASL F200 thermometer over the entire range of temperature measurement is always better than the conformity limit.




Figure 3: Evolution of the yearly deviation (average of the absolute deviation and maximum of the absolute deviation) as a function of time.

4. Current Calibration Correction

In order to perform an accurate temperature measurement with the ASL F200 reference thermometer, it is necessary to correct the readings of the instrument with the adequate correction term. This correction term depends on the temperature reading of the thermometer. The temperature-dependent correction terms (calculated from the recalibration report) are shown in *Figure 4* and can be found in Appendix *Table 5*. The correction term corresponding to the instrument's reading must be added to the instrument's reading:

$$\text{Corrected measurement} = \text{Instrument reading} + \text{Correction}$$

Because the recalibration report comprises only correction terms at every 10 °C, it is recommended to use a linear interpolation or a cubic Hermite spline interpolation function to calculate the correction term at the specific temperature corresponding to the reading of the instrument. Alternatively, for very smooth and regular calibration correction curves such as the one from the 2020 recalibration (see *Figure 4*), one can use a polynomial fitting function on the entire set of recalibration data to interpolate the correction term at the specific temperature.




Figure 4: Current calibration correction curve for the ASL F200 reference thermometer to be used from 2020 onwards.

5. Conclusion

From the evolution of the thermometer's yearly deviation presented in *Figure 3*, one can conclude that the ASL F200 thermometer presents a very good long-term stability over the last 2 years and can thus be trusted as a reference thermometer for the calibration of other temperature sensors.

6. References

- [1] WIKA Instruments Limited
<https://www.wika.co.uk/>
- [2] Aalborg University, Department of Built Environment, Aalborg, Denmark.
<https://www.build.aau.dk/>
- [3] Danish Technological Institute (DTI), The Temperature Laboratory.
<https://www.dti.dk/testing/temperature/37566>

7. Appendix

Table 1: Results of the calibration report of 2005.

Reference temperature [°C]	Instrument reading [°C]	Error [°C]	Correction [°C]
-20.0042	-20.0050	-0.0008	0.0008
-10.0089	-10.0075	0.0014	-0.0014
0.0000	-0.0100	-0.0100	0.0100
10.0100	10.0060	-0.0040	0.0040
20.0060	20.0080	0.0020	-0.0020
30.0070	30.0130	0.0060	-0.0060
40.0057	40.0155	0.0098	-0.0098
50.0078	50.0205	0.0127	-0.0127
60.0008	60.0145	0.0137	-0.0137
70.0037	70.0180	0.0143	-0.0143
80.0016	80.0160	0.0144	-0.0144
90.0127	90.0100	-0.0027	0.0027
100.0048	100.0060	0.0012	-0.0012

Table 2: Results of the calibration report of 2009.

Reference temperature [°C]	Instrument reading [°C]	Error [°C]	Correction [°C]
-20.1026	-20.0795	0.0231	-0.0231
-10.0195	-10.0080	0.0115	-0.0115
0.0000	-0.0040	-0.0040	0.0040
9.9906	9.9905	-0.0001	0.0001
20.0370	20.0430	0.0060	-0.0060
29.9930	30.0020	0.0090	-0.0090
39.9760	39.9870	0.0110	-0.0110
50.0159	50.0280	0.0121	-0.0121
60.0720	60.0835	0.0115	-0.0115
69.9874	70.0025	0.0151	-0.0151
80.0050	80.0170	0.0120	-0.0120
89.9920	90.0100	0.0180	-0.0180
99.9900	100.0190	0.0290	-0.0290

Table 3: Results of the calibration report of 2012.

Reference temperature [°C]	Instrument reading [°C]	Error [°C]	Correction [°C]
-20.0005	-20.0295	-0.0290	0.0290
-10.0583	-10.0760	-0.0177	0.0177
0.0000	-0.0080	-0.0080	0.0080
9.9981	9.9970	-0.0011	0.0011
19.9940	20.0010	0.0070	-0.0070
30.0330	30.0450	0.0120	-0.0120
39.9830	40.0000	0.0170	-0.0170
49.9774	49.9970	0.0196	-0.0196
59.9770	59.9985	0.0215	-0.0215
69.9764	69.9985	0.0221	-0.0221
79.9927	80.0155	0.0228	-0.0228
90.0480	90.0740	0.0260	-0.0260
100.0620	100.0860	0.0240	-0.0240

Table 4: Results of the calibration report of 2018.

Reference temperature [°C]	Instrument reading [°C]	Error [°C]	Correction [°C]
-20.6814	-20.6995	-0.0181	0.0181
-10.6286	-10.6350	-0.0064	0.0064
0.0000	0.0030	0.0030	-0.0030
10.3294	10.3365	0.0071	-0.0071
20.2992	20.3095	0.0103	-0.0103
30.2714	30.2845	0.0131	-0.0131
40.0836	40.1010	0.0174	-0.0174
50.0565	50.0780	0.0215	-0.0215
60.0318	60.0535	0.0217	-0.0217
70.0103	70.0325	0.0222	-0.0222
79.9909	80.0105	0.0196	-0.0196
90.0043	90.0245	0.0202	-0.0202
100.0124	100.0290	0.0166	-0.0166

Table 5: Results of the calibration report of 2020.

Reference temperature [°C]	Instrument reading [°C]	Error [°C]	Correction [°C]
-20.0117	-20.0310	-0.0193	0.0193
-10.0040	-10.0150	-0.0110	0.0110
0.0000	-0.0050	-0.0050	0.0050
10.0098	10.0130	0.0032	-0.0032
19.9999	20.0085	0.0086	-0.0086
29.9913	30.0045	0.0132	-0.0132
39.9901	40.0060	0.0159	-0.0159
50.0033	50.0220	0.0187	-0.0187
59.9993	60.0190	0.0197	-0.0197
69.9973	70.0170	0.0197	-0.0197
79.9885	80.0080	0.0195	-0.0195
89.9954	90.0120	0.0166	-0.0166
100.0025	100.0165	0.0140	-0.0140

Recent publications in the DCE Technical Report Series

Hicham Johra. Thermal properties of common building materials. DCE Technical Reports No. 216. Department of Civil Engineering, Aalborg University, 2019.

Hicham Johra. Project CleanTechBlock 2: Thermal conductivity measurement of cellular glass samples. DCE Technical Reports No. 263. Department of Civil Engineering, Aalborg University, 2019.

Hicham Johra. Cleaning Procedure for the Guarded Hot Plate Apparatus EP500. DCE Technical Reports No. 265. Department of Civil Engineering, Aalborg University, 2019.

