



Aalborg Universitet

AALBORG UNIVERSITY
DENMARK

A Chlamydia trachomatis 23S rRNA G1523A variant escaping detection in the Aptima Combo 2 assay (Hologic) was widespread across Denmark in July-September 2019

Hadad, Ronza; Skov Jensen, Jørgen; Westh, Henrik; Grønbaek, Ida; Jessen Schwartz, Lasse; Nielsen, Lene; Müller Vang, Tobias; Nielsen, Rikke; Sandborg Weinreich, Lenette; Skov, Marianne N; Olsen, Marlene; Kjølseth Møller, Jens; Kolmos, Birte; Unemo, Magnus; Hoffmann, Steen

Published in:

APMIS - Journal of Pathology, Microbiology and Immunology

DOI (link to publication from Publisher):

[10.1111/apm.13043](https://doi.org/10.1111/apm.13043)

Publication date:

2020

Document Version

Accepted author manuscript, peer reviewed version

[Link to publication from Aalborg University](#)

Citation for published version (APA):

Hadad, R., Skov Jensen, J., Westh, H., Grønbaek, I., Jessen Schwartz, L., Nielsen, L., Müller Vang, T., Nielsen, R., Sandborg Weinreich, L., Skov, M. N., Olsen, M., Kjølseth Møller, J., Kolmos, B., Unemo, M., & Hoffmann, S. (2020). A Chlamydia trachomatis 23S rRNA G1523A variant escaping detection in the Aptima Combo 2 assay (Hologic) was widespread across Denmark in July-September 2019. *APMIS - Journal of Pathology, Microbiology and Immunology*, 128(6), 440-444. Advance online publication. <https://doi.org/10.1111/apm.13043>

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal -

Corresponding Author Email ID: magnus.unemo@orebroll.se

A *Chlamydia trachomatis* 23S rRNA G1523A variant escaping detection in the Aptima Combo 2 assay (Hologic) was widespread across Denmark in July-September 2019

RONZA HADAD¹, JØRGEN SKOV JENSEN², HENRIK WESTH³, IDA GRØNBÆK³, LASSE JESSEN SCHWARTZ², LENE NIELSEN⁴, TOBIAS MÜLLER VANG⁴, RIKKE NIELSEN⁵, LENETTE SANDBORG WEINREICH⁵, MARIANNE N. SKOV⁶, MARLENE OLSEN⁶, JENS KJØLSETH MØLLER⁷, BIRTE KOLMOS⁷, MAGNUS UNEMO^{1,*,**} and STEEN HOFFMANN^{2,**}

¹WHO Collaborating Centre for Gonorrhoea and Other STIs, National Reference Laboratory for STIs, Department of Laboratory Medicine, Microbiology, Faculty of Medicine and Health, Örebro University, Örebro, Sweden; ²Department for Bacteria, Parasites and Fungi, Infectious Diseases Preparedness, Statens Serum Institut, Copenhagen, Denmark; ³Department of Clinical Microbiology, Hvidovre University Hospital, Hvidovre, Denmark; ⁴Department of Clinical Microbiology, Copenhagen University Hospital, Herlev, Denmark; ⁵Department of Clinical Microbiology, Aalborg University Hospital, Aalborg, Denmark; ⁶Department of Clinical Microbiology, Odense University Hospital, Odense, Denmark; ⁷Department of Clinical Microbiology, Vejle University Hospital, Vejle, Denmark

*Corresponding author: WHO Collaborating Centre for Gonorrhoea and Other STIs, National Reference Laboratory for STIs, Department of Laboratory Medicine, Microbiology, Örebro University Hospital, SE-701 85 Örebro, Sweden. Phone: +46-19-6022038. Fax: +46-19-127416. e-mail: magnus.unemo@regionorebrolan.se

**Joint Senior Authors

Running head: APTIMA COMBO 2 DIAGNOSTIC-ESCAPE MUTANTS IN DENMARK

Word Count: Abstract: 224 words; Text: 1669 words

This article has been accepted for publication and undergone full peer review but has not been through the copyediting, typesetting, pagination and proofreading process, which may lead to differences between this version and the [Version of Record](#). Please cite this article as [doi: 10.1111/APM.13043](https://doi.org/10.1111/APM.13043)

This article is protected by copyright. All rights reserved

SUMMARY

Hadad R, Jensen JS, Westh H, Grøn­bæk I, Schwartz LJ, Nielsen L, Vang TM, Nielsen R, Weinreich LS, Skov MN, Olsen M, Møller JK, Kolmos B, Unemo M, Hoffmann S. A *Chlamydia trachomatis* 23S rRNA G1523A variant escaping detection in the Aptima Combo 2 assay (Hologic) was widespread across Denmark in July-September 2019

Chlamydia trachomatis infection is the most common bacterial sexually transmitted infection globally, and nucleic acid amplification tests (NAATs) are recommended for highly sensitive and specific diagnosis. In early 2019, the Finnish new variant of *Chlamydia trachomatis* (FI-nvCT) was identified. The FI-nvCT has a C1515T mutation in the 23S rRNA gene, making it escaping detection in the Aptima Combo 2 (AC2; Hologic) NAAT, and the FI-nvCT has been subsequently reported in Sweden and Norway. In the present study, we investigated the presence of the FI-nvCT

and other AC2 diagnostic-escape CT mutants in July-September 2019 in Denmark. The FI-nvCT was present but rare in Denmark. However, another AC2 diagnostic-escape CT mutant (with a 23S rRNA G1523A mutation) was found to be widespread across Denmark, accounting for 95% (76/80) of AC2 diagnostic-escape nvCT samples from five Danish CT-diagnostic laboratories. This nvCT-G1523A has previously only been detected in one single sample in the United Kingdom and Norway, respectively. It is vital to monitor the continued stability of the NAAT targets in local, national and international settings and monitor as well as appropriately analyse incidence, unexplained shifts in diagnostics rates, and/or annual collections of samples diagnosed as negative/equivocal using NAATs with different target(s). Furthermore, diagnostic CT NAATs with dual target sequences are crucial and fortunately, an updated Hologic AC2 assay including one additional target sequence is in advanced development.

Key words: *Chlamydia trachomatis*, FI-nvCT, New variant, Aptima Combo 2 assay, 23S rRNA, C1515T, G1523A, Denmark

Magnus Unemo, WHO Collaborating Centre for Gonorrhoea and other STIs, Department of Laboratory Medicine, Microbiology, Örebro University Hospital, SE-701 85 Örebro, Sweden. e-mail: magnus.unemo@regionorebrolan.se

INTRODUCTION

Chlamydia trachomatis infection is the most common bacterial sexually transmitted infection globally (1), and nucleic acid amplification tests (NAATs) are recommended for highly sensitive and specific diagnosis (2,3). In early 2019, the Finnish new variant of *Chlamydia trachomatis* (FI-nvCT) was identified (4,5). The FI-nvCT has a C1515T mutation in the CT 23S rRNA gene, making it escaping detection in the Aptima Combo 2 (AC2; Hologic Inc., San Diego, CA, USA) NAAT that targets CT 23S rRNA (4-9). The FI-nvCT has been subsequently reported in Sweden (6) and Norway (7). Two additional very rare AC2 diagnostic-escape nvCTs were described in single specimens in Norway (7) and the United Kingdom (8). As the FI-nvCT, these two rare diagnostic-escape nvCTs also had a single nucleotide polymorphism (SNP) in the CT probe detection 23S rRNA sequence of AC2, that is, 23S rRNA C1514T or G1523A (7,8). The AC2 detection is measured in relative light units (RLUs) and the 23S rRNA mutations in these nvCTs result in low CT RLU values, mostly 15-99, that are interpreted as negative or equivocal in AC2 (4-9). The Aptima *C. trachomatis* NAAT (ACT; Hologic Inc.), which targets CT 16S rRNA, detects all these diagnostic-escape nvCTs, and reflex testing of AC2 samples with CT RLU values of 15-99 using ACT was implemented in May-June 2019 in many European countries (9,10).

Considering the spread of particularly the FI-nvCT in Finland (4,5), Sweden (6) and Norway (7); the close geographic proximity and frequent travelling between the Nordic countries; and the rapid risk assessment published by the European Centre for Disease Prevention and Control (ECDC) on 17 June 2019 (10), it was urgent to appropriately investigate if the FI-nvCT and/or additional AC2 diagnostic-escape nvCTs were spreading also in Denmark. In Denmark, with over 5.8 million inhabitants, the majority (62% in 2019) of CT samples are analysed using AC2 on Panther instruments (Hologic Inc.).

MATERIALS AND METHODS

The numbers of mandatorily reported CT cases in Denmark in 2018 and 2019 were obtained at the Statens Serum Institut, Copenhagen, Denmark. Consecutive clinical AC2 specimens obtained at

five CT-diagnostic clinical microbiology laboratories across Denmark, i.e. in Hvidovre, Herlev, Odense, Vejle and Aalborg (all AC2 laboratories in Denmark and together responsible for 62% of the CT diagnostic testing in 2019), from 1 July to 30 September 2019 were assessed. Samples with AC2 CT RLU values of 15-97 were reflex tested using the ACT. AC2 and ACT were performed in accordance with the instructions from the manufacturer (Hologic Inc., San Diego, CA, USA). Confirmation of AC2 diagnostic-escape nvCTs (one sample per patient) was performed by sequencing of the V2 region of the 23S rRNA gene at the WHO Collaborating Centre for Gonorrhoea and other STIs, Sweden, as described previously (5).

RESULTS

Diagnostics and incidence of *Chlamydia trachomatis* in Denmark

The number of CT cases in Denmark in 2019 compared to 2018 was similar for each month until ACT reflex testing of AC2 samples with CT RLU values of 15-99, in accordance with the ECDC rapid risk assessment (10), was initiated in mid- to late-June 2019 (FIG. 1). After ACT reflex testing started, an increase in the number of CT positive samples was seen compared to the corresponding months in 2018, i.e. an increase by 461 CT positive cases in July 2019, 486 cases in August and 364 cases in September (FIG. 1). However, the significance of these findings is not clear due to the fluctuations of number of positive samples, particularly in different months and seasons, such as summer, and over the years.

Possible cases of AC2 diagnostic-escape nvCTs in Denmark

Consecutive clinical AC2 specimens obtained at five CT-diagnostic clinical microbiology laboratories across Denmark, i.e. in Hvidovre, Herlev, Odense, Vejle and Aalborg, from 1 July to 30 September 2019 were assessed. During these months, 66552 samples (Hvidovre: 34737, Herlev: 12717, Odense: 6850, Vejle: 4184, Aalborg: 8064) were examined and 6117 samples (9.2%) were positive for CT in AC2 (Hvidovre: 2726, Herlev: 1210, Odense: 687, Vejle: 504, Aalborg: 990). In total, 707 (1.1%) samples with AC2 CT RLU values of 15-97 were reflex tested using the ACT. Of these samples (n=707), 204 (28.9%) were positive in ACT (RLUs 94-8891) (FIG. 2) and accordingly possible cases of AC2 diagnostic-escape nvCTs (Hvidovre: 107, Herlev: 35, Odense: 37, Vejle: 6, Aalborg: 19).

Confirmed cases of AC2 diagnostic-escape nvCTs in Denmark

In total, samples from 150 of the above mentioned possible cases of AC2 diagnostic-escape nvCTs (Hvidovre: 85, Herlev: 20, Odense: 28, Vejle: 4, Aalborg: 13) were available for sequencing (the vast majority from July-August 2019). Eighty (53.3%) of these samples contained an nvCT with a SNP in the AC2 CT probe detection sequence of the 23S rRNA gene (12 (8.0%) contained CT wild type and 58 (38.7%) had too low bacterial load for sequencing). Notably, nine (6%) of the 150 possible cases of AC2 diagnostic-escape nvCTs were also positive for *Neisseria gonorrhoeae*. Of the 80 AC2 diagnostic-escape nvCT samples, 95% (n=76) contained nvCT with the G1523A SNP in the 23S rRNA gene (RLUs: 16-86), 2.5% (n=2) 23S rRNA C1514T (RLUs: 17 and 25), and 2.5% (n=2) 23S rRNA C1515T (RLUs: 21 and 26), that is, the FI-nvCT. The AC2 diagnostic-escape nvCT G1523A was widely spread across Denmark and detected in Hvidovre (n=43), Herlev (n=15), Odense (n=7), Vejle (n=4), and Aalborg (n=7). The nvCT C1514T and the FI-nvCT was found in one sample each in Hvidovre and Odense, respectively. Gender distribution was 97 females, 50 males and three unknown. No information regarding the age, sexual orientation or recent travel of the patients was available for the present study.

DISCUSSION

In the present study, we identified in Denmark all three previously described AC2 diagnostic-escape nvCTs causing false-negative or equivocal AC2 results (4-8). The FI-nvCT was initially published in late-May 2019 in Finland (4) and the prevalence of FI-nvCT appeared to be relatively high in particular in South-West Finland (4,5). FI-nvCT was subsequently reported to be present in Sweden (6) and widely spread in Norway (7). Two additional AC2 diagnostic-escape nvCTs, with 23S rRNA C1514T or G1523A gene mutation, were also found in one sample each in Norway (7) and the United Kingdom (8), respectively. Already in mid-June 2019, recommendations to ACT reflex test all specimens with AC2 CT RLU values of 15-99 (possible AC2 diagnostic-escape nvCT) were published (9,10). Most worryingly, in Denmark 0.3% of all samples and 3.3% of all CT positive samples examined with AC2 from July to September 2019 were possible cases of an AC2 diagnostic-escape nvCT. Furthermore, 95% (n=76) of the confirmed AC2 diagnostic-escape nvCT samples, corresponding to 0.2% of all examined AC2 samples or 1.9% of all true AC2 CT positive samples (likely underestimated because 39% of samples could not be sequenced) from July to August 2019, contained the 23S rRNA G1523A nvCT. Surprisingly, this widely spread nvCT-1523A has still not been detected in the neighbouring country Sweden (6,unpublished) and only two samples in Denmark contained the FI-nvCT despite that this variant is widely spread in

Finland (4,5) and Norway (7), and not exceedingly rare in Sweden (6,unpublished). This indicates that the spread of CT strains between Denmark and the additional Nordic countries is more limited than anticipated, which was also indicated as one of the contributing factors more than 10 years ago when the Swedish nvCT was widely spreading in Sweden but very rare in Denmark (11-13). An Aptima-based FI-nvCT specific assay has now been developed and will be used in a pan-European surveillance study (9). Based on the present results from Denmark, it is important to not only use the Aptima-based FI-nvCT specific assay in the pan-European study, but additionally perform 23S rRNA gene sequencing on AC2/ACT discordant samples from countries where other diagnostic-escape CT variants are widespread, such as in Denmark, where the proportion of AC2/ACT discordant samples negative in the FI-nvCT specific assay is high, or where the distribution of diagnostic-escape CT variants or AC2/ACT discordant samples is completely unknown. The results of the present study show that the ACT reflex testing currently recommended for AC2 samples with CT RLUs of 15-99 should be continuously performed (9,10), and the CT RLU values caused by the nvCT-1523A monitored, i.e. to ensure that the $RLU \geq 15$ cut-off captures also all nvCT-1523A samples. Ultimately, a validated and quality assured revised version of AC2 that detects all the diagnostic-escape CT variants is very important, and such NAAT is fortunately in advanced development (14).

Whole genome sequencing of the different AC2 diagnostic-escape nvCTs is in progress and this work will contribute to the understanding on how CT is evolving including the emergence, clonality, spread, and fitness advantages/disadvantages of these nvCTs. The presence of several different reported AC2 diagnostic-escape nvCTs illustrates the vulnerability of using a single target for detection of an infectious agent with nucleic acid amplification tests (NAATs). This was seen already with the experience of the Swedish nvCT initially detected in 2006 in Sweden where a deletion in the cryptic plasmid that included the target sequence for the Roche and Abbott CT NAATs led to diagnostic escape (11-13). Previous genomic, phenotypic and epidemiological studies on the Swedish nvCT did not reveal any altered biological fitness and instead showed that the rapid spread was most likely mainly due to the advantage of a strong diagnostic selection, introduction in a high-frequency transmitting population, and the lack of subsequent tracing and treatment of sexual contacts (11).

In conclusion, it is vital to monitor the continued stability of the NAAT targets in local, national and international settings and monitor as well as appropriately analyse incidence, unexplained shifts in diagnostics rates, and/or annual collections of samples diagnosed as

negative/equivocal using NAATs with different target(s). Furthermore, diagnostic CT NAATs with dual target sequences are crucial and fortunately, an updated Hologic AC2 assay including one additional target sequence is in advanced development (14). Finally, international and national surveillance programmes capturing NAAT diagnostic-escape variants for CT and other infectious agents are imperative.

Acknowledgements

We are grateful to all the involved staff at the contributing Danish laboratories for collecting data and performing the required laboratory analyses.

REFERENCES

1. Rowley J, Vander Hoorn S, Korenromp E, Low N, Unemo M, Abu-Raddad LJ, et al. Chlamydia, gonorrhoea, trichomoniasis and syphilis: global prevalence and incidence estimates, 2016. Bull World Health Organ 2019;**97**:548-62P.
2. Lanjouw E, Ouburg S, de Vries HJ, Stary A, Radcliffe K, Unemo M. 2015 European guideline on the management of *Chlamydia trachomatis* infections. Int J STD AIDS 2016;**27**:333-48.
3. Workowski KA, Bolan GA; Centers for Disease Control and Prevention. Sexually transmitted diseases treatment guidelines, 2015. MMWR Recomm Rep 2015;**64**(RR-03):1-137.
4. Rantakokko-Jalava K, Hokynar K, Hieta N, Keskitalo A, Jokela P, Muotiala A, et al. *Chlamydia trachomatis* samples testing falsely negative in the Aptima Combo 2 test in Finland, 2019. Euro Surveill 2019;**24**(22):1900298. <https://www.eurosurveillance.org/content/10.2807/1560-7917.ES.2019.24.22.1900298> PMID: 31164192
5. Hokynar K, Rantakokko-Jalava K, Hakanen A, Havana M, Mannonen L, Jokela P, et al. The Finnish new variant of *Chlamydia trachomatis* with a single nucleotide polymorphism in the 23S rRNA target escapes detection by the Aptima Combo 2 test. Microorganisms 2019;**7**(8). <https://www.mdpi.com/2076-2607/7/8/227> PMID: 31370214
6. Unemo M, Hansen M, Hadad R, Lindroth Y, Fredlund H, Puolakkainen M, et al. Finnish new variant of *Chlamydia trachomatis* escaping detection in the Aptima Combo 2 assay also present in Örebro County, Sweden, May 2019. Euro Surveill 2019;**24**(26).

<https://www.eurosurveillance.org/content/10.2807/1560-7917.ES.2019.24.26.1900370>

PMID: 31266590

7. Johansen TB, Kløvstad H, Rykkvin R, Herrfurth-Erichsen EB, Sorthe J, Njølstad G, et al. The 'Finnish new variant of *Chlamydia trachomatis*' escaping detection in the Aptima Combo 2 assay is widespread across Norway, June to August 2019. *Euro Surveill* 2019;**24**(42). <https://www.eurosurveillance.org/content/10.2807/1560-7917.ES.2019.24.42.1900592> PMID: 31640843
8. Roberts DJ, Davis GS, Cole MJ, Naik D, Maru H, Woodford N, et al. Prevalence of new variants of *Chlamydia trachomatis* escaping detection by the Aptima Combo 2 assay, England, June to August 2019. *Euro Surveill* 2019;**24**(38). <https://www.eurosurveillance.org/content/10.2807/1560-7917.ES.2019.24.38.1900557> PMID: 31552817
9. Unemo M, Getman D, Hadad R, Cole M, Thomson N, Puolakkainen M, et al. Letter to the editor: *Chlamydia trachomatis* samples testing falsely negative in the Aptima Combo 2 test in Finland, 2019. *Euro Surveill* 2019;**24**(24). <https://www.eurosurveillance.org/content/10.2807/1560-7917.ES.2019.24.24.1900354> PMID: 31213219
10. European Centre for Disease Prevention and Control (ECDC). Rapid risk assessment: *Chlamydia trachomatis* false-negative test results by Aptima Combo 2 CT/NG assay (Hologic) in the EU/EEA. Stockholm: ECDC; 2019
11. Unemo M, Seth-Smith HM, Cutcliffe LT, Skilton RJ, Barlow D, Goulding D, et al. The Swedish new variant of *Chlamydia trachomatis*: genome sequence, morphology, cell tropism and phenotypic characterization. *Microbiology* 2010;**156**(Pt 5):1394-404. <https://www.microbiologyresearch.org/content/journal/micro/10.1099/mic.0.036830-0> PMID: 20093289
12. Ripa T, Nilsson PA. A *Chlamydia trachomatis* strain with a 377-bp deletion in the cryptic plasmid causing false-negative nucleic acid amplification tests. *Sex Transm Dis* 2007;**34**:255-6.
13. Hoffmann S, Jensen JS. Mutant *Chlamydia trachomatis* in Denmark. *Euro Surveill* 2007;**12**(10):E7-8. Erratum in: *Euro Surveill*. 2007 Nov;**12**(11):E071108.4.

14. Getman D. Mechanism of F-nvCT effect on AC2 performance, design and validation of F-nvCT surveillance IUO assay. IUSTI 2019 European Congress, 5-7 September 2019, Tallinn, Estonia.

Figure legends

FIG 1. Number of *Chlamydia trachomatis* (CT) cases reported in Denmark from January to September 2018 and 2019, respectively. Reflex testing using the Aptima CT assay started in mid-June 2019, which is indicated by a black line.

FIG 2. Number of samples tested positive for *Chlamydia trachomatis* (CT) in Aptima Combo 2 assay (AC2; black), samples that qualified for reflex testing by Aptima CT assay (ACT; dark grey), and samples that tested positive in ACT (light grey). The number of samples examined in AC2 was 18499 in July, 24555 in August, and 23498 in September.

Figure 1




Figure 2

