



Aalborg Universitet

AALBORG UNIVERSITY
DENMARK

A simple, bed-side tool to assess evoked pressure pain intensity

Hostrup, Søren Nicolai Frederiksen; O'Neill, Søren Francis Dyhrberg; Larsen, Jesper Bie; Arendt-Nielsen, Lars; Petersen, Kristian Kjær

Published in:
Scandinavian Journal of Pain

DOI (link to publication from Publisher):
[10.1515/sjpain-2022-0055](https://doi.org/10.1515/sjpain-2022-0055)

Publication date:
2023

Document Version
Publisher's PDF, also known as Version of record

[Link to publication from Aalborg University](#)

Citation for published version (APA):
Hostrup, S. N. F., O'Neill, S. F. D., Larsen, J. B., Arendt-Nielsen, L., & Petersen, K. K. (2023). A simple, bed-side tool to assess evoked pressure pain intensity. *Scandinavian Journal of Pain*, 23(2), 382-388. Advance online publication. <https://doi.org/10.1515/sjpain-2022-0055>

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal -

Take down policy

If you believe that this document breaches copyright please contact us at vbn@aub.aau.dk providing details, and we will remove access to the work immediately and investigate your claim.

Original Experimental

Søren Nicolai Frederiksen Hostrup, Søren Francis Dyhrberg O'Neill, Jesper Bie Larsen, Lars Arendt-Nielsen and Kristian Kjær Petersen*

A simple, bed-side tool to assess evoked pressure pain intensity

<https://doi.org/10.1515/sjpain-2022-0055>

Received April 8, 2022; accepted July 31, 2022;

published online August 16, 2022

Abstract

Objectives: Existing equipment for quantitative sensory testing is generally expensive and not easily applicable in a clinical setting thus simple bed-side devices are warranted. Pressure hyperalgesia is a common finding in patients with musculoskeletal pain and an experimental model is delayed-onset muscle soreness (DOMS). DOMS is characterised by muscle hyperalgesia and some studies report facilitation of temporal summation of pain. This study aimed to detect DOMS induced muscle hyperalgesia and temporal summation of pain using a newly developed bed-side quantitative sensory testing device to deliver standardised pressure.

Methods: Twenty-two healthy participants participated in two sessions with the second session approximately 48 h after baseline. Pressure pain intensities were assessed from the gastrocnemius muscle with four probes calibrated to apply 2, 4, 6 and 8 kg, respectively. Temporal summation of pain (10 stimuli delivered at 0.5 Hz using the 6 kg probe) intensities were assessed from the same location. DOMS was evoked in the gastrocnemius muscle by an eccentric exercise. Sleepiness and physical activity were measured with the Epworth Sleepiness Scale and the Global Physical Activity Questionnaire to investigate if they were associated with the quantitative sensory testing measures.

Results: Pressure pain intensity was significantly increased 48 h after induction of DOMS when compared to baseline for all four probes ($p < 0.05$). Temporal summation of pain was not statistically significant affected by DOMS and sleep quality and physical activity did not associate with any of the measures.

Conclusions: This study introduces a simple, bed-side assessment tool for the assessment of pressure pain intensity and hence hyperalgesia and temporal summation of pain.

Keywords: bed-side assessments; delayed onset muscle soreness; hyperalgesia; Pressure pain thresholds; quantitative sensory testing.

*Corresponding author: Associate Professor Kristian Kjær Petersen, Dr. Med., PhD, M.Sc, Center for Neuroplasticity and Pain, SMI, Department of Health Science and Technology, Faculty of Medicine, Aalborg University, Aalborg, Denmark; and Center for Mathematical Modeling of Knee Osteoarthritis (MathKOA), Department of Material and Production, Faculty of Engineering and Science, Aalborg University, Fredrik Bajers Vej 7 D3, DK-9220, Aalborg, Denmark, Phone: +45 9940 7529, Fax: +45 9815 4008, E-mail: KKP@hst.aau.dk

Søren Nicolai Frederiksen Hostrup and Jesper Bie Larsen, Department of Health Science and Technology, Faculty of Medicine, Aalborg University, Aalborg, Denmark, E-mail: hostrup110@live.dk (S.N.F. Hostrup), jbl@hst.aau.dk (J.B. Larsen)

Søren Francis Dyhrberg O'Neill, Spine Centre of Southern Denmark, University Hospital of Southern Denmark, Middelfart, Denmark, E-mail: Soeren.ONeill@rsyd.dk

Lars Arendt-Nielsen, Center for Neuroplasticity and Pain, SMI, Department of Health Science and Technology, Faculty of Medicine, Aalborg University, Aalborg, Denmark; Department of Medical Gastroenterology, Mech-Sense, Aalborg University Hospital, Aalborg, Denmark; and Center for Mathematical Modeling of Knee Osteoarthritis (MathKOA), Department of Material and Production, Faculty of Engineering and Science, Aalborg University, Aalborg, Denmark, E-mail: lan@hst.aau.dk

Introduction

The estimated prevalence of chronic pain is 20% [1–3], it accounts for 15–20% of consultations to general practitioners [4, 5] and has negative consequences for the individuals' quality of life [2, 6].

The underlying mechanisms of chronic pain can be explored by mechanistic, quantitative sensory testing (QST) [7, 8]. Several standardised QST protocols exists [9–11] but many are time consuming, expensive and difficult to implement in clinical practice.

Pressure stimuli are among the most widely used QST measures specifically for assessing muscle hyperalgesia in

chronic musculoskeletal pain [7, 12, 13]. Different pressure algometers have been developed [14–16], but most commercially available pressure algometers are generally expensive and are mainly used for research purposes. If a simple, cheap and quick bed-side pressure algometer existed it might be more widely used in clinical settings for monitoring and quantification.

Previous studies have found lowered pressure pain thresholds (PPTs), and facilitated temporal summation of pain (TSP) in patients with e.g., severe osteoarthritis [17, 18], neck pain [19, 20], or chronic low back pain [21] although conflicting evidence exist for e.g., young adults with patellofemoral pain [22–24]. PPTs assessed over a painful area is considered to reflect peripheral sensitisation whereas PPTs assessed outside of a painful area can reflect widespread hyperalgesia [21]. TSP is a phenomenon of increased pain perception evoked by repetitive painful stimulation with fixed intensity, and it is considered to compose the perceptual equivalent of the neuronal wind-up responses in dorsal horn seen in animals, which is an indicator for central sensitisation [21].

A frequently utilised experimental muscle pain model for evoked pressure hyperalgesia, is delayed onset muscle soreness (DOMS) [25–32]. DOMS is evoked by intense and unaccustomed muscle contraction from physical activity and exposes the individual to a sub-clinical pain lasting 5–7 days, with peak intensity 24–72 h after exercise [33]. Studies have demonstrated that PPTs decrease as a consequence of DOMS [25–27, 30, 31] and some studies have found that TSP is facilitated by DOMS [28, 29]. The general level of physical activity for the individual might affect the degree to which DOMS is manifested after exercise [33]. Furthermore, PPTs has been demonstrated to decrease subsequently to 24 h of total sleep deprivation, while TSP was facilitated under the same conditions [34], suggesting that sleep might be another important factor to consider when assessing PPTs and TSP.

The aims of this study were (1) to investigate if a newly developed, bed-side, pressure algometer could quantify pressure pain intensity and TSP in healthy participants following DOMS and (2) if sleep quality and physical activity influenced pressure hyperalgesia and TSP.

Methods

This experimental study was composed of two sessions separated by approximately 48 h.

Session 1 (baseline): The participants completed the Epworth Sleepiness Scale (ESS) and the Global Physical Activity Questionnaire (GPAQ) and were assessed for pressure pain intensity using the four

probes, followed by assessment of TSP. Session 1 was concluded by having the participant perform an eccentric calf-raise exercise to evoke DOMS in the lower leg. *Session 2* (48 h after induction of delayed onset muscle soreness, DOMS): This session was similar to session 1, except for no GPAQ and calf-raise exercise.

Participants

A previous study found an average decrease in PPT from 375 kPa (SD 93) to 325 kPa (SD 128) with DOMS [30]. Consequently, the present sample size calculations (conducted using the G*power version 3.1 software, Kiel University, Germany) were performed with an effect size at 0.55, 0.05 alpha level and 80% power, thus 22 participants were required. A total of 25 participants were recruited to account for potential drop-outs.

Healthy participants aged 18–45 years were recruited through notices on social media and community boards. Exclusion was warranted if they reported any of the following: drug or alcohol addiction; current use of medications, which might affect the trial (e.g., analgesics and anti-inflammatory drugs); previous or current history of chronic musculoskeletal, neurological, pulmonary, cardiac, or chronic pain conditions as well as mental illness; recent or acute pain; consumption of stimulants or painkillers on the morning of the experiment; lack of ability to cooperate.

Participants who reported mild or no perception of DOMS were excluded, where mild DOMS were defined as “Presence of tenderness, but no pain or difficulties associated with movement and daily activities”. The study was approved by the North Denmark Region Committee on Health Research Ethics, case number N-20170088, and conducted in accordance with the Helsinki Declaration. All participants signed informed consent prior to participation.

Pressure probe assessments

The probes were composed of a cylinder containing a spring whose tension could be adjusted by a screw thread (Figure 1). A wing nut secured the spring tension level which determined the resulting pressure delivered by the probe.

Each probe (sphere of 15 mm diameter) was tested each day before application and calibrated, if needed, to ensure the delivered pressure matched the intended pressure. The probes were calibrated using a 10 kg Strain Gauge Load Cell force sensor from Adafruit (Adafruit Industries, New York City, New York USA). The sensor measurements were read by an Arduino UNO through an HX711 Load Cell Amplifier and ADC Module (SparkFun Electronics, Niwot, Colorado USA).

The participants lay on their abdomen during the pressure stimuli. The stimuli were applied to the muscle belly of the medial gastrocnemius muscle. The exact point of stimulation was determined through palpation to locate a point of stability. All assessments were performed by the same assessor (Søren Nicolai Frederiksen Hostrup).

Pressure pain was applied sequentially in order of 2, 4, 6, and 8 kg to the calf on the dominant leg. The evoked pain intensity by each pressure probe was rated by the participant on a Visual Analog Scale (VAS) ranging from 0 to 10, where 0 corresponded to no pain and 10 corresponded to worst pain imaginable.

TSP was applied to the dominant calf with the 6 kg probe and similar to previous studies, 10 stimulations were given with approximately 1 s



Figure 1: The four probes utilised in the present study with the respective calibrated pressure. The probes are approximately 0.23 m long. Developed and calibrated by University of Southern Denmark, Odense, Denmark.

duration and 1 s between stimulations [35]. The pain intensities of the initial and final stimulation were rated on the VAS and the difference between the ratings was defined as the TSP score.

DOMS

DOMS was evoked by having the participants performing calf-raises standing on an elevation with the dominant leg, with no additional load to the body weight, similar to previous studies [31, 32]. The toes were at the edge of the elevation whilst the heel was alternately raised and lowered. The participants were asked to do a maximum of 4×30 repetitions (30 s break between sets) with full range of motion. The participant was supported in case of fatigue before reaching the maximum repetitions, however, similar to previous methodology [32] the set was concluded if the participant was incapable of performing the exercise with full range motion.

Questionnaires

The ESS [36] assessed sleep quality as daytime sleepiness. The ESS has differentiated between individuals diagnosed with sleep disorders and healthy individuals [36]. The propensity of falling asleep was evaluated in the range of 0–3 in eight everyday scenarios, where 0 corresponded to “would never doze” and 3 corresponded to “high chance of dozing”. Thus, total scores range from 0 to 24, with higher scores indicating worse quality of sleep.

The GPAQ is a questionnaire constructed by The World Health Organization designed to assess physical activity [37]. The GPAQ measures physical activity across three domains of everyday life in moderate and vigorous intensity levels, thus providing a wide measure of the individual’s activity level [37]. The results of the GPAQ have been found to provide an acceptable valid measure of physical activity levels [37–39]. Each question was posed to the participant, and the questionnaire was answered in continuous dialog between

the experimenter and the participant. Each intensity level in each domain was converted to a Metabolic Equivalent (MET)-value as prescribed by the GPAQ guidelines [30]. The sum of MET-values composed the resulting activity level (GPAQ score) of the participant, with higher scores indicating greater physical activity.

Statistical analysis

All statistical analyses were performed with the IBM SPSS Statistics 26 software, with $p < 0.05$ considered as a statistically significant finding. All VAS data were analysed with a Shapiro–Wilk test for normality. The data are presented as mean \pm standard deviation.

Pressure pain intensity and TSP in session 1 and 2 were compared with paired t-tests and in case of non-gaussian data distribution Wilcoxon Signed Rank Tests were utilised to assess the difference between the measures. To account for the possible impact of physical activity and sleep quality, one-way repeated measures ANCOVAs were applied to the pressure pain intensity and TSP paradigm. Physical activity was denoted as the calculated MET-values and daytime sleepiness was denoted as the difference in ESS score between session 1 and session 2. VAS baseline and VAS DOMS were dependent variables, time were the independent variable and GPAQ score or ESS score were respectively covariate variables in the ANCOVAs.

Results

Twenty-five participants were recruited, 24 completed both sessions and two were excluded from the analysis as the exercise failed to evoke DOMS. DOMS were evoked in the remaining 22 participants that were included for statistical analysis. Table 1 illustrates demographic information about the participants.

Pressure pain intensity and temporal summation of pain

Significant higher pressure pain intensities were found in the presence of DOMS for the 2 kg ($p=0.014$), 4 kg ($p=0.016$), 6 kg ($p=0.031$) and 8 kg ($p=0.002$) probes when compared to baseline (Figure 2).

Table 1: Participant demographics and questionnaire responses.

Characteristic	Values
Age	24.8 \pm 1.8
Sex	Male 11; Female 11
BMI	24.1 \pm 2.3
Dominant leg	Right 21; Left 1
ESS score	Baseline 5.0 \pm 3.2; DOMS 4.8 \pm 3.2
GPAQ score	2,763 \pm 2,065

Data are reported as mean \pm standard deviation. BMI, body mass index; ESS, Epworth Sleepiness Scale; GPAQ, global physical activity questionnaire.




Figure 2: Pressure pain intensity assessed by the visual analog scale (VAS) from the application of different probes comparing session 1 (baseline) with session 2 (48 h after induction of delayed onset muscle soreness, DOMS). *illustrates $p < 0.05$ comparing baseline and DOMS sessions.

Figure 2 shows the pressure pain intensity comparison between session 1 and 2.

No statistically significant difference was found in TSP between baseline 1.32 ± 2.34 and after induction of DOMS 1.50 ± 2.04 ($p = 0.639$).

Sleep and physical activity

Pressure pain intensity measurements and TSP were analysed for differences between baseline and DOMS controlling for covariance of sleep quality (ESS score) and physical activity (GPAQ score) and no statistically significant association was found (Table 2).

Discussion

This study applied a newly developed, bed-side, assessment tool for assessing pressure pain intensity and temporal

Table 2: The column ‘Pressure probe’ refers to the specific utilised pressure probe and the pain assessment paradigm by which it was applied.

Pressure probe	Time-ESS [P-Value]	Time-GPAQ [P-Value]
2 kg pressure	0.962	0.976
4 kg pressure	0.766	0.817
6 kg pressure	0.618	0.950
8 kg pressure	0.274	0.942
6 kg pressure used for temporal summation of pain	0.229	0.651

The column ‘Time-ESS’ is the resulting p -values from the ANCOVA with ESS as covariate for baseline-DOMS comparison, the same is shown for the column ‘Time-GPAQ’, though with GPAQ as covariate. ESS, Epworth Sleepiness Scale; GPAQ, global physical activity questionnaire.

summation of pain. DOMS was found to significantly increase pressure pain intensity when utilising each of the four calibrated probes. Sleep quality and physical activity levels showed no statistically significant impact on pressure pain intensity and TSP.

Pressure hyperalgesia and temporal summation of pain

Many studies have demonstrated that pressure pain thresholds decreases in the presence of DOMS [25–27, 30, 31] making this a validated experimental muscle pain model. The present assessment tool was capable of assessing muscle hyperalgesia after induction of DOMS and hence may have potential for clinical applications. Future studies are warranted to confirm the reliability of the probes and potentially to assess the importance of e.g., other probe diameters.

TSP has been an experimental proxy for assessment of sensitisation most likely of a central origin [40]. Facilitated TSP has been found in a range of patients with chronic pain including fibromyalgia [41], osteoarthritis [42], low back and neck pain [19], irritable bowel syndrome patients [43] and some neuropathic chronic pain patients [44], and evidence suggest that TSP is predictive for chronic post-operative pain and response to weak analgesics [7, 45–47]. Thus, including TSP in bed-side settings might provide complementary clinical information and have predictive value. Conflicting evidence suggests that TSP is facilitated [28, 29], or unaffected [48, 49] by DOMS. Previous studies have argued that facilitation of TSP require continuous peripheral drive for a long duration (years) [8], such as seen in e.g., severe osteoarthritis [8, 18]. DOMS is a short-lasting pain model at low intensity, which could explain why the current study does not find DOMS to alter the TSP response.

Sleep and physical activity

Impaired sleep quality is associated with many chronic pain conditions [50], with a minimum of 50% chronic pain patients affected by sleep impairments [51]. Additionally, co-occurrence of sleep impairments might enhance chronic pain symptoms [51]. Assessment of pressure pain tolerance thresholds shows that total sleep deprivation induces pressure pain sensitivity [34, 52]. By contrast, the present study showed no evidence of facilitated pain sensitivity in relation to sleep deprivation. However, the participants in this study were not sleep deprived as indicated by ESS scores similar to those seen in healthy

controls [36]. Since DOMS is the product of unaccustomed eccentric exercise, it might be generally associated with physical activity levels [33]. However, the present findings showed no significant relation between DOMS related evoked pain and physical activity levels. Nonetheless, DOMS was successfully induced in all included participants which might explain why physical activity levels did not account for any significant variance in pain measures between baseline and DOMS.

Limitations

DOMS was only reported as present or absent in the current study and the degree of DOMS per se was not quantified. No participants could be categorised as sleep deprived, and the measure for sleep quality simply reflected the difference in self-reported daytime sleepiness between the two sessions, and the measure is not sufficiently sensitive to assess minor changes in sleepiness.

This preliminary study was initiated to assess if the newly developed bed-side pressure algometer could detect pressure hyperalgesia. It is however important to assess reliability of the assessments and to investigate the clinical usefulness prior to implementation.

Conclusions

This study evaluated a newly developed bed-side pressure algometer and confirmed that pressure hyperalgesia can be assessed in healthy individuals following delayed onset muscle soreness. Sleep quality and physical activity levels did not affect pressure pain intensity and temporal summation of pain. Future studies are needed to assess the reliability of the tool.

Research funding: The authors thank Aalborg University Talent Management Programme (j.no. 771126) and the European Union, Horizon 2020 Grant (Project title: Painless) for providing the funding to initiate this work. Center for Neuroplasticity and Pain (CNAP) is supported by the Danish National Research Foundation (DNRF121). Center for Mathematical Modeling of Knee Osteoarthritis (MathKOA) is funded by the Novo Nordisk Foundation (grant no. NNF21OC0065373).

Author contributions: The authors have accepted responsibility for the entire content of this manuscript and approved its submission.

Competing interests: The authors have no conflicts of interest to declare.

Informed consent: Informed consent has been obtained from all individuals included in this study.

Ethical approval: This research, involving human subjects, complied with all relevant national regulations, institutional policies and is in accordance with the tenets of the Helsinki Declaration (as amended in 2013), and has been approved by the authors and by the North Denmark Region Committee on Health Research Ethics, case number N-20170088.

References

- Gureje O, Von Korff M, Simon GE, Gater R. Persistent pain and well-being: a World Health Organization study in primary care. *JAMA* 1998;280:147–51.
- Breivik H, Collett B, Ventafridda V, Cohen R, Gallacher D. Survey of chronic pain in Europe: prevalence, impact on daily life, and treatment. *Eur J Pain* 2006;10:287–333.
- Dahlhamer J, Lucas J, Zelaya C, Nahin R, Mackey S, DeBar L, et al. Prevalence of chronic pain and high-impact chronic pain among adults — United States, 2016. *Morb Mortal Wkly Rep* 2018;67: 1001–6.
- Mäntyselkä P, Kumpusalo E, Ahonen R, Kumpusalo A, Kauhanen J, Viinamäki H, et al. Pain as a reason to visit the doctor: a study in Finnish primary health care. *Pain* 2001;89: 175–80.
- Koleva D, Krulichova I, Bertolini G, Caimi V, Garattini L. Pain in primary care: an Italian survey. *Eur J Publ Health* 2005;15: 475–9.
- McCarberg BH, Nicholson BD, Todd KH, Palmer T, Penles L. The impact of pain on quality of life and the unmet needs of pain management: results from pain sufferers and physicians participating in an internet survey. *Am J Therapeut* 2008;15: 312–20.
- Petersen KK, Vaegter HB, Stubhaug A, Wolff A, Scammell BE, Arendt-Nielsen L, et al. The predictive value of quantitative sensory testing: a systematic review on chronic postoperative pain and the analgesic effect of pharmacological therapies in patients with chronic pain. *Pain* 2021;162:31–44.
- Graven-Nielsen T, Arendt-Nielsen L. Assessment of mechanisms in localized and widespread musculoskeletal pain. *Nat Rev Rheumatol* 2010;6:599–606.
- Dyck PJ, Zimmerman I, Gillen DA, Johnson D, Karnes JL, O'Brien PC. Cool, warm, and heat-pain detection thresholds: testing methods and inferences about anatomic distribution of receptors. *Neurology* 1993;43:1500.
- Rolke R, Magerl W, Campbell KA, Schalber C, Caspari S, Birklein F, et al. Quantitative sensory testing: a comprehensive protocol for clinical trials. *Eur J Pain* 2006;10:77–88.
- Wilder-Smith OHG. A paradigm-shift in pain medicine; 2013: 1–107 pp. Available from: http://riverpublishers.com/book_details.php?book_id=222 [Accessed 16 Oct 2021].
- Neziri AY, Limacher A, Jüni P, Radanov BP, Andersen OK, Arendt-Nielsen L, et al. Ranking of tests for pain hypersensitivity according to their discriminative ability in chronic neck pain. *Reg Anesth Pain Med* 2013;38:308–20.

13. Georgopoulos V, Akin-Akinyosoye K, Zhang W, McWilliams DF, Hendrick P, Walsh DA. Quantitative sensory testing (QST) and predicting outcomes for musculoskeletal pain, disability and negative affect: a systematic review and meta-analysis. *Pain* 2019;160:1920–32.
14. Kinser AM, Sands WA, Stone MH. Reliability and validity of a pressure algometer. *J Strength Condit Res* 2009;23:312–4.
15. Vaughan B, McLaughlin P, Gosling C. Validity of an electronic pressure algometer. *Int J Osteopath Med* 2007;10:24–8.
16. Jerez-Mayorga D, dos CFA, de CMM, Fernandes IG, Aedo-Muñoz E, Intelangelo L, et al. Instrumental validity and intra/inter-rater reliability of a novel low-cost digital pressure algometer. *PeerJ* 2020;8:e10162.
17. Petersen KK, Arendt-Nielsen L, Finocchietti S, Hirata RP, Simonsen O, Laursen MB, et al. Age interactions on pain sensitization in patients with severe knee osteoarthritis and controls. *Clin J Pain* 2017;33:1081–7.
18. Arendt-Nielsen L, Egsgaard LL, Petersen KK, Eskehave TN, Graven-Nielsen T, Hoeck HC, et al. A mechanism-based pain sensitivity index to characterize knee osteoarthritis patients with different disease stages and pain levels. *Eur J Pain* 2015;19:1406–17.
19. Manresa JAB, Neziri AY, Curatolo M, Arendt-Nielsen L, Andersen OK. Reflex receptive fields are enlarged in patients with musculoskeletal low back and neck pain. *Pain®* 2013;154:1318–24.
20. Rebeck T, Moloney N, Azoozy R, Hübscher M, Waller R, Gibbons R, et al. Clinical ratings of pain sensitivity correlate with quantitative measures in people with chronic neck pain and healthy controls: cross-sectional study. *Phys Ther* 2015;95:1536–46.
21. Arendt-Nielsen L, Morlion B, Perrot S, Dahan A, Dickenson A, Kress HG, et al. Assessment and manifestation of central sensitisation across different chronic pain conditions. *Eur J Pain* 2018;22:216–41.
22. Rathleff MS, Rathleff CR, Stephenson A, Mellor R, Matthews M, Crossley K, et al. Adults with patellofemoral pain do not exhibit manifestations of peripheral and central sensitization when compared to healthy pain-free age and sex matched controls – an assessor blinded cross-sectional study. *PLoS One* 2017;12:e0188930.
23. Rathleff MS, Petersen KK, Arendt-Nielsen L, Thorborg K, Graven-Nielsen T. Impaired conditioned pain modulation in young female adults with long-standing patellofemoral pain: a single blinded cross-sectional study. *Pain Med* 2016;17:980–8.
24. Bartholomew C, Edwards L, Lack S. Pressure pain thresholds in adults with patellofemoral pain and patellofemoral joint osteoarthritis: a case-control study. *Scand J Pain* 2019;19:713–23.
25. Lau WY, Muthalib M, Nosaka K. Visual analog scale and pressure pain threshold for delayed onset muscle soreness assessment. *J Musculoskelet Pain* 2013;21:320–6.
26. Binderup AT, Arendt-Nielsen L, Madeleine P. Pressure pain threshold mapping of the trapezius muscle reveals heterogeneity in the distribution of muscular hyperalgesia after eccentric exercise. *Eur J Pain* 2010;14:705–12.
27. Andersen H, Arendt-Nielsen L, Danneskiold-Samsøe B, Graven-Nielsen T. Pressure pain sensitivity and hardness along human normal and sensitized muscle. *Somatosens Mot Res* 2006;23:97–109.
28. Nie H, Arendt-Nielsen L, Madeleine P, Graven-Nielsen T. Enhanced temporal summation of pressure pain in the trapezius muscle after delayed onset muscle soreness. *Exp Brain Res* 2006;170:182–90.
29. Nie H, Madeleine P, Arendt-Nielsen L, Graven-Nielsen T. Temporal summation of pressure pain during muscle hyperalgesia evoked by nerve growth factor and eccentric contractions. *Eur J Pain* 2009;13:704–10.
30. Black CD, Tynes BK, Gonglach AR, Waddell DE. Local and generalized endogenous pain modulation in healthy men: effects of exercise and exercise-induced muscle damage. *Pain Med* 2016;17:2422–33.
31. Gervasio S, Finocchietti S, Stevenson AJT, Mrachacz-Kersting N. Delayed muscle onset soreness in the gastrocnemius muscle attenuates the spinal contribution to interlimb communication. *Eur J Appl Physiol* 2018;118:2393–402.
32. Kristensen NS, Hertel E, Skadhauge CH, Kronborg SH, Petersen KK, McPhee ME. Psychophysical predictors of experimental muscle pain intensity following fatiguing calf exercise. *PLoS One* 2021;16:e0253945.
33. Cheung K, Hume PA, Maxwell L. Delayed onset muscle soreness. *Sports Med* 2003;33:145–64.
34. Staffe AT, Bech MW, Clemmensen SLK, Nielsen HT, Larsen DB, Petersen KK. Total sleep deprivation increases pain sensitivity, impairs conditioned pain modulation and facilitates temporal summation of pain in healthy participants. *PLoS One* 2019;14:e0225849.
35. Petersen KK, Jensen MB, Graven-Nielsen T, Hauerslev LV, Arendt-Nielsen L, Rathleff MS. Pain catastrophizing, self-reported disability, and temporal summation of pain predict self-reported pain in low back pain patients 12 weeks after general practitioner consultation: a prospective cohort study. *Clin J Pain* 2020;36:757–63.
36. Johns MW. A new method for measuring daytime sleepiness: the Epworth Sleepiness Scale. *Sleep* 1991;14:540–5.
37. Armstrong T, Bull F. Development of the World Health Organization global physical activity questionnaire (GPAQ). *J Public Health* 2006;14:66–70.
38. Cleland CL, Hunter RF, Kee F, Cupples ME, Sallis JF, Tully MA. Validity of the global physical activity questionnaire (GPAQ) in assessing levels and change in moderate-vigorous physical activity and sedentary behaviour. *BMC Publ Health* 2014;14:1255.
39. Laeremans M, Dons E, Avila-Palencia I, Carrasco-Turigas G, Orjuela JP, Anaya E, et al. Physical activity and sedentary behaviour in daily life: a comparative analysis of the global physical activity questionnaire (GPAQ) and the sensewear armband. *PLoS One* 2017;12:e0177765.
40. Woolf CJ. Central sensitization: implications for the diagnosis and treatment of pain. *Pain* 2011;152(3 Suppl):S2–15.
41. Staud R, Cannon RC, Mauderli AP, Robinson ME, Price DD, Vierck CJ. Temporal summation of pain from mechanical stimulation of muscle tissue in normal controls and subjects with fibromyalgia syndrome. *Pain* 2003;102:87–95.
42. Arendt-Nielsen L, Skou ST, Nielsen TA, Petersen KK. Altered central sensitization and pain modulation in the CNS in chronic joint pain. *Curr Osteoporos Rep* 2015;13:225–34.
43. Rössel P, Drewes AM, Petersen P, Nielsen J, Arendt-Nielsen L. Pain produced by electric stimulation of the rectum in patients with irritable bowel syndrome: further evidence of visceral hyperalgesia. *Scand J Gastroenterol* 1999;34:1001–6.

44. Maier C, Baron R, Tölle TR, Binder A, Birbaumer N, Birklein F, et al. Quantitative sensory testing in the German research network on neuropathic pain (DFNS): somatosensory abnormalities in 1236 patients with different neuropathic pain syndromes. *Pain* 2010; 150:439–50.
45. Petersen KK, Arendt-Nielsen L, Simonsen O, Wilder-Smith O, Laursen MB. Presurgical assessment of temporal summation of pain predicts the development of chronic postoperative pain 12 months after total knee replacement. *Pain* 2015;156:55–61.
46. Petersen KK, Olesen AE, Simonsen O, Arendt-Nielsen L. Mechanistic pain profiling as a tool to predict the efficacy of 3-week nonsteroidal anti-inflammatory drugs plus paracetamol in patients with painful knee osteoarthritis. *Pain* 2019;160: 486–92.
47. Izumi M, Petersen KK, Laursen MB, Arendt-Nielsen L, Graven-Nielsen T. Facilitated temporal summation of pain correlates with clinical pain intensity after hip arthroplasty. *Pain* 2017;158: 323–32.
48. Lo Vecchio S, Petersen LJ, Finocchietti S, Gazerani P, Arendt-Nielsen L, Graven-Nielsen T. The effect of combined skin and deep tissue inflammatory pain models. *Pain Med* 2015;16:2053–64.
49. McPhee M, Graven-Nielsen T. Alterations in temporal summation of pain and conditioned pain modulation across an episode of experimental exercise-induced low back pain. *J Pain* 2019;20: 264–76.
50. Mathias JL, Cant ML, Burke ALJ. Sleep disturbances and sleep disorders in adults living with chronic pain: a meta-analysis. *Sleep Med* 2018;52:198–210.
51. Smith MT, Haythornthwaite JA. How do sleep disturbance and chronic pain inter-relate? Insights from the longitudinal and cognitive-behavioral clinical trials literature. *Sleep Med Rev* 2004;8:119–32.
52. Onen SH, Alloui A, Gross A, Eschallier A, Dubray C. The effects of total sleep deprivation, selective sleep interruption and sleep recovery on pain tolerance thresholds in healthy subjects. *J Sleep Res* 2001;10:35–42.