



Aalborg Universitet

AALBORG UNIVERSITY
DENMARK

Risk of complications following external cardioversion in cardiovascular implantable electronic devices with and without generator replacement

Elgaard, AF; Dinesen, PT; Riahi, S; Lundbye-Christensen, S; Johansen, JB; Nielsen, JC; Larsen, JM

Published in:
Europace

DOI (link to publication from Publisher):
[10.1093/europace/euad122.075](https://doi.org/10.1093/europace/euad122.075)

Creative Commons License
CC BY-NC-ND 4.0

Publication date:
2023

Document Version
Publisher's PDF, also known as Version of record

[Link to publication from Aalborg University](#)

Citation for published version (APA):
Elgaard, AF., Dinesen, PT., Riahi, S., Lundbye-Christensen, S., Johansen, JB., Nielsen, JC., & Larsen, JM. (2023). Risk of complications following external cardioversion in cardiovascular implantable electronic devices with and without generator replacement. *Europace*, 25(Suppl. 1), Article euad122.075. <https://doi.org/10.1093/europace/euad122.075>

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal -

Take down policy

If you believe that this document breaches copyright please contact us at vbn@aub.aau.dk providing details, and we will remove access to the work immediately and investigate your claim.

Risk of complications following external cardioversion in cardiovascular implantable electronic devices with and without generator replacement

Mr Elgaard AF; Ms Dinesen PT; Professor Riahi S; Associate Professor Lundbye-Christensen S; Associate Professor Johansen JB; Professor Nielsen JC; Associate Professor Larsen JM.

Aalborg University Hospital, Department of Cardiology and Department of Clinical Medicine, Aalborg, Denmark

Aalborg University Hospital, Department of Anaesthesia and Intensive Care Medicine, Aalborg, Denmark

Aalborg University Hospital, Department of Research data and Statistics, Aalborg, Denmark

Odense University Hospital, Department of Cardiology, Odense, Denmark

Aarhus University Hospital, Department of Cardiology and Department of Clinical Medicine, Aarhus N, Denmark

Funding Acknowledgements: Type of funding sources: Foundation. Main funding source(s): Karl G. Andersen Foundation

Introduction: External cardioversion (ECV) with transthoracic shock is a standard procedure to control atrial fibrillation and flutter in patients regardless of whether cardiovascular implantable electronic devices (CIED) are in situ. Studies have demonstrated infrequent ECV-related CIED malfunction, requiring device re-interventions, but the mechanisms seem heterogeneous. We hypothesized, that replaced CIED generators are more vulnerable to ECV procedures, maybe caused by insulation defects. The aim of this study is to investigate the long-term risk of complications following ECV in CIEDs with or without previous generator replacement.

Methods: All CIED implants and re-operations in Denmark were identified in the Danish Pacemaker and ICD Register and ECVs were identified in the Danish National Patient Registry from January 2010 to February 2019. We identified CIED patients with and without previous generator replacement. Inclusion was at the time of first ECV. Time to generator replacement with the same type and lead re-intervention with the same lead type after ECV were calculated independently, using Cox regression with death, generator extraction, and system up-/down-grade being competing events. The cause-specific hazard ratios (HR) were calculated using inverse probability of treatment balancing patient age and sex and stratified for the different types of CIED and lead, allowing different baseline hazard functions (1).

Results: We identified 515 CIED patients with previous generator replacement and 1,878 CIED patients without generator replacement. Mean age of patients at inclusion was 68.7 ± 11.7 years with 74% men and the median generator and lead implant times were 1.6 years (0.9 years without replacement) and 3.0 years (0.7 years without replacement) before ECV. The types of CIED included 50% of pacemakers, 30% of Implantable Cardioverter Defibrillators, 6% of Cardiac Resynchronization Therapy-Pacemaker, and 14% of Cardiac Resynchronization Therapy-Defibrillators. During the first 3 years of follow-up, 92 (18%) of the patients with replaced generator had an additional generator replacement compared to 161 (9%) of the patients without generator replacement. The type of leads included 38% atrial leads, 26% right ventricle leads, 13% left ventricle leads, and 23% high voltage leads. During the first 3 years of follow-up, 71 (8%) of the patients with a replaced generator had lead re-intervention compared to 191 (5%) of the patients without a replaced generator. The HR of generator replacement was 2.00 (95%CI:1.51;2.64, $p < 0.001$) and 1.60 (95%CI:1.21;2.12, $p = 0.001$) balancing generator implant time. The HR of lead re-intervention was 1.62 (95%CI:1.21;2.16, $p = 0.001$).

Conclusion: Wearing a contemporary cardiovascular implantable electronic device with a prior generator replacement is associated with a two-fold increased risk of additional generator replacement and an increased risk of lead re-intervention following external cardioversion.

10.4.3 - Rhythm Control, Cardioversion

Figure 1

