

Efficient Contactless Heartbeat Rate Measurement for Health Monitoring

Haque, Mohammad Ahsanul; Nasrollahi, Kamal; Moeslund, Thomas B.

Published in:
International Journal of Integrated Care

DOI (link to publication from Publisher):
[10.5334/ijic.2266](https://doi.org/10.5334/ijic.2266)

Publication date:
2015

[Link to publication from Aalborg University](#)

Citation for published version (APA):
Haque, M. A., Nasrollahi, K., & Moeslund, T. B. (2015). Efficient Contactless Heartbeat Rate Measurement for Health Monitoring. *International Journal of Integrated Care*, 15. <https://doi.org/10.5334/ijic.2266>

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal -

Take down policy

If you believe that this document breaches copyright please contact us at vbn@aub.aau.dk providing details, and we will remove access to the work immediately and investigate your claim.

Efficient contactless heartbeat rate measurement for health monitoring

Mohammad A. Haque, Kamal Nasrollahi and Thomas B. Moeslund, Aalborg University, Aalborg, Denmark

Correspondence to: Mohammad A. Haque, T +45-9940-2663, mah@create.aau.dk

Motivations:

Heartbeat Rate (HR) is an important physiological parameter required in health monitoring. Recent methods use videos captured by simple webcams to track subtle periodic changes of facial skin color for contactless HR measurement. These methods denoise the raw Heartbeat Signal (HS) and then transform the signal to frequency domain to measure the HR from the highest frequency component. However, simple denoising does not provide visible heartbeat peaks for analysis, and the highest frequency component does not always represent heartbeat. Thus, we propose a novel way for HS processing, which solves the aforementioned problems and provides a precise measurement of HR for health monitoring.

Method:

We, first, extract the HS from facial video by tracing the facial color changes. As heartbeats are cyclic, the HS is passed through a Hodrick–Prescott filter to reduce trends. The signal is then decomposed into some Intrinsic Mode Functions (IMF) by employing a Complete Ensemble EMD with Adaptive Noise (CEEMDAN) [1]. As the raw HS from facial video is composed of heartbeat and noise components, the decomposition generates an IMF that keeps the heartbeat peak in an uncontaminated form. We empirically select that IMF and employ a peak detection (local maxima) method to detect the number of heartbeat peaks in a time interval. The number of peaks provides us with the measurement of HR.

Results and discussions:

We tested a Matlab implementation of the system on the well-known MAHNOB-HCI database [2]. The database has 491 videos, each video longer than 30 seconds which are collected from 23 subjects in realistic scenarios. The proposed method obtained the Root Mean Square Error (RMSE) 8.54 bpm and Mean Error Rate (MER) 11.96% while comparing to the ground truth ECG. In addition, we compared the results of the proposed method with the state-of-the-art-methods proposed in [3]–[5]. The proposed approach outperformed these methods both by RMSE 5.06, 16.56 and 12.46 bpm and MER 1.24%, 11.30% and 8.11%, respectively. The results imply that the proposed method can provide a highly accurate (around 90%) system for HR measurement and can be effectively used for health care and monitoring application.

Keywords:

heartbeat rate, facial video, health monitoring, empirical mode decomposition

References:

- [1] M. E. Torres, M. A. Colominas, G. Schlotthauer, and P. Flandrin, “A complete ensemble empirical mode decomposition with adaptive noise,” in 2011 IEEE International Conference on Acoustics, Speech and Signal Processing (ICASSP), 2011, pp. 4144–4147.
- [2] X. Li, J. Chen, G. Zhao, and M. Pietikainen, “Remote Heart Rate Measurement From Face Videos Under Realistic Situations,” in IEEE Conference on Computer Vision and Pattern Recognition (CVPR), 2014, pp. 4321–4328.
- [3] M.-Z. Poh, D. J. McDuff, and R. W. Picard, “Advancements in Noncontact, Multiparameter Physiological Measurements Using a Webcam,” IEEE Trans. Biomed. Eng., vol. 58, no. 1, pp. 7–11, Jan. 2011.
- [4] S. Kwon, H. Kim, and K. S. Park, “Validation of heart rate extraction using video imaging on a built-in camera system of a smartphone,” in 2012 Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC), 2012, pp. 2174–2177.
- [5] G. Balakrishnan, F. Durand, and J. Guttag, “Detecting Pulse from Head Motions in Video,” in IEEE Conference on Computer Vision and Pattern Recognition (CVPR), 2013, pp. 3430–3437.