

# **Aalborg Universitet**

# Real-Time Interferometric Refractive Index Change Measurement for the Direct **Detection of Enzymatic Reactions and the Determination of Enzyme Kinetics**

Jepsen, Søren T: Jørgensen, Thomas M: Sørensen, Henrik S: Kristensen, Søren R

Published in: Sensors

DOI (link to publication from Publisher): 10.3390/s19030539

Creative Commons License CC BY 4.0

Publication date: 2019

Document Version Publisher's PDF, also known as Version of record

Link to publication from Aalborg University

Citation for published version (APA):

Jepsen, S. T., Jørgensen, T. M., Sørensen, H. S., & Kristensen, S. R. (2019). Real-Time Interferometric Refractive Index Change Measurement for the Direct Detection of Enzymatic Reactions and the Determination of Enzyme Kinetics. Sensors, 19(3), Article 539. https://doi.org/10.3390/s19030539

#### **General rights**

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
   You may not further distribute the material or use it for any profit-making activity or commercial gain
   You may freely distribute the URL identifying the publication in the public portal -

If you believe that this document breaches copyright please contact us at vbn@aub.aau.dk providing details, and we will remove access to the work immediately and investigate your claim.

Downloaded from vbn.aau.dk on: December 05, 2025





Article

# Real-Time Interferometric Refractive Index Change Measurement for the Direct Detection of Enzymatic Reactions and the Determination of Enzyme Kinetics

Søren T. Jepsen <sup>1,2,\*</sup>, Thomas M. Jørgensen <sup>3</sup>, Henrik S. Sørensen <sup>4</sup> and Søren R. Kristensen <sup>1,2</sup>

- Department of Clinical Biochemistry, Aalborg University Hospital, Sdr. Skovvej, DK-9000 Aalborg, Denmark; srk@rn.dk
- Department of Clinical Medicine, Aalborg University, Sdr. Skovvej, DK-9000 Aalborg, Denmark
- Department of Applied Mathematics and Computer Science, Technical University of Denmark, DK-2800 Kongens Lyngby, Denmark; tmjq@dtu.dk
- Department of Photonics Engineering, Technical University of Denmark, DK-2800 Kongens Lyngby, Denmark; biz4doc@gmail.com
- \* Correspondence: sotj@rn.dk

Received: 30 December 2018; Accepted: 22 January 2019; Published: 28 January 2019



Abstract: Back scatter interferometry (BSI) is a sensitive method for detecting changes in the bulk refractive index of a solution in a microfluidic system. Here we demonstrate that BSI can be used to directly detect enzymatic reactions and, for the first time, derive kinetic parameters. While many methods in biomedical assays rely on detectable biproducts to produce a signal, direct detection is possible if the substrate or the product exert distinct differences in their specific refractive index so that the total refractive index changes during the enzymatic reaction. In this study, both the conversion of glucose to glucose-6-phosphate, catalyzed by hexokinase, and the conversion of adenosine-triphosphate to adenosine di-phosphate and mono-phosphate, catalyzed by apyrase, were monitored by BSI. When adding hexokinase to glucose solutions containing adenosine-triphosphate, the conversion can be directly followed by BSI, which shows the increasing refractive index and a final plateau corresponding to the particular concentration. From the initial reaction velocities,  $K_M$  was found to be 0.33 mM using Michaelis–Menten kinetics. The experiments with apyrase indicate that the refractive index also depends on the presence of various ions that must be taken into account when using this technique. This study clearly demonstrates that measuring changes in the refractive index can be used for the direct determination of substrate concentrations and enzyme kinetics.

**Keywords:** refractive index; interferometry; enzyme; *dn/dc* 

#### 1. Introduction

The detection and quantification of enzymatic reactions is widely used in many areas of industrial production and biomedical assays. Consequently, there is also a wide range of well-established methods and sensors for measuring such reactions, including but far from limited to colorometric, polariometric and amperometric sensors [1]. Many of these sensors rely on a coupled reaction scheme with secondary reactants that form a detectable biproduct, e.g., NAD+/NADH, or enable a transfer of electrons to electrodes to produce a detectable signal. The implementation of such sensors into miniaturized microfluidic systems is therefore challenged by the need to include additional reactants and to ensure the precise mixing of sample and reactant. Furthermore, such coupled reactions may interfere with the reaction rates if secondary products either inhibit the enzyme or have a lower

Sensors **2019**, 19, 539

catalytic rate of formation than the primary product, making them less desirable when one wants to study the kinetic parameters of the enzyme itself.

Direct detection can be possible if either the substrate or the product exert distinct physiochemical properties such as chirality, intrinsic fluorescence or in this case, a different refractive index (RI). The refractive index of a dense medium is a macroscopic quantity that is related to the microscopic polarizability through the Lorentz–Lorentz equation:

$$\frac{n^2 - 1}{n^2 + 1} = \frac{4\pi}{3} N\alpha \tag{1}$$

where n is the refractive index, N is the number density and  $\alpha$  is the polarizability.

Because of the non-invasive nature and ease of miniaturization, RI-based sensors are often used to detect protein and other biomolecules of interest in the emerging area of nano- and microfluidics [2,3] and the possibility of expanding the use of refractive index sensors to also include enzymatic reactions poses an interesting possibility [4–7].

Previously, we briefly demonstrated that the direct detection of the enzymatic phosphorylation of glucose is possible by measuring the change in the refractive index of a solution, using both an interferometric setup and a conventional differential refractometer [8]. Enzymatic reactions have previously been characterized from RI changes using mass sensitive surface plasmon resonance sensors, however, the enzymes were bound to a solid surface and the change in RI occurred on localized surfaces as a result of mass density changes [9,10]. Likewise, refractometry has been used to study the enzymatic degradation of small micelle-like particles [11]. To the best of our knowledge the characterization of small molecule enzymatic kinetics from RI changes of a bulk fluid has not previously been studied.

In this study, we pursued the use of a refractive index sensor based on so-called back scatter interferometry (BSI), to directly detect and quantify enzymatic reactions in a microfluidic environment. Studies have shown that BSI can detect low concentrations of both ions and saccharides in capillary electrophoresis [12–14] and an early study even used a polariometric-BSI configuration to detect an enzymatic reaction with hydroxybutyrate dehydrogenase [15]. There have also been multiple studies on the molecular interactions between protein-sized molecules using BSI [16], but the effect of such interactions on the bulk refractive index and origin of the observed signal has been debated [12,17–19].

In the present paper, we limited our study to two enzymatic reactions: The first reaction was the phosphorylation of glucose to glucose-6-phosphate (G-6-P) with adenosine-triphosphate (ATP) as a bisubstrate, catalyzed by the enzyme hexokinase, which we have previously shown [8]. In this study, we examined if the observed change in RI during the initial reaction rate was suitable for determining the kinetic Michaelis–Menten constant ( $K_M$ ). Secondly, given that the former reaction involves multiple products, we sought to differentiate the refractive index contributions of the individual components by studying the enzymatic hydrolysis of adenosine-triphosphate to adenosine di-phosphate and mono-phosphate (ATP, ADP and AMP, respectively) catalyzed by the enzyme apyrase.

Our detection principle was based on the assumption that there is a measurable difference in the refractive index between the solute substrates and products formed in the enzymatic reaction. The observed change in the RI of a solution as a function of solute concentration is termed the refractive index increment (dn/dc). If the dn/dc value of the product is different from that of the substrate, one would expect a change in the bulk RI as the reaction proceeds, proportional to the difference in the dn/dc value between the substrate and the product. We determined the dn/dc values of ATP and ADP from a series of standard dilutions and discuss these values with the observed RI changes from the enzymatic reactions. The dn/dc values for glucose and G-6-P have previously been determined using BSI [8].

Sensors **2019**, 19, 539

#### 2. Materials and Methods

The back scatter interferometry setup was identical to that used in our previous work [8,19]. Briefly, the beam from a HeNe laser impinges a glass capillary with inner and outer diameters of 1.4 and 1.9 mm respectively, which contains the sample of interest. The probe volume is in itself quite small, less than 5  $\mu$ L, but the total sample volume injected must be sufficient to ensure full fluid displacement. As light from the laser traverses through the sample liquid, it obtains a phase shift proportional to the RI of the liquid, and as light is reflected from both the inner and outer capillary surfaces, an interference pattern consisting of high contrast periodic fringes is formed. This fringe pattern is recorded with a charge-coupled device and a Fourier transformation is applied to the image. Changes to the sample RI produce a spatial translation of the fringes that is quantified by obtaining the phase of the Fourier transformed image in units of radians (rad).

When observing the directly backwards-reflected fringes the sensitivity of the interferometer  $(d\phi/dRI)$ , i.e., the phase response per change in sample RI, is given by the following relationship [20]:

$$d\phi/dRI = \left(\frac{2\pi}{\lambda}\right) 2nd\tag{2}$$

where  $\lambda$  is the laser wavelength (632.8 nm), n is the refractive index and d is the inner diameter of the capillary (1.4 mm), multiplied by 2 because the backwards reflected light passes the sample twice. In the current setup, this yields a theoretical sensitivity, where  $d\phi/dRI = 27,802$  rad/RIU. Conversion from radians into refractive index units (RIU) is achieved by the division of the sensitivity. The limit of detection and limiting factors thereof was examined in our previous work and was found to be  $7 \times 10^{-7}$  RIU for the current system [19].

The refractive index of a fluid is largely affected by the temperature  $(10^{-4})^{\circ}$ C) and stable temperature control over time is essential for accurately detecting small RI changes during the course of a measured reaction. The current temperature control of the system does not produce baseline deviations greater than  $10^{-7}$  over a period of hours [19]. All measurements were performed at 25 °C.

### 2.1. Reagents

D-Glucose anhydrous of high purity (>99%) was obtained from Merck (Darmstadt, Germany). Glucose-6-phosphate (G7772), hexokinase from *Saccharomyces cerevisiae* (H4502), apyrase from *Solanum tuberosum*, (A6535), adenosine 5'-triphosphate disodium salt hydrate (A2383), and adenosine 5'-diphosphate sodium salt (A2754) were all acquired from Sigma-Aldrich (St. Louis, MO, USA).

#### 2.2. Determination of the dn/dc Value

The dn/dc value was determined using weighed solutions or ATP and ADP prepared in a MES buffer (identical to that used in the enzymatic reactions with apyrase, see below). The dn/dc value was found by linear regression of the measured signal vs. the concentration.

#### 2.3. Measurement of Enzymatic Reactions

Substrate and enzyme mixtures were rapidly mixed by pipetting and 400  $\mu$ L was injected into the BSI system using a peristaltic pump (BT100-2J, Longer Precision Pump Co., Ltd., Hebei, China). The total mixing and injection time was 40 s, in order to ensure full fluid displacement inside the capillary. The hexokinase measurements were performed using a 1 nM hexokinase solution in a 0.05 M tris buffer, pH 8.0, with 2 mM ATP and 15 mM MgCl<sub>2</sub>. The initial reaction velocity as a function of the substrate concentration ( $V_{[S]}$ ) in units of rad/min was measured over a range of glucose substrate concentrations [0.1–2.0 mM] and the Michaelis–Menten constant ( $K_M$ ) was derived by non-linear parameter estimation using the equation:

$$V_{[S]} = V_{max}([S]/[S] + K_M)$$
(3)

Sensors **2019**, 19, 539 4 of 8

where [S] is the glucose concentration,  $V_{max}$  is the highest reaction velocity obtained, and  $V_{[S]}$  and  $K_M$  are as explained above.

Apyrase measurements were performed using a 5 nM apyrase solution in a 0.1 M MES buffer with pH 6.0, containing 5 mM  $CaCl_2$ .

#### 3. Results

#### 3.1. The dn/dc Value

All standard curves showed a linear relationship ( $R^2 > 0.99$ ) between signal and concentration. As reported previously, the dn/dc values were found to be 0.145 mL/g and 0.156 RIU/(g/mL) for glucose and for G-6-P, respectively [8]. The value for glucose was in good agreement with other studies reporting values for glucose, ranging from 0.142 to 0.145 RIU/(g/mL), depending on temperature and wavelength [21,22]. The dn/dc values for ATP and ADP were found to be 0.144 RIU/(g/mL) and 0.163 RIU/(g/mL), respectively. The results are summarized, including the molar values, in Table 1. Based on these findings, an *a priori* assumption was that the enzymatic phosphorylation of glucose produces an increase in RI, whereas the conversion from ATP to ADP should produce a decrease in RI. Thus, the total signal from the enzymatic phosphorylation of 1 millimole of glucose, including the equimolar bisubstrate conversion of ATP to ADP, should yield a total change in RI equal to  $(4.07-2.61) + (6.95-7.95) = 0.45 \times 10^{-5}$  RIU.


**Table 1.** Summarized findings of the dn/dc value and the slope of the endpoint values vs. substrate concentration. n is the number of solutions in each standard curve and substrate solutions in the enzymatic reactions.

Substance	n	$dn/dc$ (RIU/mM) $ imes 10^{-5}$	$dn/dc \text{ RIU/(g/mL)}^{-1} \times 10^{-5}$
Glucose	9	2.61	0.145
G-6-P	9	4.07	0.156
ATP	7	7.95	0.144
ADP	7	6.95	0.163
Enzymatic Reaction		Slope of Endpoint (RIU/mM) $\times$ 10 <sup>-5</sup>	
Glu→G-6-P (ATP→ADP)	4	0.99	
$\overrightarrow{ATP} \rightarrow AMP + 2P_i$	4	1.51	
$ADP \rightarrow AMP + P_i$	4	0.68	

#### 3.2. Enzymatic Phosphoryalation of Glucose by Hexokinase

The measured signal from the enzymatic phosphorylation of glucose catalyzed by hexokinase, as seen in Figure 1a, displayed an initial linear increase in the signal, and the slope of this initial linear period was identical for the higher concentrations within the first minute, indicating zero order kinetics (substrate concentrations  $>> K_M$ ). Eventually, as the glucose substrate was catalyzed to G-6-P, the signal rate decreased until a maximum plateau (end-point) was reached and the reaction was complete, in agreement with the general Michaelis–Menten kinetics. The measurements were started after 40 s of injection time and, therefore, we lack the signal generated in this time period. Because the reaction proceeded at maximum velocity, this initial lacking signal is linear, and is essentially a constant. Using backwards linear extrapolation on the data, as shown in Figure 1a, we found that this lacking signal was relatively small, equaling 0.02 and 0.03 radians for the smallest and the highest substrate concentrations, respectively. The relationship between the end-point signal and initial glucose concentration was linear, with a slope of 0.99  $\times$  10<sup>-5</sup> RIU/mM, as seen in Figure 1b.


Sensors **2019**, 19, 539 5 of 8



**Figure 1.** The enzymatic phosphorylation of glucose to glucose-6-phosphate by hexokinase, measured using BSI. (a) The real-time measurement of four glucose solutions (0.4, 1.0, 1.6 and 2.0 mM). (b) The end-point values in refractive index units vs. initial glucose concentration. Linear fit ( $R^2 > 0.99$ , slope  $0.99 \times 10^{-5}$  RIU/mM). Figure 1a is reproduced from reference [8] with permission from Elsevier.

# 3.3. K<sub>M</sub> Hexokinase

To determine the  $K_M$  for hexokinase, the initial velocities from the experiments with glucose concentrations ranging from 0.1 to 2.0 mM were plotted against the substrate concentrations (Figure 2) and  $K_M$  was found to be 0.33 mM (95% CI: 0.19–0.47).




**Figure 2.** Hexokinase substrate saturation curve: Initial reaction velocity (rad/min) plotted against initial glucose concentrations (mM). Red line shows non-linear regression ( $R^2 = 0.93$ ) to the parameters in Equation 3 from which  $K_M$  is obtained.

## 3.4. Enzymatic Hydrolosis of ATP and ADP by Apyrase

The enzymatic hydrolysis of ATP and ADP by apyrase, as seen in the examples shown in Figure 3a, was in accordance with enzymatic reactions, showing an initial linear increase where the reaction proceeded at  $V_{max}$ , and a final plateau with endpoint values that showed a linear relationship when plotted against the initial substrate concentration, as seen in Figure 3b. The estimated signal that was lost during injection (the initial 40 s) was less than 0.01 rad and was disregarded. Surprisingly, this reaction showed a marked increase in the RI, which was contrary to the expected decrease based on the dn/dc values of ATP and ADP. From the linear regression, the  $\Delta$ RI was found to be 1.51  $\times$  10<sup>-5</sup> for ATP and 0.68  $\times$  10<sup>-5</sup> for ADP. As is discussed below, the reason for this increase in RI could be attributed to the release of free inorganic phosphate (Pi) from the enzymatic hydrolysis of ATP and ADP.

Sensors **2019**, 19, 539 6 of 8



**Figure 3.** The enzymatic hydrolysis of ATP and ADP to AMP by apyrase. (a) The real-time measurement of 1.6 mM ATP and ADP (duplicate runs are shown for both ATP and ADP). (b) The linear fit of end-point values in refractive index units vs. initial substrate concentration. ATP (blue)  $R^2 > 0.99$ , slope  $1.51 \times 10^{-5}$  RIU/mM, ADP (green)  $R^2 > 0.99$ , slope  $0.68 \times 10^{-5}$  RIU/mM.

#### 4. Discussion

Our results from the enzymatic reactions showed changes of the RI in the order of  $10^{-5}$ , well above the RI detection limit of our set-up ( $7 \times 10^{-7}$ ) [19] for substrate concentrations in the milimolar range. We also found a linear relationship between the substrate and endpoint signal, as seen in Figure 1b, which implies that it would be possible to determine unknown concentrations of glucose from enzymatic reactions using BSI.

The  $K_M$  for hexokinase was found to be 0.33 mM (95% CI: 0.19–0.47), which was in reasonable accordance with the published values of 0.13–0.24 mM [2,4], with variations attributed to experimental conditions such as ionic strength. The relatively large uncertainty in  $K_M$  was caused by the limitations of the current experimental setup. For low substrate concentrations ([S]< $K_M$ ), a significant amount of substrate will be catalyzed during injection (40 s) and the enzyme does not operate at  $V_{max}$ , which will effectively cause a slight overestimation of  $K_M$ . However, our results demonstrated that refractive index sensing can be used in the estimation of enzyme kinetics, although the system should be further optimized to handle low substrate concentrations.

### Possible Effect of Ions

The enzymatic hydrolysis of ATP performed in our experiments resulted in a large increase in the refractive index upon the hydrolyzation of ATP, which suggests that the inorganic phosphate (Pi) released by the hydrolysis of the  $\gamma$ - and  $\beta$ -phosphate groups of adenosine phosphates made a significant contribution to the refractive index. The refractive index of a solution containing ions is largely dependent on the polarizability of the ions, and there are several studies on this issue [23–26]. Marcus found the following polarizabilities for ions in solution:  $Ca^{2+}$  0.63 and  $PO_4^{3-}$  $5.99 \times 10^{-24} \text{ cm}^3$  [24]. Phosphate especially stands out, with a relatively large polarizability, which supports our findings that the release of P<sub>i</sub> made a significant contribution to the RI of a solution. Although equimolar amounts of AMP are produced from both ATP and ADP as a substrate, two phosphate groups ( $\gamma$ - and  $\beta$ -phosphate) are released from ATP, but only one ( $\beta$ -phosphate) is released from ADP. However, the end-point values for ATP hydrolysis are more than twice as high as those for ADP (1.51  $\times$  10<sup>-5</sup> vs. 0.68  $\times$  10<sup>-5</sup>), which suggests that the hydrolysis of the  $\gamma$ -phosphate group produces a larger increment in the RI than the hydrolysis of the β-phosphate. This difference in the RI contribution from the phosphate groups could be caused by an inequivalent dissociation of divalent ions chelated with adenosine phosphates. The Ca2+ and Mg2+ stability constants for the complexes with ATP, ADP, AMP were  $log(K_{Mg})$  4.0, 3.0 and 1.6; and  $log(K_{Ca})$  3.5, 2.9 and 1.5, Sensors **2019**, 19, 539 7 of 8

respectively [27–29]. Thus, a larger amount of cations were dissociated upon  $\beta$ -hydrolysis compared to  $\gamma$ -hydrolysis, which may explain the disproportionate change in refractive index between  $\beta$ - and  $\gamma$ -hydrolysis.

As ATP and ADP are only commercially available as salts, the standard solutions used to determine the dn/dc value contained dissociated ions that may affect the refractive index significantly and the experimental determined dn/dc values for ATP and ADP needs further validation. The theoretical polarizabilities of the adenine phosphates are ATP 35.6, ADP 31.7 and AMP 27.7 ( $\times 10^{-24}$  cm<sup>3</sup>) [30], which is in accordance with our results that showed that the dn/dc value for ATP was larger than for ADP.

In the case of the enzymatic phosphorylation of glucose by hexokinase, we also found a discrepancy between the increase in the RI measured from the enzymatic reaction (end-point value:  $0.99 \times 10^{-5}$  RIU/mM) and the increase estimated from the dn/dc value ( $0.45 \times 10^{-5}$  RIU/mM), which further suggests that the dn/dc value of substrates and of products cannot accurately predict the RI change from enzymatic reactions.

#### 5. Conclusions

We demonstrated that the measurement of RI changes can be used to directly measure selected enzymatic reactions. Specifically, this was achieved using BSI with small sample volumes contained in a simple glass capillary. The change in RI was proportional to the formation of the products and could be used to determine the substrate concentrations and to obtain enzymatic kinetic parameters, but the effect of the various components in the solution on the bulk RI was complex and requires further study.

**Author Contributions:** Conceptualization and methodology, S.T.J., T.M.J., H.S.S. and S.R.K.; Supervision, review and editing, S.R.K., H.S.S. and T.M.J.; Writing—Original draft preparation, S.T.J. Funding acquisition, S.R.K.

Funding: This research received no external funding.

**Conflicts of Interest:** The authors declare no conflict of interest.

#### References

- 1. Rahman, M.M.; Ahammad, A.J.S.; Jin, J.H.; Ahn, S.J.; Lee, J.J. A comprehensive review of glucose biosensors based on nanostructured metal-oxides. *Sensors* **2010**, *10*, 4855–4886. [CrossRef]
- 2. Cooper, M.A. Optical biosensors in drug discovery. Nat. Rev. Drug Discov. 2002, 1, 515–528. [CrossRef]
- 3. Marazuela, M.; Moreno-Bondi, M. Fiber-optic biosensors—An overview. *Anal. Bioanal. Chem.* **2002**, 372, 664–682. [CrossRef]
- 4. Nemati, M.; Santos, A.; Kumeria, T.; Losic, D. Label-Free Real-Time Quantification of Enzyme Levels by Interferometric Spectroscopy Combined with Gelatin-Modified Nanoporous Anodic Alumina Photonic Films. *Anal. Chem.* **2015**, *87*, 9016–9024. [CrossRef]
- 5. Sun, J.; Chan, C.C.; Zhang, Y.F.; Zheng, L.T.; Ho, H.L.; Dong, X.Y.; Shao, L.Y. Glucose optical biosensor with sol-gel-coated long-period gratings. *Int. Soc. Opt. Photonics* **2008**, *7134*, 713443. [CrossRef]
- Fang, S.; Lee, H.J.; Wark, A.W.; Kim, H.M.; Corn, R.M. Determination of Ribonuclease H Surface Enzyme Kinetics by Surface Plasmon Resonance Imaging and Surface Plasmon Fluorescence Spectroscopy. *Anal. Chem.* 2005, 77, 6528–6534. [CrossRef]
- 7. Markov, D.; Begari, D.; Bornhop, D.J. Breaking the  $10^{-7}$  barrier for RI measurements in nanoliter volumes. *Anal. Chem.* **2002**, 74, 5438–5441. [CrossRef]
- 8. Jørgensen, T.M.; Jepsen, S.T.; Sørensen, H.S.; di Gennaro, A.K.; Kristensen, S.R. Back scattering interferometry revisited—A theoretical and experimental investigation. *Sens. Actuators B Chem.* **2015**, 220, 1328–1337. [CrossRef]
- 9. Iwasaki, Y.; Horiuchi, T.; Niwa, O. Detection of Electrochemical Enzymatic Reactions by Surface Plasmon Resonance Measurement. *Anal. Chem.* **2001**, *73*, 1595–1598. [CrossRef]

Sensors **2019**, 19, 539

 Wegner, G.J.; Wark, A.W.; Lee, H.J.; Codner, E.; Saeki, T.; Fang, S.; Corn, R.M. Real-Time Surface Plasmon Resonance Imaging Measurements for the Multiplexed Determination of Protein Adsorption/Desorption Kinetics and Surface Enzymatic Reactions on Peptide Microarrays. *Anal. Chem.* 2004, 76, 5677–5684.
 [CrossRef]

- 11. Lam, H.; Gong, X.; Wu, C. Novel Differential Refractometry Study of the Enzymatic Degradation Kinetics of Poly(ethylene oxide)-*b*-poly(ε-caprolactone) Particles Dispersed in Water. *J. Phys. Chem. B* **2007**, 111, 1531–1535. [CrossRef]
- 12. Bornhop, D.J.; Kammer, M.N.; Kussrow, A.; Flowers, R.A.; Meiler, J. Origin and prediction of free-solution interaction studies performed label-free. *Proc. Natl. Acad. Sci. USA* **2016**, *113*, E1595–E1604. [CrossRef]
- 13. Swinney, K.; Markov, D.; Bornhop, D.J. Ultrasmall volume refractive index detection using microinterferometry. *Rev. Sci. Instrum.* **2000**, *71*, 2684. [CrossRef]
- 14. Dunn, R.C. Wavelength Modulated Back-Scatter Interferometry for Universal, On-Column Refractive Index Detection in Picoliter Volumes. *Anal. Chem.* **2018**, *90*, *6789–6795*. [CrossRef]
- Swinney, K.; Bornhop, D.J. D-β-Hydroxybutyrate Reaction Kinetics Studied in Nanoliter Volumes Using a Capillary Polarimeter. Appl. Spectrosc. 2000, 54, 2–6. [CrossRef]
- 16. Bornhop, D.J.; Latham, J.C.; Kussrow, A.; Markov, D.A.; Jones, R.D.; Sørensen, H.S. Free-solution, label-free molecular interactions studied by back-scattering interferometry. *Science* **2007**, *317*, 1732–1736. [CrossRef]
- 17. Varma, M.M. Article does not explain the origin of free-solution protein interaction signals. *Proc. Natl. Acad. Sci. USA* **2016**, *113*, E4930. [CrossRef]
- 18. Bornhop, D.J.; Kammer, M.N.; Kussrow, A.; Flowers, R.A. Reply to Varma: Elucidation of the signal origin for label-free, free-solution interactions. *Proc. Natl. Acad. Sci. USA* **2016**, *113*, E4931–E4932. [CrossRef]
- 19. Jepsen, S.T.; Jørgensen, T.M.; Zong, W.; Trydal, T.; Kristensen, S.R.; Sørensen, H.S. Evaluation of back scatter interferometry, a method for detecting protein binding in solution. *Analyst* **2014**, *140*, 895–901. [CrossRef]
- 20. Becker, A.; Köhler, W.; Müller, B. A Scanning Michelson Interferometer for the Measurement of the Concentration and Temperature Derivative of the Refractive Index of Liquids. *Ber. Bunsen-Ges. Phys. Chem.* **1995**, *99*, 600–608. [CrossRef]
- 21. Larin, K.V.; Akkin, T.; Esenaliev, R.O.; Motamedi, M.; Milner, T.E. Phase-sensitive optical low-coherence reflectometry for the detection of analyte concentrations. *Appl. Opt.* **2004**, *43*, 3408–3414. [CrossRef]
- 22. Tumolo, T.; Angnes, L.; Baptista, M.S. Determination of the refractive index increment (dn/dc) of molecule and macromolecule solutions by surface plasmon resonance. *Anal. Biochem.* **2004**, *333*, 273–279. [CrossRef]
- Berlind, T.; Pribil, G.K.; Thompson, D.; Woollam, J.A.; Arwin, H. Effects of ion concentration on refractive indices of fluids measured by the minimum deviation technique. *Phys. Status Solidi* 2008, 5, 1249–1252. [CrossRef]
- 24. Marcus, Y. Ionic Liquid Properties; Springer: Berlin/Heidelberg, Germany, 2016.
- 25. Pacák, P. Molar refractivity and interactions in solutions. 1. Molar refractivity of some monovalent ions in aqueous and dimethyl sulfoxide solutions. *Chem. Pap.* **1989**, 43, 489–500.
- 26. Bauer, N.; Fajans, K. The Molar Dispersion and Refraction of Free and Bonded Ions<sup>1</sup>. *J. Am. Chem. Soc.* **1942**, 64, 3023–3034. [CrossRef]
- 27. Heyde, M.E.; Rimai, L. A Raman Spectroscopic Study of the Interaction of Ca<sup>2+</sup> and Mg<sup>2+</sup> with the Triphosphate Moiety of Adenosine Triphosphate in Aquas Solution. *Biochemistry* **1971**, *10*, 1121–1128.
- 28. Alberty, R.A. Effect of pH and metal ion concentration on the equilibrium hydrolysis of adenosine triphosphate to adenosine diphosphate. *J. Biol. Chem.* **1968**, 243, 1337–1343.
- 29. Wilson, J.; Chin, A. Chelation of divalent cations by ATP, studied by titration calorimetry. *Anal. Biochem.* **1991**, *19*, 16–19. [CrossRef]
- 30. Chemspider. Available online: http://www.chemspider.com/Chemical-Structure.5742.html (accessed on 1 November 2018).



© 2019 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (http://creativecommons.org/licenses/by/4.0/).