



Aalborg Universitet

AALBORG UNIVERSITY
DENMARK

Improvement in image quality of Tc-99m-based ventilation/perfusion single-photon emission computed tomography in patients with chronic obstructive pulmonary disease through pretest continuous positive airway pressure treatment

Paludan, Jens Peder Dreyer; Andresen, Sven Robert; Abrahamsen, Jan; Petersen, Lars Jelstrup; Høyer, Christian

Published in:
World Journal of Nuclear Medicine

DOI (link to publication from Publisher):
[10.4103/wjnm.WJNM_81_18](https://doi.org/10.4103/wjnm.WJNM_81_18)

Creative Commons License
CC BY-NC-SA 4.0

Publication date:
2019

Document Version
Publisher's PDF, also known as Version of record

[Link to publication from Aalborg University](#)

Citation for published version (APA):

Paludan, J. P. D., Andresen, S. R., Abrahamsen, J., Petersen, L. J., & Høyer, C. (2019). Improvement in image quality of Tc-99m-based ventilation/perfusion single-photon emission computed tomography in patients with chronic obstructive pulmonary disease through pretest continuous positive airway pressure treatment. *World Journal of Nuclear Medicine*, 18(2), 185-186. https://doi.org/10.4103/wjnm.WJNM_81_18

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal -

Take down policy

If you believe that this document breaches copyright please contact us at vbn@aub.aau.dk providing details, and we will remove access to the work immediately and investigate your claim.

Downloaded from vbn.aau.dk on: February 08, 2025

Case Report

Improvement in image quality of Tc-99m-based ventilation/perfusion single-photon emission computed tomography in patients with chronic obstructive pulmonary disease through pretest continuous positive airway pressure treatment

ABSTRACT

Ventilation/perfusion single-photon emission computed tomography performed using an aerosol of carbon-coated technetium is frequently used for diagnosing pulmonary embolism. Certain patients may suffer from chronic obstructive pulmonary disease (COPD); for such patients, the formation of mucus clots in airways can cause accumulation of the aerosol in the larger airways. This centralized deposition of the aerosol leads to insufficient activity in peripheral lung segments and subsequently results in ventilation images of substandard or even nondiagnostic quality. Continuous positive airway pressure (CPAP) therapy improves airway dynamics and quality of life for COPD patients. We report for the first time the results for two patients for whom initial ventilation scans were of insufficient quality, but diagnostic-quality images were obtained after CPAP therapy.

Keywords: Chronic obstructive pulmonary disease (MeSH), continuous positive airway pressure ventilation (MeSH), emission-computed, pulmonary embolism/diagnostic imaging (MeSH), single-photon (MeSH), tomography

INTRODUCTION AND CASE REPORT

Pulmonary embolism is a potentially life-threatening and common condition with a clinical symptomatology that makes it difficult to distinguish from other pulmonary or cardiac diseases. Imaging techniques such as ventilation/perfusion single-photon emission computed tomography/computed tomography (SPECT/CT) are therefore pivotal in the diagnosis of this condition.^[1,2] In patients with severe chronic obstructive pulmonary disease (COPD), the image quality produced by SPECT is often of substandard or even nondiagnostic quality. These patients exhibit increased mucus production that results in accumulation of the radioactive tracer in “hot spots” when ventilation imaging is performed through inhalation of a carbon-coated technetium aerosol.^[3,4] Continuous positive airway pressure (CPAP) physiotherapy is frequently used

for COPD patients to assist with mucus mobilization.^[5] We present two cases [Figures 1 and 2] in which pretest CPAP reduced deposition of the tracer in central airways, thereby improving imaging quality. This approach could particularly benefit patients for whom other imaging modalities, such

**JENS PEDER DREYER PALUDAN,
SVEN ROBERT ANDRESEN, JAN ABRAHAMSEN,
LARS JELSTRUP PETERSEN¹, CHRISTIAN HØYER**


Department of Clinical Physiology, Viborg Regional Hospital, Viborg, ¹Department of Nuclear Medicine, Aalborg University Hospital, Aalborg, Denmark

Address for correspondence: Jens Peder Dreyer Paludan, Department of Clinical Physiology, Viborg Regional Hospital, Heibergs Allé 4, DK-8800 Viborg, Denmark.
E-mail: jens.paludan@midt.rm.dk

This is an open access journal, and articles are distributed under the terms of the Creative Commons Attribution-NonCommercial-ShareAlike 4.0 License, which allows others to remix, tweak, and build upon the work non-commercially, as long as appropriate credit is given and the new creations are licensed under the identical terms.

For reprints contact: reprints@medknow.com

How to cite this article: Paludan JP, Andresen SR, Abrahamson J, Petersen LJ, Høyer C. Improvement in image quality of Tc-99m-based ventilation/perfusion single-photon emission computed tomography in patients with chronic obstructive pulmonary disease through pretest continuous positive airway pressure treatment. *World J Nucl Med* 2019;18:185-6.

Access this article online	
Website: www.wjnm.org	Quick Response Code 
DOI: 10.4103/wjnm.WJNM_81_18	




Figure 1: An 84-year-old man with severe chronic obstructive pulmonary disease with centrilobular emphysema, aortic valve stenosis, and pulmonary hypertension was admitted for exacerbation. Infection was excluded, and pulmonary embolism was suspected. Ventilation/perfusion single-photon emission computed tomography was performed after the patient had inhaled 40 MBq/1.1 mCi of Tc-99 m-labelled carbon (Technegas[®], Cyclomedica Ireland, Ltd., Dublin, Ireland) and was intravenously administered 160 MBq/4.3 mCi of Tc-99 m-labelled macroaggregated albumin. The maximum projection images for the ventilation single-photon emission computed tomography (a) showed tracer deposition in the central airways (arrows). The corresponding perfusion imaging (not shown) revealed subsegmental perfusion defects that only partially matched the ventilation scan findings; thus, the test was deemed to be inconclusive. The test was repeated 2 days later after pretest continuous positive airway pressure treatment (10 min of administration, flow of 10 L O₂, resistance of 7.5 cm H₂O). In this second test (b), central tracer accumulation was reduced, and peripheral tracer deposition was increased. The final conclusion was no suspicion of acute pulmonary embolism, in contrast to the equivocal findings from the initial test




Figure 2: A 60-year-old woman with severe chronic obstructive pulmonary disease who was receiving a long-term oxygen therapy was admitted due to increasing dyspnea and fatigue. (a) Blood tests showed no indication of infection, and pulmonary embolism was suspected. Ventilation/perfusion single-photon emission computed tomography was performed using a similar protocol to that described above; the ventilation single-photon emission computed tomography results are presented in Figure 2a. This test showed inhomogeneous tracer distribution, including no tracer deposits in the lower lobes of either lung. The test was ruled to be inconclusive due to the findings of perfusion defects and poor image quality for the ventilation single-photon emission computed tomography. The test was repeated 2 days later after pretest CPAP (b). Overall tracer deposition remained inhomogeneous, but the peripheral deposition of the tracer was improved. Furthermore, fewer lung sections were found to be hypoventilated (arrows), and tracer accumulation in the lower lobes had emerged. All of the perfusion defects were matched to those indicated by the ventilation scan; therefore, there was a reduced probability of pulmonary embolism. These two cases suggest that tracer deposition can be optimized through pretest CPAP, although our findings and the effects of day-to-day variation remain to be explored in standardized trials. Studies have shown that approximately 16% of chronic obstructive pulmonary disease patients with unexplained exacerbation suffer from pulmonary embolism. Optimizing single-photon emission computed tomography by administering pretest CPAP could particularly benefit the subset of chronic obstructive pulmonary disease patients for whom other imaging modalities, such as CT angiography, are not an option due to intolerance to contrast agents. We are currently investigating CPAP in a larger cohort of chronic obstructive pulmonary disease patients with nondiagnostic ventilation scans

as CT angiography, are not an option.^[6] This method has not previously been described in the literature.

Declaration of patient consent

The authors certify that they have obtained all appropriate patient consent forms. In the form the patient(s) has/have given his/her/their consent for his/her/their images and other clinical information to be reported in the journal. The patients understand that their names and initials will not be published and due efforts will be made to conceal their identity, but anonymity cannot be guaranteed.

Financial support and sponsorship

Nil.

Conflicts of interest

There are no conflicts of interest.

REFERENCES

1. Mortensen J, Gutte H. SPECT/CT and pulmonary embolism. *Eur J Nucl Med Mol Imaging* 2014;41 Suppl 1:S81-90.

2. Hess S, Frary EC, Gerke O, Madsen PH. State-of-the-art imaging in pulmonary embolism: Ventilation/perfusion single-photon emission computed tomography versus computed tomography angiography – Controversies, results, and recommendations from a systematic review. *Semin Thromb Hemost* 2016;42:833-45.
3. Wollmer P. Aerosolised radionuclides for functional imaging in COPD/asthma. *Clin Transl Imaging* 2014;2:403-13.
4. Schembri GP, Roach PJ, Bailey DL, Freeman L. Artifacts and anatomical variants affecting ventilation and perfusion lung imaging. *Semin Nucl Med* 2015;45:373-91.
5. Dhand R. Aerosol therapy in patients receiving noninvasive positive pressure ventilation. *J Aerosol Med Pulm Drug Deliv* 2012;25:63-78.
6. Aleva FE, Voets LW, Simons SO, de Mast Q, van der Ven AJ, Heijdra YF. Prevalence and localization of pulmonary embolism in unexplained acute exacerbations of COPD: A systematic review and meta-analysis. *Chest* 2017;151:544-54.