



Correlation Between Valve Event Amplitudes in the Seismocardiogram and VO2-max

Sørensen, Kasper; Poulsen, Mathias Krogh; Karbing, Dan Stieper; Søgaard, Peter; Struijk, Johannes Jan; Schmidt, Samuel Emil

Creative Commons License
Unspecified

Publication date:
2019

Document Version
Accepted author manuscript, peer reviewed version

[Link to publication from Aalborg University](#)

Citation for published version (APA):
Sørensen, K., Poulsen, M. K., Karbing, D. S., Søgaard, P., Struijk, J. J., & Schmidt, S. E. (2019). *Correlation Between Valve Event Amplitudes in the Seismocardiogram and VO2-max*. Abstract from Engineering in Medicine and Biology 2019, Berlin, Germany.

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal -

Take down policy

If you believe that this document breaches copyright please contact us at vbn@aub.aau.dk providing details, and we will remove access to the work immediately and investigate your claim.

Correlation Between Valve Event Amplitudes in the Seismocardiogram and VO₂-max

Kasper Sørensen¹ (*Member IEEE*), Mathias Krogh Poulsen¹, Dan Stieper Karbing¹ (*Member IEEE*), Peter Søgaard², Johannes J. Struijk¹, and Samuel Emil Schmidt¹ (*Member IEEE*)

Abstract—Cardio Respiratory Fitness (CRF) is a strong predictor of mortality. Because of this the American Heart Association recommends that ideally all patients should have measured CRF with a maximal effort test on a regular basis. However, this is not practically possible. Seismocardiography (SCG) is a measure of the vibration produced by the beating heart, measured on the sternum. This study investigates the correlation between amplitudes in the SCG and VO₂-max from 17 females and 9 males. All amplitudes investigated correlated significantly with VO₂-max.

I. INTRODUCTION

According to an American Heart Association (AHA) Scientific Statement from 2016, CRF is potentially a stronger risk predictor of mortality than smoking, diabetes and hypertension [1]. CRF can be estimated by maximal oxygen consumption (VO₂-max) with a maximal effort cycle ergometer test. This test is not well suited for the elderly or obese patients. In this study we investigated using recordings of vibrations on the chest produced by the beating of the heart, SCG, and its correlation to VO₂-max. SCG is a deterministic signal recorded non-invasively on the sternum with an accelerometer [2].

II. METHODS

Seventeen females and 9 males were recruited for the study. All subjects signed written informed consent form. Demographic data, SCG and electrocardiography (ECG) was recorded for 5 minutes at rest. VO₂-max was measured with a Vyntus CPX during a maximal effort bike ergometer test. The female subjects participated in a follow up recording session, after 8 weeks of structured high intensity training for 2 hours a week. Based on the ECG R-peak, the SCG was divided into individual beats and one mean SCG beat was calculated for each subject. From the mean SCG beats the peak-to-peak amplitudes between the location of the mitral valve closure (E_s) and the following valley F_s, F_s to aortic valve opening (G_s), aortic valve closure (B_d) to the following valley C_d and lastly C_d to D_d was measured (see Fig 1) [3]. The amplitudes were correlated with the VO₂-max using the Pearson’s Correlation.

III. RESULTS

The mean VO₂-max for the females was 28.7 (± 4.5) (mL/kg)/min prior to the 8 weeks of training, compared to




Fig. 1: A normal SCG with notation of the fiducial points.

31.2 (± 6.5) (mL/kg)/min after training ($p < 0.05$). The mean VO₂-max for the males was 40.1 (± 8.4) (mL/kg)/min. The correlations between the amplitudes of the valve events and VO₂-max are listed in Table I.

TABLE I. PEARSON’S CORRELATION COEFFICIENT AND P-VALUE BETWEEN VALVE EVENT AMPLITUDES AND VO₂-MAX

Event	n	Correlation [r]	p-value
E _s to F _s	48	-0.47	< 0.001
F _s to G _s	48	0.47	< 0.001
B _d to C _d	48	-0.72	< 0.001
C _d to D _d	48	0.81	< 0.001

IV. DISCUSSION & CONCLUSION

The amplitude of C_d to D_d could be caused by the heart moving towards its apex when the aortic valve closes [2]. The amplitude of this event could be manipulated by the speed of the pressure drop in the left ventricle. Faster ventricular relaxation has previously been shown in athletes compared to normal subjects.

This study shows a correlation between amplitudes measured in the SCG and VO₂-max, presenting a novel non-exercise method for estimating CRF.

REFERENCES

- [1] R. Ross *et al.*, “Importance of Assessing Cardiorespiratory Fitness in Clinical Practice: A Case for Fitness as a Clinical Vital Sign: A Scientific Statement from the American Heart Association,” *Circulation*, vol. 134, no. 24, pp. e653–e699, 2016.
- [2] P. Mounsey, “Praecordial ballistocardiography,” *Br. Heart J.*, vol. 19, no. 2, pp. 259–271, 1957.
- [3] K. Sørensen *et al.*, “Definition of Fiducial Points in the Normal Seismocardiogram,” *Sci. Rep.*, vol. 8, no. 1, p. 15455, 2018.

¹Aalborg University, Denmark, Health Science and Technology. Contact: ks@hst.aau.dk

²Aalborg University Hospital, Denmark. Department of Cardiology.