



Aalborg Universitet

AALBORG  
UNIVERSITY

## Radiotherapy in Patients with Pacemakers and Implantable Cardioverter Defibrillators

Zaremba, Tomas

DOI (*link to publication from Publisher*):  
[10.5278/vbn.phd.med.00021](https://doi.org/10.5278/vbn.phd.med.00021)

Publication date:  
2015

Document Version  
Publisher's PDF, also known as Version of record

[Link to publication from Aalborg University](#)

Citation for published version (APA):  
Zaremba, T. (2015). *Radiotherapy in Patients with Pacemakers and Implantable Cardioverter Defibrillators*. Aalborg Universitetsforlag. <https://doi.org/10.5278/vbn.phd.med.00021>

### General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal -

### Take down policy

If you believe that this document breaches copyright please contact us at [vbn@aub.aau.dk](mailto:vbn@aub.aau.dk) providing details, and we will remove access to the work immediately and investigate your claim.

# **RADIOTHERAPY IN PATIENTS WITH PACEMAKERS AND IMPLANTABLE CARDIOVERTER DEFIBRILLATORS**

**BY  
TOMAS ZAREMBA**

**DISSERTATION SUBMITTED 2015**



**AALBORG UNIVERSITY  
DENMARK**

# **Radiotherapy in Patients with Pacemakers and Implantable Cardioverter Defibrillators**

**Tomas Zaremba**

**Academic advisors**

*Sam Riahi, MD, PhD (main supervisor)*

Department of Cardiology,

Aalborg University Hospital

*Anna Margrethe Thøgersen, MD, DMSc*

Department of Cardiology,

Aalborg University Hospital

*Mette Søgaard, DVM, PhD*

Department of Clinical Epidemiology,

Aarhus University Hospital

**Members of the assessment committee**

*Professor Ursula G. Falkmer, MD, PhD (chairman)*

Department of Oncology,

Aalborg University Hospital

*Milos Kesek, MD, PhD*

Department of Cardiology,

Norrlund University Hospital, Umeå, Sweden

*Christian Gerdes, MD, PhD*

Department of Cardiology,

Aarhus University Hospital

PhD Series: Faculty of Medicine, Aalborg University

ISSN (online): 2246-1302

ISBN (online): 978-87-7112-270-1

Published by:

Aalborg University Press

Skjernvej 4A, 2nd floor

DK – 9220 Aalborg Ø

Phone: +45 99407140

[aauf@forlag.aau.dk](mailto:aauf@forlag.aau.dk)

[forlag.aau.dk](http://forlag.aau.dk)

Printed in Denmark by Rosendahls, 2015

The Faculty of Medicine, Aalborg University, Denmark has approved this PhD dissertation for public defense. The public lecture and defense will take place May 29, 2015 at 13.00 in the Auditorium in the House of Research, Aalborg University Hospital.

## Preface

This journey began in 2009 during my cardiology training at Aalborg University Hospital. I was fortunate enough to be approached by Anna Margrethe Thøgersen who proposed me to write a case report on a patient with a pacemaker functioning normally despite exposure to a relatively high radiation dose. At that time, it appeared that there were a rather limited number of studies present on the effects of ionizing radiation on heart rhythm devices. During the initial stage of working on the case report, cooperation with Sam Riahi and Annette Ross Jakobsen was established as well. This lead to our next work, where we conducted a survey of Danish radiotherapy and cardiology departments aiming to elucidate practical treatment of patients with a pacemaker or an implantable cardioverter defibrillator undergoing radiotherapy. Afterwards, we resolved to keep working in the field, as the number of pacemaker patients undergoing radiotherapy appeared to be increasing, and there still seemed to be numerous aspects to be explored.

Enrolled as a PhD student in September 2012, I had the great privilege to have Sam Riahi as my main academic advisor. He guided and supported me through the whole project with great enthusiasm. Sam's inspiring commitment and his ability to show the way by seeing solutions in every challenge made this endeavor an exciting experience. I am grateful to Sam for introducing me to Mette Søgaard, who as my academic advisor played a key role in the epidemiological aspects of the study and made invaluable contributions during the writing process. I would also like to acknowledge and thank Anna Margrethe Thøgersen for sharing her great ideas and for being a great academic advisor always ready to discuss both small details and large questions of the project. Her practical electrophysiology skills were essential during our experiments.

Very special thanks go to Annette Ross Jakobsen who has been a fantastic colleague and partner, and who used tremendous amounts of her time and efforts during the experiments and data collection. Being the only physicist in our group, she played a pi-

votal role in guiding me through the technological aspects of radiotherapy. I owe also many thanks to Benedict Kjærgaard and his great team from the Biomedical Research Laboratory, Aalborg University Hospital, for their commitment, immense practical help during the in vivo study, and for making complex things possible and fun.

Furthermore, I thank all colleagues at cardiology and radiotherapy departments in Western Denmark for their enthusiasm and help during the data collection. It has been a great pleasure to work with you during this project. I am especially indebted to Peter Skogholt, Oncology Department, Vejle Hospital, for assisting with extraction of archived radiotherapy data.

In particular I would like to thank Søren Pihlkjær Hjortsøj, the head of the Department of Cardiology, Aalborg University Hospital, for supporting me during the study and for allocating time for the project when it was most needed. I am also grateful to professor Erik Berg Schmidt for his assistance and practical guidance during the PhD study and to Lars Oddershede, Martin Berg Johansen, and Søren Lundbye-Christensen for their statistical support. I thank Lærke Bruun Madsen for her help during data collection in the epidemiology study. I also owe a big thank to Hanne Madsen for revising the manuscripts. Great thanks go as well to the rest of my colleagues at the Department of Cardiology, Aalborg University Hospital; amongst them, I thank fellow PhD student Jacob Moesgaard Larsen for the encouragement and helpful insights.

I am grateful to the manufacturers for donating pacemakers and implantable cardioverter defibrillators for the project: Biotronik, Boston Scientific, Medtronic, Sorin, and St. Jude Medical.

Last, but definitely not the least, I would like to thank my charming wife Sandra and our two great sons Gustas and Ignas. Only your love and tremendous support have made it all possible.

*Tomas Zaremba*  
February 2015

# Table of Contents

Preface	3
Table of Contents	4
List of Papers	4
Abbreviations	4
Introduction	5
Background	6
Aims and hypotheses	7
Materials and methods	8
Results	13
Discussion	18
Conclusions	29
Summary	30
Dansk resumé	31
Appendices	32
References	35
Paper I	39
Paper II	40
Paper III	41

## List of Papers

This thesis was based on the following papers:

1. Zaremba T, Jakobsen AR, Thøgersen AM, Oddershede L, Riahi S. The effect of radiotherapy beam energy on modern cardiac devices: an in vitro study. *Europace*. 2014; 16(4):612–6.
2. Zaremba T, Jakobsen AR, Thøgersen AM, Riahi S, Kjaergaard B. Effects of high-dose radiotherapy on implantable cardioverter defibrillators: an in vivo porcine study. *Pacing and Clinical Electrophysiology: PACE*. 2013; 36(12):1558–63.
3. Zaremba T, Jakobsen AR, Søgaard M, Thøgersen AM, Johansen MB, Madsen LB, Riahi S. Risk of device malfunction in cancer patients with implantable cardiac device undergoing radiotherapy: a population-based cohort study. *Pacing and Clinical Electrophysiology: PACE*. 2015; 38(3):343–56.

# Abbreviations

AAI	single chamber atrial pacemaker
AAPM TG-34	American Association of Physicists in Medicine Task Group No. 34
AMI	acute myocardial infarction
ATP	antitachycardia pacing
bpm	beats per minute
CI	confidence interval
CMOS	complementary metal oxide semiconductor
Co	cobalt
CRT	cardiac resynchronization therapy
CRT-D	cardiac resynchronization therapy defibrillator
CRT-P	cardiac resynchronization therapy pacemaker
CT	computed tomography
DDD	dual chamber pacemaker
DDD-ICD	dual chamber implantable cardioverter defibrillator
DNPR	Danish National Patient Registry
EMI	electromagnetic interference
ERI	elective replacement indicator
eV	electronvolt
Gy	gray
HR	hazard ratio
ICD	implantable cardioverter defibrillator
IQR	interquartile range
kV	kilovolt
LINAC	linear accelerator
LET	linear energy transfer
MeV	megaelectronvolt
MRI	magnetic resonance imaging
MV	megavolt
OR	odds ratio
PM	pacemaker
RT	radiotherapy
RRT	recommended replacement time
VDD	single chamber pacemaker with ventricular lead and dual chamber sensing
VF	ventricular fibrillation
VHR	ventricular high-rate episode
VT	ventricular tachycardia
VVI	single chamber pacemaker with ventricular lead
VVI-ICD	single chamber implantable cardioverter defibrillator

## Introduction

Since the first implantation of a pacemaker (PM) in humans by Senning and Elmqvist in 1958, implantable electronic devices have evolved into the mainstay of the treatment of cardiac rhythm disturbances.<sup>1-3</sup> Initially used for management of bradycardias, these devices have during the last decades been increasingly used for treatment of tachyarrhythmias as well.<sup>4</sup> An implantable PM consists of an impulse generator which is typically placed subcutaneously in the pectoral region and is connected to endocardium via one or two transvenous leads. Powered by a lithium battery, modern PMs rely on complementary metal oxide semiconductor (CMOS) technology, permitting incorporation of up to millions of transistors which in turn enable the usage of sophisticated programmable algorithms in the management of cardiac rhythm disturbances.<sup>5</sup> Besides single and dual chamber PMs, cardiac resynchronization therapy PMs (CRT-P) are implanted in selected patients suffering from systolic dysfunction.<sup>6,7</sup> In addition, due to treatment modalities such as antitachycardia pacing (ATP) and shock therapy, implantable cardioverter defibrillators (ICDs) were proven effective in preventing sudden cardiac death in patients at risk of life threatening ventricular arrhythmias.<sup>8,9</sup> In some devices, both defibrillator and resynchronization functions are used concurrently [cardiac resynchronization therapy defibrillators (CRT-D)].<sup>6</sup>

With more than 700,000 new PMs and more than 200,000 new ICDs implanted worldwide each year, the rate of PM/ICD implantations is increasing both on a global scale and in Europe.<sup>10,11</sup> In

Denmark, 4,725 PMs were implanted in 2013, of which 3,543 (75.0%) were first implants.<sup>11</sup> The corresponding numbers for ICDs were 1,285 and 890 (69.3%), respectively. In addition, 1,001 cardiac resynchronization therapy (CRT) units were implanted in Denmark in 2013.<sup>11</sup>

As the functionality of modern PMs/ICDs to a high extent relies on sensing the intrinsic electrical signals of the heart, these devices may be susceptible to extraneous signals.<sup>12</sup> In order to mitigate these effects, the manufacturers have introduced protective measures such as shielding in hermetic metal cases, signal filtering, interference rejection circuits, modern alternatives to reed switches, and use of bipolar leads.<sup>12,13</sup> However, hazardous factors may still be present in the medical environment: e.g. electrosurgery, direct current external defibrillation, magnetic resonance imaging (MRI), neurostimulation, radiofrequency catheter ablation, lithotripsy, diagnostic radiation, and cancer radiotherapy (RT).<sup>14</sup> As interference from these factors may lead to malfunction of the cardiac rhythm devices, the ability to predict and reduce these negative effects plays a central role for safe treatment of PM/ICD patients in these circumstances.

This project focuses on external beam RT for cancer in PM/ICD patients. Based on three research papers, the present work emphasizes the epidemiological, clinical, and safety aspects in the management of patients with PM/ICD undergoing RT.

## Background


Due to the ageing and growth of the world population, the burden of cancer is increasing.<sup>15-17</sup> The number of new cancer cases is expected to increase worldwide from 12.7 million in 2008 to 21.4 million by 2030,<sup>18</sup> with lung cancer being the most frequently diagnosed cancer type in males and breast cancer in females.<sup>16</sup> Likewise, the incidence of cardiovascular diseases is increasing.<sup>19,20</sup> Being responsible for 31% of deaths worldwide,<sup>21</sup> cardiovascular diseases are projected to maintain this leading position by 2030.<sup>22</sup> As a result of the age being a risk factor for both cardiovascular diseases and cancer, there is a growing probability that some of PM/ICD patients will develop malignancies and receive RT treatment.

Among cancer treatment options, RT has become an established therapy method in oncology, in both curative and palliative intent, with at least 50% of all cancer patients requiring RT during the course of their illness.<sup>23</sup> RT uses high energy radiation to kill or damage cancer cells and stop them from growing and multiplying. Radiation doses used in cancer therapy are measured in grays (1 Gy = 1 joule of absorbed energy of ionizing radiation per 1 kg of matter). Commonly, RT is given as a course of several treatments over days or weeks with daily fractions of typically 1.8-2 Gy. This serves two main purposes. Firstly, normal cells are allowed to recover between fractions. Secondly, the survived tumor cells may have entered a more radiation-sensitive phase of the cell cycle before a subsequent fraction. Cumulative doses of curative RT for solid tumors generally range from 50 to 80 Gy.<sup>24-26</sup> A typical radiation dose for breast cancer is about 50 Gy, while cumulative doses of at least 60 Gy are administered for lung cancer.<sup>27,28</sup> Lower doses of 20-40 Gy may be applied in treatment of lymphomas,<sup>29,30</sup> whereas RT for bone metastases usually consists of 8-30 Gy in 1-10 fractions.<sup>31,32</sup>

At present, the most commonly used types of radiation in RT are photons or electrons, which are generated and delivered by a RT machine called linear accelerator (LINAC). Along with other characteristics, the radiation beams are often described by their depth dose curves (Figure 1). By increasing the beam energy of the LINAC, the depth of the maximal delivered radiation dose also increases. Hence, photons in megavolt (MV) range (commonly 6-20 MV) are used for more deeply located tumors, whereas electrons due to their sharp decline with increasing depth and hence finite range are typically used for superficially located tumors. Kilovolt (kV) photons are also often used for superficial lesions such as skin cancer.

RT is normally delivered according to an individually designed treatment plan based on image data derived from modalities such as computed tomography (CT), MRI, positron emission tomography, and ultrasound. When creating this treatment plan, beam energy is chosen according to depth of the tumor, also taking other parameters such as the number of RT fields and their angles into account.

Figure 1: Depth doses for different radiation types and energies.



kV = kilovolt; MeV = megaelectronvolt; MV = megavolt.

Besides injuring the tissues, ionizing radiation may cause damage to the circuits in electronic implants. In the early years of treatment with PMs, RT did not pose any considerable threat, as the devices from the 1960s and the beginning of the 1970s were based on discrete bipolar transistors and were found to be highly resistant to ionizing radiation.<sup>33-35</sup> In comparison, modern PMs and ICDs rely on CMOS circuitry which has the advantages of greater reliability and lower power consumption.<sup>36</sup> However, these devices have been reported to be more susceptible to malfunctions at exposure to ionizing radiation.<sup>37</sup>

Official guidelines for managing PM patients undergoing RT were published by The American Association of Physicists in Medicine Task Group No. 34 (AAPM TG-34) in 1994<sup>38</sup> and give no recommendations for RT in ICD patients. Furthermore, the recommendations by PM/ICD manufacturers vary regarding both tolerable dose of ionizing radiation and follow-up.<sup>39,40</sup> Reimplantation of a cardiac device or use of a temporary PM is currently advocated before RT if the maximal dose to the PM exceeds 2-10 Gy, while removal of the ICD is recommended at even lower radiation doses to the device.<sup>36,38-41</sup> While these dose levels are lower than the cumulative target doses used in cancer treatment, every additional surgical intervention to the PM/ICD exposes the patient to a substantial hazard of infectious and surgical complications<sup>42-44</sup> and likely augments healthcare costs. Importantly, while some devices are able to resist radiation doses considerably higher than recommended as safe,<sup>45,46</sup> other PMs/ICDs may malfunction despite exposure to only scattered radiation from RT to an anatomically remote area.<sup>47-51</sup>

Besides external beam RT, RT is in some cases delivered as short range brachytherapy.<sup>52</sup> While the literature on the effects of this treatment modality on PMs/ICDs is limited, no device malfunctions during brachytherapy have been reported so far.<sup>53-55</sup> Meanwhile, although there seems to be no solid evidence that kV photons should harm modern PMs/ICDs,<sup>56</sup> a few prior case reports have described PM malfunctions during diagnostic (kV) radiation, e.g. in relation to CT.<sup>57</sup>

# Aims and hypotheses

## Aims

### **Study I**

- To assess the influence of high-energy (18 MV) photon beams on modern PMs and ICDs compared to low-energy (6 MV) photon beams.

### **Study II**

- To evaluate the effects of cumulative radiation dose and beam energy on ICDs *in vivo*.
- To determine the feasibility of a porcine model to study the effects of ionizing radiation on ICDs.

### **Study III**

- To quantify the annual rates of RT in patients with PM/ICD.
- To elucidate safety measures used in clinical practice during RT in PM/ICD patients.
- To quantify the frequency of PM/ICD malfunctions during RT.
- To identify the predictors of PM/ICD malfunctions during RT.

## Hypotheses

- Modern PMs/ICDs can resist higher doses of ionizing radiation than generally anticipated (Study I and II).
- Animal models are feasible for studying the effects of RT on ICDs *in vivo* (Study II).
- The rate of RT in PM/ICD patients in the general population is increasing (Study III).
- The use of safety measures varies during RT in PM/ICD patients in clinical practice (Study III).
- PM/ICD malfunctions can be predicted based on parameters of RT and/or type of the device (Study I and III).

# Materials and methods

## Study I

The purpose of the study was to evaluate the effects of high-energy photon beams on modern PMs/ICDs compared to the effects of low-energy photon beams in a realistic clinical scenario mimicking the actual RT doses used in treatment of a breast cancer.

### Devices

Ten unused PMs and two explanted fully functional ICDs were exposed to either 6 MV or 18 MV photons (Table 1).

**Table 1: Devices irradiated in vitro in Study I.**

6 MV photons		18 MV photons	
Device type	Manufacturer and model	Device type	Manufacturer and model
DDD	Biotronik Evia DR-T	DDD	Biotronik Evia DR-T
DDD	Boston Scientific Altrua 60	DDD	Boston Scientific Altrua 60
DDD	Medtronic Adapta L	DDD	Medtronic Adapta
DDD	Sorin Esprit DR	DDD	Sorin Esprit DR
DDD	St. Jude Medical Zephyr XL DR	AAI/VVI	St. Jude Medical Zephyr XL SR
VVI-ICD	Medtronic Secura VR	DDD-ICD	Medtronic Maximo II DR


AAI = single chamber atrial pacemaker; DDD = dual chamber pacemaker; DDD-ICD = dual chamber implantable cardioverter defibrillator; MV = megavolt; VVI = single chamber pacemaker with ventricular lead; VVI-ICD = single chamber implantable cardioverter defibrillator.

The PMs were programmed with standard settings, e.g. DDDR 60–130 beats per minute (bpm), output 3.5 V / 0.4 ms on both channels. Regarding the ICDs, antitachycardia pacing and shock therapies were inactivated. Ventricular tachycardia (VT) monitor zones were programmed active, e.g. VT zone from 167 bpm and ventricular fibrillation (VF) zone from 214 bpm. All lead connector ports were closed with pin plugs.

### Irradiations

Each device was irradiated repeatedly with 2 Gy daily for five days followed by a two-day break. The photon beams were generated by a Clinac iX LINAC (Varian Medical Systems, Inc., Palo Alto, CA, USA) and delivered with a dose rate of 600 monitor units/min. During irradiations, the devices were placed in a custom manufactured polymethyl methacrylate phantom (Figure 2) placed between adequate build-up material of solid water boards. This permitted locating the PMs/ICDs at the depth of dose maximum for each photon energy, as the depth where maximum dose is delivered correlates with beam energy and field size. The distance from the radiation source to the surface of the phantom including build-up material was 100 cm. The irradiation field was 10 cm x 10 cm for the PMs and 15 cm x 15 cm for the ICDs. RT treatment planning software (Eclipse v. 10.0, Varian Medical Systems, Inc., Palo Alto, CA, USA) was used to plan the irradiations.

**Figure 2: Pacemaker located in the radiotherapy field in a polymethyl methacrylate phantom.**



After reaching a cumulative dose of 70 Gy, the doses per fraction in the 6-MV group were increased. They consisted of 10, 10, 20, and 40 Gy and were delivered during the same day. In the 18-MV group, single doses were increased after reaching 50 Gy to 10, 10, 10, 20, 20, and 30 Gy. After reaching 80 Gy in this group, the intervals between the irradiations were prolonged to a median of 55 days [inter-quartile range (IQR) 28–75]. The irradiations were chosen not to be performed during the same day in order to avoid exposing the investigators to the increased in-room level of induced radioactivity due to secondary neutrons. The intervals between irradiations were also prolonged due to logistic constraints at our institution. In both the 6- and the 18-MV group, cumulative radiation doses of 150 Gy per device were delivered.

### Interrogations

The PMs and ICDs were interrogated after every radiation dose either on the same or on the following day, using manufacturer-specific standard telemetry equipment. Presence or absence of the following events was recorded:

- Noise during RT sessions;
- Spontaneous change in programmed device parameters without reset to backup mode;

- Reset to backup mode or other error, recoverable using the programmer;
- Error, not recoverable using the programmer;
- Clinically significant reduction in battery capacity;
- Inappropriate antitachycardia pacing or delivery of shock therapy in the ICDs in spite of deactivation of these functions;
- Loss of telemetry.

When all irradiations were completed, the devices were interrogated at least twice during a period of at least two months.

## Study II

The study was performed as a porcine *in vivo* experiment of accelerated RT delivered to implanted modern ICDs simulating a worst case scenario of a device irradiated directly in the RT field.

### Implantation procedures

Five pigs (1 Göttingen minipig and 4 Danish Landrace pigs), all weighing around 40 kg, were implanted with ICD systems in our Biomedical Research Laboratory. The reason for switching from one pig race to another was purely logistic.

Prior to the implantation procedures, the animals were pre-anesthetized with intramuscular injection of Zoletil. Zoletil is a veterinarian medicine consisting of a mixture of two dissociative anesthetics (Ketamine 6.25 mg/ml and Tiletamine 6.25 mg/ml), a benzodiazepine (Zolazepam 6.25 mg/ml), a synthetic opioid (Butorphanol 1.25 mg/ml), and Xylazine (6.25 mg/ml).

The animals were intubated and ventilated with Sevoflurane 1% using a Dameca Dream anesthesia machine (Dameca, Rodovre, Denmark). Volume-controlled respiration was used. During surgery, the anesthesia was maintained with intravenous infusion of Fentanyl 50 µg/ml at 10 ml/h rate and infusion of Midazolam 5 mg/ml at 10 ml/h rate. The blood pressure was monitored invasively via femoral artery cannulation. If needed, single doses of intravenous Ketamine 50-100 mg (50 mg/ml) as an anesthetic, potassium chloride 10 mmol to correct hypokalemia, and Lidocaine 100 mg (10 mg/ml) for ventricular arrhythmias were administered during the anesthesia. Antibiotic prophylaxis consisted of a single dose of periprocedural intramuscular Gentamycin 80 mg (40 mg/ml) and intramuscular Benzylpenicillin 5 millions IU, the latter continued q.d. for 3 days.

The following ICD generators from five different manufacturers were implanted:

- Biotronik Lumax 540 DR-T, unused;
- Boston Scientific Energen ICD F141, unused;
- Medtronic Maximo II CRT-D D284TRK, unused;
- Sorin Paradym SonR CRT-D 8770, explanted;
- St. Jude Medical Unify CD3235-40, unused.

The Medtronic ICD was implanted in a Göttingen minipig, while the remaining devices were implanted in Danish Landrace pigs.

Intravenous access for placement of the leads was acquired by Seldinger technique through a puncture of the brachiocephalic vein. Medtronic Capturfix Novus 5076 52 cm active fixation leads (Medtronic Inc., Minneapolis, MN, USA) were used as right atrium

leads. St. Jude Medical Durata 7120, 65 cm active fixation leads (St. Jude Medical, Inc., St. Paul, MN, USA) were implanted in the right ventricular septum in all animals. In the Sorin ICD, left-sided Medtronic Attain Ability Plus 4296, 88 cm (Medtronic, Inc., Minneapolis, MN, USA) lead was implanted to a branch in the coronary sinus as well. Due to price constraints, the left ventricle connector ports in the Medtronic and St. Jude Medical ICDs were left unused and were closed with pinplugs. The positions of the leads in the heart were verified by fluoroscopy. The generators were fixed and placed in a subcutaneous pocket after making an incision on the left side of the cranial part of the sternum. Shock therapy was not tested at the implantations.

In the postoperative period, all animals were housed in the laboratory and observed for signs of infection and failure to thrive. Intramuscular injections of Ketoprofen 300 mg (150 mg/ml) t.i.d. for 3 days were used as a pain killer.

### Irradiations

After an average observation time of 9 days (range 4-18), the animals were anesthetized with intramuscular injections of Zoletil and transported to one of the treatment rooms in our Radiotherapy Department. The initial dose of Zoletil was 4-5 ml, supplemented with 2 ml as needed during the study. During the transportation and the irradiations, the pigs were intubated and ventilated with Sevoflurane 1% by a portable anesthesia machine Siemens SV 900 (Siemens AG, Munich, Germany).


The animals were positioned in a supine position supported by a vac lock bag (Par Scientific A/S, Odense, Denmark) on the accelerator couch. A Varian Clinac iX LINAC (Varian Medical Systems, Inc., Palo Alto, CA, USA) was used (Figure 3). During the study, the heart rhythm was monitored with an ECG monitor visible in the operator room.

The ICDs were interrogated, and all antitachycardia therapies were programmed OFF, while detection was left ON. Two or three zones were programmed, e.g. a VT zone between 150 bpm and 200 bpm, a fast VT zone between 200 bpm and 250 bpm, and a VF zone above 250 bpm.

RT with 6 MV photons and field dimensions of 15 cm x 15 cm, with the ICD generator in the center, was planned with a source-to-surface distance of 100 cm. The area above the ICD generator was covered with 5 mm Superflap build-up material (Mick Radio Nuclear Instruments Inc., Mt. Vernon, NY, USA) in order to achieve the adequate dose to the ICD. The gantry was rotated to ensure a perpendicular direction of the beam toward the ICD generator.

Incremental doses of 6 MV photons to the ICD generators were delivered: 0.5 Gy, 1.0 Gy, 2.0 Gy, 5.0 Gy, and 10.0 Gy, with a total radiation dose of 18.5 Gy. Photon energy was then increased to a maximum of 18 MV. The area over the ICD generators was covered with additional 10 mm of Superflap build-up material to a total of 15 mm in order to achieve the adequate dose for the ICDs. Irradiation with correspondingly increasing identical doses to the ICDs was performed, thus reaching a cumulative dose of 37.0 Gy in all cases.

Figure 3: Practical setup during the *in vivo* irradiations.



ICD = implantable cardioverter defibrillator.

### Interrogations

The ICDs were interrogated after every dose step. Programmed parameters were assessed, and battery status, lead impedance, sense, and capture threshold were measured. After the irradiations, the shock function of the ICDs was tested. Before this, all pigs were treated with extra anesthetics and with Rocuronium, a muscle relaxant. The shock function of the ICDs was tested with both single- and dual-coil setting. VF was induced by T-wave shock or 50 Hz pacing.

Upon completion of the shock testing, a new control of the devices was performed. Afterwards, the animals were killed with an overdose of intravenous Pentobarbital 6 g (300 mg/ml), and the ICDs were removed. All ICDs were also interrogated 2, 4, and 17-18 days later. Time period from the irradiations to last additional interrogation ranged from 75 to 402 days.

## Study III

### Study population

The Danish National Patient Registry (DNPR) contains information on all inpatient hospitalizations at Danish non-psychiatric hospitals since 1977 and on all emergency room and hospital outpatient specialist clinic visits since 1995.<sup>58</sup> Each hospital visit is recorded by physicians with one primary diagnosis and one or more secondary diagnoses classified according to the International Classification of Diseases, 8th edition (ICD-8) until the end of 1993, and ICD-10 thereafter. The DNPR also includes codes for performed procedures: Danish Hospital Sector Classification System (used since 2000); Nordic Medico-Statistical Committee (NOMESCO) Classification of Surgical Procedures (used in 1996-2000); and procedure codes used historically before NOMESCO.

We used the DNPR to identify all patients in Western Denmark with at least one PM/ICD-related procedure or diagnosis code registered from January 1, 1977 to December 31, 2012 (all codes are provided in Appendix A) who also had a code for RT (planning of RT or external RT) recorded between January 1, 2003 and December 31, 2012. Patients with RT performed prior to insertion of their PM/ICD were excluded.

### Radiotherapy data

Data on RT treatments were collected from RT-planning systems at Aalborg University Hospital, Aarhus University Hospital, Herning Regional Hospital, and Vejle Hospital. In cases of incomplete data in the planning systems, the medical files were reviewed. If needed, data from the DNPR were subsequently used to identify the start and/or end dates and the number of fractions of RT course.

Data included information on the hospital where RT was delivered, start date, end date, anatomical region irradiated, cumulative tumor dose, number of fractions, fraction dose (maximal applied during the RT course), beam type, and beam energy (maximal applied during the RT course). The anatomical regions were classified as: head and neck, thorax, esophagus, abdomen and pelvis, spine (thoracic and lumbar), upper extremity, and lower extremity. In case two anatomical regions were treated simultaneously, the one closest to the PM/ICD generator was recorded. If both photons and electrons were applied, the treatment was classified as photon RT.

### PM/ICD data

Data on PMs/ICDs were collected from implanting cardiology departments at the following hospitals in Western Denmark: Aalborg University Hospital, Aarhus University Hospital, Esbjerg Hospital, Haderslev Hospital, Herning Regional Hospital, Kolding Hospital, Vejle Hospital, and Viborg Regional Hospital. The data included information on device class (PM, CRT-P, ICD, or CRT-D), device type [single chamber atrial PM (AAI), single chamber PM with ventricular lead (VVI), single chamber pacemaker with ventricular lead and dual chamber sensing (VDD), dual chamber pacemaker (DDD), CRT-P, single chamber ICD (VVI-ICD), dual chamber ICD (DDD-ICD), or CRT-D], generator manufacturer, model, hospital of implantation, and follow-up.

### Safety measures

For each RT course we collected information on potential safety measures: evaluations at the PM/ICD clinic including visits before,

during, and after RT. Scheduling of supplementary visits up to three months after RT was recorded as well. Other recorded safety measures were reprogramming of the device before RT, application of a magnet to the ICD during RT, use of a temporary PM, and surgical relocation of the generator.

Complications related to application of a temporary PM or relocation of the device occurring during a six-month period were also recorded. The following complications were defined as major: lead-related re-intervention, local infection requiring re-intervention, device-related systemic infection/endocarditis, pneumothorax requiring drainage, cardiac perforation (without or with intervention), pocket revision because of pain, generator-lead interface problem with re-intervention, haematoma requiring re-intervention, deep venous thrombosis, Twiddler's syndrome, wound revision, stroke, acute myocardial infarction (AMI), procedure-related death. Minor complications included haematoma without re-intervention, resulting in a prolonged hospital stay, hospital re-admission, or additional out-patient visit, wound infection treated with antibiotics, conservatively treated pneumothorax, and lead dislodgement without re-intervention, as suggested by Kirkfeldt et al.<sup>59</sup>

## Outcome

Information on PM/ICD malfunctions potentially attributable to RT was obtained through reviews of cardiology records, including PM/ICD follow-up charts. Remote monitoring controls were included if documented in the patient file. The follow-up period ended at first PM/ICD evaluation after completion of RT course or on December 31, 2013, whichever came first.

PM/ICD malfunctions were categorized as:

- Electrical reset to backup mode or other minor software error;
- Electrical reset or other software error requiring reprogramming of the device by the manufacturer;
- Unexpected decrease in battery capacity without reaching elective replacement indicator (ERI);
- Unexpected ERI;
- Loss of telemetry;
- Change in one or several lead parameters eventually resulting in supplementary visits or lead replacement (only the changes suspected at the subsequent device control to have occurred due to RT and not explained by other variables, such as changes in antiarrhythmic drugs, were recorded);
- Noise oversense without symptomatic pacing inhibition, ATP, or shock therapy;
- Oversense with symptomatic pacing inhibition, ATP, or shock therapy.

## Statistical analyses

### Study I

Using the cumulative dose of ionizing radiation as a substitute for time scale, an equivalent of survival analysis was performed until first potentially clinically hazardous failure for every device.

The data were interval censored as the exact radiation dose at the exact time malfunction occurred was unknown. To accommodate for this, the events were placed either at the starting-point of the interval, at the mid-point, or at the end-point of the interval.

Kaplan-Meier curves were plotted to illustrate this approach. Using a Cox proportional hazard regression model, the mid-points were compared. In the same manner, start-point events in the 6-MV group were compared to end-point events in the 18-MV group, and reversely. Due to the low number of events, p-values do not have any practical interpretation and confidence intervals (CIs) may not have 95% coverage. Hence we refrain from reporting p-values, and emphasize that caution should be taken when interpreting CIs.

The incidence rate of all potentially hazardous malfunctions was compared between the two groups with regard of the cumulative dose. To accommodate for correlation within pacemakers a population averaged repeated measures logistic regression model was applied to detect potential differences between groups.<sup>60</sup> This requires a balanced design between groups. The dose per fraction were not the same for the two groups, hence a balanced design was achieved by collapsing non-overlapping intervals.

### Study II

In this descriptive explorative study, continuous variables were expressed as absolute values or means. Changes in PM/ICD battery voltages were analyzed by linear regression.

### Study III

Continuous variables were reported as medians and IQRs. Categorical variables were expressed as counts and percentages. Continuous variables were compared by Wilcoxon rank sum test. The annual rate of RT in PM/ICD patients was calculated by using the total Western Denmark population (obtained from Statistics Denmark).

The device malfunctions and safety measures were compared at RT treatment course (consisting of one or more fractions) level. Only RT courses with a later device control were included in the analysis of the malfunctions.

Odds ratios (ORs) with 95% CIs of PM/ICD malfunctions were computed using logistic regression. Independent variables in the model were type of device (ICD vs. PM), anatomical region irradiated (below vs. above the diaphragm), cumulative radiation dose to the tumor (10 Gy increments), fraction dose (1 Gy increments), and beam energy ( $\geq 15$  MV vs.  $< 15$  MV). The cut-off value of 15 MV for beam energy was chosen entirely based on sensitivity and specificity, since this value gave a high sensitivity to detect device malfunctions (79%) and a higher specificity (61%) than other values with equally high sensitivity. Adjusted ORs were adjusted for beam energy.

As some patients received more than one RT course, the RT courses were not completely independent. To accommodate for this dependence, the method of generalized estimating equations was used in a generalized linear model.<sup>61</sup>

Statistical analyses were performed using Stata versions IC 11.2 and 13.1 (StataCorp LP, College Station, TX, USA).

## Ethics

Prior to the experiments in Study II, a written permission to conduct the study was obtained from the Danish Animal Experiments Inspectorate (permission number 2011/561-59). Study III was approved by the Danish Health and Medicines Authority (record number 3-3013-300/1) and the Danish Data Protection Agency

(record number 2008-58-0028) allowing the researchers to access registry data and to review medical records of the study patients.

# Results

## Study I

Detected PM and ICD malfunctions are summarized in Table 2. In the 6-MV group, no malfunctions were detected in Biotronik, Boston Scientific, St. Jude Medical, or Sorin PMs. The Medtronic PM suddenly lost telemetry capability after reaching a cumulative dose of 150 Gy. The telemetry capabilities were neither present 6 nor 29 days later. However, the device was able to communicate at a supplementary interrogation 269 days after the last irradiation. The PM reported an electrical reset 81 days after last RT.

At interrogations, all devices from Medtronic reported multiple ventricular high-rate episodes (VHRs). These episodes usually lasted a few seconds and were not related in time to RT.

In the 18-MV group, the Medtronic ICD lost its preprogrammed patient data after reaching 44 Gy. No other malfunctions in this device were recorded, except for the above mentioned susceptibility to report artifacts as VHRs. All PMs in the 18-MV group exhibited some degree of potentially hazardous failure. The most common abnormal behavior was electrical reset, which is a fallback to backup or "safe" mode. The PMs could be reprogrammed to initial settings by using automatic algorithms in the programmers except the St. Jude Medical PM (after 150 Gy). In the Medtronic PM, battery depletion was present after reaching 150 Gy. All devices in the 18-MV group preserved their telemetry capabilities. No inappropriate ATP or shock therapy was reported by the ICDs.


Figure 4 shows the radiation dose given before first malfunction. The Cox-regression analysis comparing the assumption of events occurring at the mid-point in both groups showed a hazard ratio (HR) of 9.11 (approximate 95% CI: 1.04-79.69). Comparison of events occurring at the start-point of intervals in the 6-MV group to end-point in the 18-MV group yielded the same HR and CI, as events occurred in the same order. Assuming that events occur at the end-point in 6-MV and at the start-point in the 18-MV yielded a HR of 11.32 (approximate 95% CI: 1.24-103.55).

**Table 2: Recorded pacemaker and ICD malfunctions during the irradiations. A cumulative dose of 150 Gy was reached in all devices.**

Device	Malfunctions in the 6-MV group	Malfunctions in the 18-MV group
Biotronik PM	None	Reset after 100, 120, and 150 Gy
Boston Scientific PM	None	Reset after 30 Gy
Medtronic PM	No telemetry after 150 Gy	RRT/ERI detected after 150 Gy
Sorin PM	None	Reset after 80, 120, and 150 Gy
St. Jude Medical PM	None	Reset after 32, 42, 80, 100, and 120 Gy Error after 150 Gy
Medtronic ICD	None	None, except for loss of patient data after 44 Gy

ERI = elective replacement indicator; Gy = gray; ICD = implantable cardioverter defibrillator; MV = megavolt; PM = pacemaker; RRT = recommended replacement time.

**Figure 4: Kaplan-Meier graph of dose to first malfunction.**



Typical therapeutic doses are marked on the dose axis. Gy = gray; MV = megavolt.

The incidence rates of all episodes of potentially hazardous malfunctions in the two groups in regard to cumulative dose were compared by repeated measures logistic regression. The 18-MV group showed an increased risk of malfunction with an OR of 18.29 (approximate 95% CI: 1.52-219.41).

## Study II

The animals tolerated the implantation procedures, transports, and irradiations well. No significant hemodynamic disturbances or infectious complications were present.

During all irradiations, the animals maintained sinus rhythm without any arrhythmias on the monitor. Programmed settings in all the ICDs remained stable between the fractions. During the irradiations, the devices reported no stored arrhythmias, noise, or oversense events. No reset, other unexpected behavior, or malfunction in the ICDs was observed during the irradiations. The ICDs detected VF correctly and delivered therapy as programmed in all tests. Defibrillator threshold in dual coil and active can configuration ranged from 10.4 to 41 J, and from 10 to 41 J in single coil and active can.

In the Medtronic ICD, a temporary decrease in battery voltage by 0.16 V or 5.2% was observed starting at a cumulative dose of 18.5 Gy. This decline was statistically significant by linear regression analysis, with 0.018 V between interrogations ( $p=0.028$ ). Battery voltage at interrogations 2, 4, and 18 days after the irradiations was 2.97 V, 2.98 V, and 3.06 V, respectively, thus returned to the initial value.

In the Sorin ICD, the magnet rate had decreased from 91 to 85 bpm at interrogation 2 days after the irradiations. At the same time, the battery voltage decreased from 3.0 to 2.9 V. These parameters increased gradually and reached 89 bpm and 3.0 V, respectively, at day 26.

The Biotronik ICD experienced a fallback to a back-up mode at 00:01 the night after the irradiations. The device had thus reverted

to a safe program: pace mode VVI 70 bpm, output 7.5 V / 1.5 ms, VF zone from 150 bpm (400 ms), maximum energy shock of 40 J. The ICD could not be reprogrammed using a clinical interface in the programmer, and a firmware update of the ICD had to be performed by the manufacturer. Afterwards, no malfunction was detected at controls up to 75 days from the irradiations.

Compared with data at the day 2 control, the Boston Scientific ICD showed an increase in power consumption and a decrease in remaining battery charge at the interrogation 4 days after the irradiations. The power consumption increased from 31 to 40  $\mu$ W, and the remaining battery charge decreased from 1.70 to 1.67 Ah. The parameters were at 31  $\mu$ W and 1.67 V at the day 17 interrogation and remained stable.

During the study, all lead parameters, including impedance, sense, and capture threshold, remained stable.

## Study III

### Descriptive characteristics

Among 690 PM/ICD patients recorded in the DNPR, we included 560 (81.2%) patients with 678 separate RT courses in the study (Figure 5). Among the 130 excluded patients, 79 (60.8%) only had a temporary PM before RT, while the remaining 51 (39.2%) were excluded due to not receiving RT, not having a PM/ICD at all, PM/ICD not being implanted until RT, only having a loop recorder, or RT being started after 2012.

The annual rate of RT courses in PM/ICD patients increased by 199% from 1.45 to 4.33 per 100,000 person-years from 2003 to 2012 (Figure 6).

The median age at start of RT (first RT in case of several RT courses) was 75.6 years (IQR 69.3–81.7 years), with predominance of males (68.4%). Most patients had only one device during the study period. Six patients (1.1%) had two devices (Table 3).

Bradycardia PMs constituted the majority of the devices [462 (82.5%)] of which dual chamber models were dominant [331 (59.1%) of 560 devices]. There were 25 (4.5%) CRT-Ps. Defibrillators, including ICDs [54 (9.6%)] and CRT-Ds [19 (3.4%)], represented 73 (13.0%) of the devices.


Among the 678 separate RT courses, complete data on both the device and RT were available in 658 (97.1%), of which 453 (68.8%) had at least one subsequent PM/ICD control. One-year mortality among patients with complete data on last RT course, but no device control afterwards, (n=185) was 93.5% compared with 28.2% among those who had a device control (n=358). In the RT courses with subsequent control, patients tended to be younger and were more often treated with kV photons or electrons, while tumor dose, beam energy, proportion of ICDs, as well as frequency of RT to thorax were higher compared with RT courses without control.

The most common anatomical region was thorax (36.0%) followed by head and neck (27.2%), and abdomen and pelvis (27.1%). The remaining 9.7% regions were spine, extremities, and esophagus. Median time from device implantation to start of RT was 2.7 years (IQR 1.0–5.0).

The PMs/ICDs were manufactured by 13 companies, with Medtronic most frequently represented [227 (40.1%) of 556 devices


with known manufacturer], followed by St. Jude Medical [175 (30.9%)] (Appendix B).

**Figure 5: Flowchart of the study population in Study III.**



ICD = implantable cardioverter defibrillator; PM = pacemaker; RT = radiotherapy.

**Figure 6: Annual rates of radiotherapy courses in patients with pacemakers or implantable cardioverter defibrillators in Western Denmark, 2003–2012.**



ICD = implantable cardioverter defibrillator; RT = radiotherapy.

**Table 3: Baseline characteristics of 560 patients with pacemaker or implantable cardioverter defibrillator receiving radiotherapy, 2003-2012.**

Variable	
Age, years <sup>†</sup>	75.6 (69.3-81.7)
Male, n (%)	383 (68.4)
Female, n (%)	177 (31.6)
Number of RT courses per patient*	1 (1-1)
1, n (%)	470 (83.9)
2, n (%)	73 (13.0)
3, n (%)	13 (2.3)
4, n (%)	2 (0.4)
5, n (%)	1 (0.2)
10, n (%)	1 (0.2)
Age of the device at RT, years <sup>†</sup>	2.6 (0.9-4.8)
Device class <sup>†</sup>	
Single chamber PM, n (%)	130 (23.2)
Dual chamber PM, n (%)	331 (59.1)
PM, unspecified, n (%)	1 (0.2)
CRT-P, n (%)	25 (4.5)
ICD (single and dual chamber), n (%)	54 (9.6)
CRT-D, n (%)	19 (3.4)

CRT-D = cardiac resynchronization therapy defibrillator; CRT-P = cardiac resynchronization therapy pacemaker; ICD = implantable cardioverter defibrillator; PM = pacemaker; RT = radiotherapy. \*Values are shown as median (interquartile range). <sup>†</sup> During first RT course with the first device in case of several RT courses and/or several devices in the patient.

### Safety measures


Precautionary device relocation from the RT field was performed in 24 (3.5%) of 678 RT courses. In these cases, the same generator was reused in two (8.3%) patients. At least one lead was extracted in eight (33.3%), and at least one new lead was implanted in 20 (83.3%) procedures. Among RT courses to thorax with complete data (n=237), in cases where the device was relocated, both the cumulative radiation dose and the beam energy were higher compared with RT courses without PM/ICD relocation [46.5 Gy (IQR 22.5-48 Gy) and 20 Gy (IQR 10-30 Gy), respectively, p=0.0001, and 15 MV (IQR 9-18 MV) and 6 MV (IQR 6-15 MV), respectively, p=0.0011]. One of the patients suffered a major complication after device relocation (AMI). In terms of minor complications, the same patient had a displacement of the atrial lead without intervention. Another patient experienced a hematoma in the pacemaker pocket resulting in additional clinic visits. No backup temporary PM was used for any of the RTs.

Data on scheduling visits at PM clinics specifically due to upcoming RT was available for 655 RT courses in 549 patients. In 101 (15.4%) of these RT courses, the patient was seen both before and upon completion of the RT course, while 47 (7.2%) patients were only controlled before RT and 82 (12.5%) only had a control after completion. Thus, device controls due to RT before and after the RT course were performed in 148 (22.6%) and 183 (27.9%) RT courses, respectively. Among 533 RT courses consisting of more than one fraction, device control was performed at least once during the RT course in 76 (14.3%) of RT courses. A supplementary control within three months after RT was performed in 26 (4.0%) RT courses. The proportion of RT courses leading to device controls was highest among RT courses with photons in MV range, followed by electrons and kV photons (Figure 7). The median time from last RT fraction to first device control was 31 days (IQR 2-145 days) in 453 RT courses with a subsequent control. In 205 (31.2%) RT courses the device was never controlled after the RT (Figure 5).

The proportion of RT courses leading to a subsequent device evaluation was higher when the tumor was located above the


diaphragm and in cases where higher beam energy was used (Figure 8). Hence, at <15 MV, the patients were scheduled for a device evaluation after 26.2% of RT courses given above the diaphragm and in 12.9% of those under the diaphragm (24.2% in total at <15 MV). At ≥15 MV, the proportions were 42.0% and 29.1%, respectively (34.4% in total at ≥15 MV).

**Figure 7: Proportions of radiotherapy courses with device controls scheduled due to radiotherapy, classified by beam type.**



kV = kilovolt; MV = megavolt; RT = radiotherapy. \*At least one control in radiotherapy courses consisting of >1 fraction.

**Figure 8: Percentages of radiotherapy courses with a subsequently scheduled device evaluation, by the anatomical localization of the tumor and beam energy.**



MV = megavolt.

Application of a magnet to the ICD during irradiations after cardiologists recommendation was used in 8 (10.8%) of 74 treatments with implanted ICD in treatments with available data on safety measures. Reprogramming of the device prior to RT was only documented in nine PMs and one ICD (1.5% of 655 RT courses), and it mainly consisted of increasing pacing output and/or reprogramming to fixed-mode pacing. In the ICD, the antitachycardia therapies were inactivated during the RT sessions.

### Device malfunctions

Among 453 RT courses with complete data on devices and RT and a device control after RT, 14 (3.1%) device malfunctions occurred (Table 4).

**Table 4: Characteristics of radiotherapy (RT) courses resulting in device malfunction versus RT courses without malfunction, compared with RT courses without later device control, including RT courses with complete data on both the device and the RT (n = 658).**

	PM/ICD controlled after RT			PM/ICD never controlled after RT <sup>‡</sup>
	PM/ICD malfunction <sup>†</sup>	No PM/ICD malfunction <sup>†</sup>	Total <sup>†</sup>	
<b>N (%)</b>	14 (3.1)	439 (96.9)	453	205
<b>Age, years<sup>*</sup></b>	72.5 (68.4-77.8)	75.0 (68.9-80.3)	74.7 (68.9-80.3)	77.6 (70.6-82.7)
<b>Device class</b>				
PM: all classes, n (%)	10 (2.5)	384 (97.5)	394 (87.0)	188 (91.7)
Single chamber PM, n (%)	1 (1.1)	93 (98.9)	94 (20.8)	55 (26.8)
Dual chamber PM, n (%)	9 (3.3)	267 (96.7)	276 (60.9)	125 (61.0)
CRT-P, n (%)	0 (0.0)	24 (100.0)	24 (5.3)	8 (3.9)
ICD: all classes, n (%)	4 (6.8)	55 (93.2)	59 (13.0)	17 (8.3)
ICD (single and dual chamber), n (%)	2 (4.2)	46 (95.8)	48 (10.6)	11 (5.4)
CRT-D, n (%)	2 (18.2)	9 (81.8)	11 (2.4)	6 (2.9)
<b>Anatomical region</b>				
Head and neck, n (%)	1 (0.8)	118 (99.2)	119 (26.3)	60 (29.3)
Thorax, n (%)	4 (2.2)	174 (97.8)	178 (39.3)	59 (28.8)
Esophagus, n (%)	0 (0.0)	3 (100.0)	3 (0.7)	2 (1.0)
Abdomen and pelvis, n (%)	7 (5.6)	118 (94.4)	125 (27.6)	53 (25.9)
Spine, n (%)	1 (5.9)	16 (94.1)	17 (3.8)	20 (9.8)
Upper extremity, n (%)	0 (0.0)	3 (100.0)	3 (0.7)	4 (2.0)
Lower extremity, n (%)	1 (12.5)	7 (87.5)	8 (1.8)	7 (3.4)
<b>Cumulative tumor dose, Gy<sup>*</sup></b>	46.5 (20-70)	30 (20-52)	30 (20-52)	20 (20-25)
<b>Number of fractions<sup>*</sup></b>	24 (4-35)	9 (4-24)	9 (4-24)	4 (3-5)
<b>Fraction dose, Gy<sup>*</sup></b>	2 (2-5)	5 (2-5)	5 (2-5)	5 (5-5)
<b>Beam type</b>				
Photons, n (%)	14 (3.6)	371 (96.4)	385 (85.0)	191 (93.2)
Photons in MV range, n (%)	14 (4.0)	339 (96.0)	353 (77.9)	185 (90.2)
Photons in kV range, n (%)	0 (0.0)	32 (100.0)	32 (7.1)	6 (2.9)
Electrons, n (%)	0 (0.0)	68 (100.0)	68 (15.0)	14 (6.8)
<b>Beam energy, MV<sup>*</sup></b>	16.5 (15-18)	8 (6-15)	9 (6-15)	6 (6-15)
<b>Time from last RT fraction to following first PM/ICD control, days<sup>*</sup></b>	1.5 (0-15)	34 (3-158)	31 (2-145)	-
<b>Device age at RT, years<sup>*</sup></b>	2.4 (0.6-5.5)	2.8 (1.0-5.0)	2.8 (1.0-5.0)	2.6 (1.0-4.7)

CRT-D = cardiac resynchronization therapy implantable cardioverter defibrillator; CRT-P = cardiac resynchronization therapy pacemaker; Gy = gray; ICD = implantable cardioverter defibrillator; IQR = interquartile range; MV = megavolt; PM = pacemaker; RT = radiotherapy. <sup>\*</sup>Values are shown as median (interquartile range). <sup>†</sup>Percentages are row percentages to allow for comparison of RT courses with and without device malfunctions. <sup>‡</sup>Percentages are column percentages to allow for comparison of RT courses with and without later device control.

The median cumulative radiation dose in RT courses associated with device malfunctions was 46.5 Gy (IQR 20-70 Gy), and the median beam energy was 16.5 MV (IQR 15-18 MV) compared with 30 Gy (IQR 20-52 Gy) and 8 MV (IQR 6-15 MV) in treatments without malfunctions. No PM/ICD malfunctions were observed during electron RT or photon RT in kV range.

Seven (50%) malfunctions occurred during RT for tumors located in abdomen or pelvis while 4 (28.6%) involved RT of the thorax (Appendix C). The most frequent malfunction was electrical reset or transient minor software errors, occurring in 11 (78.6%) of affected devices. In these cases, the device could be reprogrammed by the staff in the clinic, whereas in 2 (14.3%) cases, assistance from the manufacturer was necessary in order to update the software of the device. In one PM, the only deviation was an increase in atrial pacing threshold from 1.25V to 2.75V. Importantly, no malfunctions required device explantation or lead revision. In terms of other adverse clinical consequences from the malfunctions, one patient experienced diaphragmatic pacing after the reset of a single-chamber PM which ceased after reprogramming of the device.

### **Predictors of device malfunctions**

Crude logistic regression analysis showed that PM/ICD malfunctions were associated with beam energy  $\geq$ 15 MV (OR 5.73, 95% CI

1.58-20.76) and location of the tumor below the diaphragm (OR 4.31, 95% CI 1.42-13.12) (Table 5).

**Table 5: Crude and adjusted logistic regression analysis of factors associated with PM/ICD malfunctions during radiotherapy courses (n=453).**

Variable	Crude OR (95% CI)	Adjusted OR (95% CI)
Device class: ICD, CRT-D vs. PM, CRT-P	2.78 (0.84-9.15)	2.93 (0.87-9.86)
Anatomical location of the tumor: below diaphragm vs. above diaphragm	4.31 (1.42-13.12)	2.27 (0.65-7.95)
Cumulative tumor dose (10 Gy increment)	1.20 (0.95-1.52)	1.13 (0.89-1.44)
Fraction dose (1 Gy increment)	0.83 (0.64-1.08)	0.92 (0.71-1.18)
Beam energy ( $\geq$ 15 MV vs. <15 MV)	5.73 (1.58-20.76)	1.0 (reference)

CI = confidence interval; CRT-D = cardiac resynchronization therapy defibrillator; CRT-P = cardiac resynchronization therapy pacemaker; Gy = gray; ICD = implantable cardioverter defibrillator; MV = megavolt; OR = odds ratio; PM = pacemaker.

The median beam energy used for treatment of tumors located above diaphragm (6 MV, IQR 6-10 MV) was lower than the median beam energy used below diaphragm (15 MV, IQR 15-18 MV). After

adjustment for beam energy, the latter remained the only significant predictor of device malfunction.

Interestingly, no significant correlation was detected between device malfunctions and cumulative radiation dose to the tumor

(adjusted OR 1.13, 95% CI 0.89-1.44) or fraction dose (adjusted OR 0.92, 95% CI 0.71-1.18). Although insignificant, device malfunctions were more frequent in ICDs (6.8%) compared with PMs (2.5%), adjusted OR 2.93, 95% CI 0.87-9.86.

## Discussion

### Prevalence of patients with PM/ICD undergoing RT

To date, no large-scale studies on the size of the population of PM/ICD patients undergoing RT have been performed. In 1991, Rodriguez et al. estimated, based on clinical experience, that 0.4–0.5% patients undergoing RT had a PM/ICD.<sup>62</sup> In 2000, Tsekos et al. reported that they saw at least 10 PM patients per year at their RT department at a university hospital.<sup>63</sup> In 2004, Solan et al. also stated that any busy RT department would see several PM/ICD patients per year.<sup>39</sup> More recently, Gossman et al. estimated, based on survey data, that the median proportion of RT patients requiring a change in the RT approach due to PM/ICD was 0.8%, and expected the total proportion of RT patients having a PM/ICD to be higher.<sup>64</sup> In the present study, the annual rate of RT courses in Western Denmark reached 4.33 per 100,000 person-years in 2012. Our findings also demonstrate a sharp increase in the annual rate by 199% during the ten-year period from 2003 to 2012.

### Probability and consequences of device malfunction

Since the publication of the AAPM TG-34 guidelines,<sup>38</sup> numerous in vitro experiments on the effects of RT on both PMs and ICDs have been reported (Table 6 and Table 7).

Variations exist in the reported rates of device malfunctions. Differences are also present in terms of radiation doses at these episodes. Hurkmans et al. irradiated 19 PMs with 6 MV photons to a dose of up to 120–130 Gy.<sup>65</sup> Points of first malfunctions varied from 10 to 120 Gy, while five devices were irradiated with full dose without any adverse effects. In the largest in vitro study so far, Mouton et al. analyzed the effects of direct irradiation with 18 MV photons on 96 PMs, reaching doses up to 200 Gy per device.<sup>66</sup> The authors classified the observed malfunctions, with three of the classes being potentially lethal: amplitude change of >10%, pauses in electrical signal of >10s, and permanent silence. These malfunctions were observed at doses starting from 2 Gy, 0.15 Gy, and 0.5 Gy, with mean doses at their occurrence of 56 Gy, 17.4 Gy, and 71 Gy, respectively. The majority (70%) of malfunctions were observed at a dose rate of 8 Gy/min, with no malfunctions occurring at or below the dose rate of 0.2 Gy/min, thus suggesting that dose rate is a potential risk factor in inducing PM/ICD malfunctions.

In terms of in vitro studies on ICDs, Hurkmans et al. exposed 11 ICDs to 6 MV photons, reaching a dose of 120 Gy.<sup>67</sup> Failures were observed in all devices with doses at first malfunction varying from 0.5 Gy to 120 Gy. Of note, sensing interference was observed in all these devices, potentially leading to shock therapy in case of a clinical situation. At the same time, exposing 20 ICDs (including 8 CRT-Ds) to 4 Gy of scattered radiation from a 6 MV photon beam, Kapa et al. reported no device malfunctions.<sup>68</sup> Hashii et al. compared the effects of different beam energies on four ICDs from one manufacturer in vitro.<sup>69</sup> An accelerated course of RT was imitated

with the ICDs outside the RT field. Interrogating the devices every 10–50 Gy, the authors found that scattered radiation with 18 MV photons led to a greater number of software errors, compared with 10 MV. In addition, dosimetry showed that during high-energy beaming, more neutrons are generated. The authors expressed their concern regarding the risk of ICD malfunction during high-energy photon irradiations even with devices located at a distance from the RT field.

Along with in vitro experiments, a number of case reports and case series have been published as well (Table 8 and Table 9).

In terms of malfunction rate, the proportion of 3.1% in Study III is in line with a retrospective study by Gossman et al., who reported device malfunctions in four (3.6%) out of 112 RT courses in PM/ICD patients.<sup>64</sup> With a median time from RT to the first device control of 31 days in 453 devices, the follow-up period in Study III corresponds to 14043 device-days, suggesting approximately 1 malfunction per 1000 device-days. Hence, this rate seems to be higher than the rate of spontaneous minor software errors of approximately 1 event per 13000 device-days, as observed by Bradley and Normand in ICDs.<sup>70</sup> The rate of device malfunctions in Study III might even have been higher, since 24 devices were relocated before the RT. The proportion of device malfunctions is comparable with the findings by Makkar et al. who observed two partial resets of ICDs, exposed to 1.23 Gy and 0.04 Gy, respectively, (2.9% of all devices) among 69 patients [50 (72.5%) PMs, 19 ICDs].<sup>50</sup> The remaining intact devices received doses ranging from 0.9 to 5.057 Gy with 6–16 MV photons with or without electrons. In a study of 62 patients [60 (96.8%) PMs, 2 ICDs], Soejima et al. observed one PM (1.6% of all devices) that reset during RT for prostate cancer with 15 MV photons.<sup>49</sup> The remaining 61 devices were exposed to radiation doses reaching as high as 20.7 Gy (>2 Gy in six cases), thus indicating that PMs/ICDs may develop malfunctions outside the RT beam and the occurrence of malfunctions may not be related to radiation dose.

Regarding proton beam RT, Oshiro et al. observed malfunctions in two (25%) out of eight PM patients with devices located outside the RT field.<sup>71</sup> In a study of 42 patients (28 PMs, 14 ICDs) undergoing proton RT with varying doses to the device, Gomez et al. observed five resets during thoracic RT in four (9.5%) patients and an expected ERI in one patient.<sup>72</sup> Hence, these limited clinical data suggest that probability of PM/ICD malfunction during proton RT might be higher compared with photon RT. However, with only five devices described in an in vitro setting of proton RT,<sup>71,73</sup> probably no direct firm conclusions can be drawn to support this statement. Clinical data on neutron beam RT in PM/ICD patients are even scarcer. To the best of our knowledge, there has only been published one case report on uncontrollable pacing occurring during neutron RT to thyroid.<sup>74</sup> Although limited to 16 PMs, in vitro data suggest that PMs/ICDs might be even more sensitive to neutron RT compared with other beam types.<sup>75,76</sup>

**Table 6: Studies examining the effects of radiotherapy on pacemakers in vitro, published in English since 1994.**

Year	Author	n	Beam type	Beam energy, MV	Maximal generator dose, Gy	Main results
1994	Souliman <sup>77</sup>	18	Photons	8	70	Complete failure at 16.8-70 Gy in 11 PMs. No effects from EMI alone.
1999	Mouton <sup>78</sup>	42	Photons (n=19); Co-60 (n=23)	4 (LINAC); 1.17-1.33 (Co-60 source)	140	No malfunction at therapeutic doses (n=15); frequency modifications (n=9) starting at 2 Gy; deprogramming and modification in battery characteristics (n=11) starting at 4 Gy; destruction of the PM (n=7) at 44-77 Gy.
2002	Mouton <sup>66</sup>	96	Photons	18	200	Amplitude change >10%: 38 PMs at 2-130 Gy; silence >10 s: 35 PMs at 0.15-74 Gy; permanent silence: 12 PMs at 0.5-170 Gy.
2005	Hurkmans <sup>65</sup>	19	Photons	6	120; 130 (n=2)	5 PMs: no malfunction; 7 PMs: no output at 80-130 Gy; 5 PMs: ERI at 120-130 Gy; 2 PMs: no communication at 20-130 Gy; 8 PMs: inhibition during direct irradiations.
2008	Oshiro <sup>71</sup>	1	Protons	250	NA (35 Gy to the lead)	No malfunctions.
2011	Koivunoro <sup>75</sup>	2	Epithermal neutrons	0.414 eV < E < 9.12 keV	1.2; 2.0	Reset due to memory changes in microprocessor; many severe bit flips and loss of telemetry. Both PMs got activated.
2012	Trigano <sup>76</sup>	14	Neutrons	3-50 (peak around 20)	NA	Electrical reset in 6 cases.
2014	Zaremba <sup>79</sup> (Study I)	10	Photons	6; 18	150	6 MV group: one episode of malfunction at 150 Gy; 18 MV group: 14 episodes of malfunction starting at 30 Gy.

Co = cobalt; E = energy; EMI = electromagnetic interference; ERI = elective replacement indicator; eV = electronvolt; Gy = gray; LINAC = linear accelerator; MV = megavolt; PM = pacemaker.

**Table 7: Studies examining the effects of radiotherapy on implantable cardioverter defibrillators in vitro, published in English since 1994.**

Year	Author	n	Beam type	Beam energy	Dose, Gy	Main results
2002	Hoecht <sup>80</sup>	5	NA	NA	Scatter; >50	No effects from EMI; scatter radiation: 1 fallback; direct exposure: malfunctions at >50 Gy, unspecified.
2005	Hurkmans <sup>67</sup>	11	Photons	6	120	Sensing interference in all ICDs, which would have resulted in shock in 4 ICDs. Failure of all devices at 0.5-120 Gy. Complete loss of function at 0.5-1.5 Gy in 4 ICDs.
2008	Kapa <sup>68</sup>	20 (8 of them CRT-D)	Photons	6	4	No malfunctions.
2012	Hashimoto <sup>73</sup>	4	Protons	200	Scatter	1 power-on reset per approximately 50 Gy, 1 soft error per approximately 15 Gy. No permanent malfunctions.
2013	Hashii <sup>69</sup>	10	Photons	10; 18	Scatter	More soft errors during irradiation with 18 MV photons compared with 10 MV photons. No hard errors or permanent malfunctions.
2014	Mollerus <sup>81</sup>	8	Photons	6	131.11	4 contemporary devices remained functional after 131.11 Gy despite minor memory faults in 3 of them; 4 legacy devices failed to deliver shock therapy after 41.11 Gy and had changes in lead impedance.
2014	Zaremba <sup>79</sup> (Study I)	2	Photons	6; 18	150	6 MV: no malfunctions; 18 MV: no malfunctions, except loss of patient data after 44 Gy.

CRT-D = cardiac resynchronization therapy defibrillator; EMI = electromagnetic interference; Gy = gray; ICD = implantable cardioverter defibrillator; MV = megavolt.

Concerning severity and clinical consequences of RT-induced PM/ICD malfunctions, removal of the device due to malfunction has been described in five published cases (two PMs, three ICDs) over the last two decades.<sup>74,80,82-84</sup> Of note, the generators were located outside the direct RT field in all of these patients, and no lethal malfunctions have been reported. In a number of other published PM/ICD malfunction cases, the devices were either successfully reprogrammed, or the effects were transient, permitting completion of the RT.<sup>47-51,63,71,72,85</sup> Similarly, no patients in Study III required device explantation, and the majority of the observed malfunctions were resets or transient software errors. In Study I, ERI was first recorded after 150 Gy in two devices, i.e. at supratherapeutic doses. In Study II, one episode of reset was observed, after which the device could be reprogrammed. Meanwhile, despite the fact that a PM/ICD after undergoing a reset is capable of delivering a basic treatment to the patient, it may be inappropriate for the patient to have the device functioning in this mode for a longer period due to deprivation of atrioventricular synchrony, rate-adaptive pacing, or biventricular pacing.<sup>86-88</sup> In

addition, one of the patients in Study III experienced diaphragmatic pacing. Regarding PM-dependent patients, while permanent or temporary loss of pacing during RT has been described in *in vitro* experiments,<sup>65-67,77</sup> this phenomenon to date has not been reported in a clinical setting. Concerns have been raised regarding noise oversense during RT, potentially leading to inappropriate shock therapies in ICD patients, as demonstrated *in vitro* by Hurkmans et al.<sup>67</sup> So far, inappropriate tachycardia sensing without shock therapy has only been reported in one ICD patient undergoing RT to femur in a series (n=15) by Elders et al.<sup>51</sup>

Although RT may affect the function of the generator of PM/ICD, the PM/ICD leads are considered to be resistant to these effects.<sup>41</sup> Only one case report presented a course of RT where malfunction of an ICD shock lead was suspected due to direct irradiation, leading to device reimplantation.<sup>82</sup> In Study III, an increase in pacing threshold was observed in two patients (0.4% of RT courses with subsequent device control), still none of the episodes required intervention.

**Table 8: Studies examining the effects of radiotherapy on pacemakers *in vivo*, published in English since 1994.**

Year	Author	Devices, n	Tumor	Beam type	Beam energy	Tumor dose, Gy	Device dose, Gy	Outcome	Clinical consequences
1994	Raitt <sup>74</sup>	1	Thyroid	Neutrons	NA	4.8	0.9	Uncontrollable ventricular pacing at 180 bpm. Corruption of the programming code of the microprocessor.	Replacement of the device.
2000	Tsekos <sup>63</sup>	1	Right lower arm and axilla	NA	NA	50.4	50.4 (in the field)	Decrease in magnet rate, returning to normal 4 months later.	None
2001	Nibhanupudy <sup>89</sup>	1	Left breast and supraclavicular field	Photons	6	50.4	1.82	No malfunctions.	-
2006	Ampil <sup>90</sup>	3	Lung	Photons	NA	20-60	NA	No malfunctions.	-
2006	Mitra <sup>91</sup>	1	Right lung and mediastinum	Photons	NA	40	1.66	No malfunctions.	-
2008	Kapa <sup>68</sup>	8 (1 of them CRT-P)	Head and neck; thorax	Photons	6	30-69.96	NA (outside field)	No malfunctions.	-
2008	Oshiro <sup>71</sup>	8	Liver; lung (n=1)	Protons	155-250	36.3-77	NA (outside field)	Reverting to "safety backup program" after 46 Gy (n=1); 2 episodes (after 23 and 26 Gy) of changing in pacing frequency (n=1).	Both devices successfully reprogrammed.
2008	Zweng <sup>84</sup>	1	Esophagus	Photons	NA	30	0.11	Runaway PM. Change from DDD to AAI with a fixed rate of 185 bpm. Corruption of the software.	Circulatory collapse. Replacement of the device.
2010	Ferrara <sup>92</sup>	37	Various	Photons and electrons (95.6%); Co-60 (4.4%)	6; 18 (59%)	8-79.2	>2 (n=5); <2 (n=32)	No malfunctions.	-
2011	Croshaw <sup>53</sup>	3	Breast	Photons	6	38.5	0.23-0.73	No malfunctions.	-
2011	Dasgupta <sup>93</sup>	1	Heart	Photons	6	37.5	0.37	Transient ventricular undersensing.	Device successfully reprogrammed.

Year	Author	Devices, n	Tumor	Beam type	Beam energy	Tumor dose, Gy	Device dose, Gy	Outcome	Clinical consequences
2011	Soejima <sup>49</sup>	60	Various	NA	NA	20-74 (range in the whole study)	20.69 in 1 patient, otherwise not exceeding 4.78	1 CRT-P was found initialized at 46 Gy and 56 Gy (treated with 74 Gy 15 MV photons for prostate cancer).	Device successfully reprogrammed.
2011	Wadasadawala <sup>94</sup>	8	Head and neck; breast; lung	Co-60 (n=3); photons (n=5)	Photons: 6 (n=3); 15 (n=2)	45-70	0.14-60 (including PM leads)	No malfunctions.	-
2012	Kesek <sup>45</sup>	1	Lung	Photons	6	80	48	No malfunctions.	-
2012	Kirova <sup>95</sup>	1	Thoracic spine	Photons	20	30	0.3 (leads irradiated directly)	No malfunctions.	-
2012	Makkar <sup>50</sup>	50	Various	Photons; photons and electrons	Photons: 6 (n=26); 16 (n=24). Both with or without electrons (6-16 MeV)	NA	0.844 +/- 0.997	No malfunctions.	-
2013	Gomez <sup>72</sup>	28	Various	Protons	NA	46.8-87.5	0.13-21 (range in the whole study)	Reset in 2 PMs, both treated to thorax, tumor doses at the episodes 4 Gy and 16.2 Gy, respectively. Distance from device to RT field 3 cm and 0.9 cm, respectively.	Both devices successfully reprogrammed.
2014	Ampil <sup>96</sup>	2	Head and neck	Photons	6	NA	NA	No malfunctions.	-
2014	Gossman <sup>64</sup>	67 (out of 107 devices)	Various	Presumably photons	Various	NA	<2 in 85%, >2 in 15%, not exceeding 6.5 (the whole study)	Failure at 0.3 Gy (n=1); increase in sensor rate during RT (n=1); irregular heartbeat leading to reprogramming (n=1); twinging in the chest wall resulting in respiratory arrest (n=1) (the whole study).	Not specified in more detail
2015	Zaremba <sup>97</sup> (Study III)	487 (25 CRT-P)	Various (thorax 36% in the whole study)	MV photons, kV photons, electrons	9 (IQR 6-15) in all interrogated devices in the study	Various	NA	Reset or deprogramming (n=9); increase in atrial pacing threshold from 1.25 to 2.75 V (n=1) out of 394 interrogated PMs.	Devices successfully reprogrammed. No device replacements.

AAI = atrial pacing and sensing; bpm = beats per minute; Co = cobalt; CRT-P = cardiac resynchronization therapy pacemaker; DDD = dual chamber pacing and sensing; Gy = gray; IQR = interquartile range; kV = kilovolt; MeV = megaelectronvolt; MV = megavolt; PM = pacemaker.

**Table 9: Studies examining the effects of radiotherapy on implantable cardioverter defibrillators *in vivo*, published in English since 1994.**

Year	Author	Devices, n	Tumor	Beam type	Maximal beam energy	Tumor dose, Gy	Device dose, Gy	Outcome	Clinical consequences
2002	Hoechtl <sup>80</sup>	4 (3 patients)	NA	NA	NA	NA	NA	2 ICDs of the same model in the same patient fell into fall back mode at <0.5 Gy to the ICD (RT to pelvis).	The device was replaced due to the first episode.
2004	John <sup>82</sup>	1	Left breast	NA	NA	50	Leads: 50, partial exposure of the generator	Shock coil failure due to structural damage during RT was suspected (shock impedance >125 ohms).	A new system was implanted.
2004	Thomas <sup>85</sup>	1	Right lung	Photons	18	56	NA (outside field)	Electrical reset.	Unspecified (asymptomatic).
2007	Nemec <sup>83</sup>	1	Left lung	NA	NA	59.4	NA (outside field)	Rapid pacing triggering polymorphic VT during the 3rd fraction of 1.8 Gy.	Collapse requiring resuscitation. Device removal afterwards.
2007	Sepe <sup>98</sup>	1	Larynx	Photons	6	60	2.5	No malfunctions.	-
2008	Kapa <sup>68</sup>	5	Various	Photons	6	18-56	NA (outside field)	No malfunctions.	-
2008	Lau <sup>47</sup>	1	Prostate	Photons	23	74	0.004	Resets during 2nd and 9th fractions of 2 Gy.	RT completed without other events. Normal ICD parameters afterwards.
2009	Gelblum <sup>48</sup>	1+33	Various	Photons; electrons (n=1); photons and electrons (n=1)	6 (photons); 6 MeV (electrons); 6 MV and 9 MeV (photons and electrons); 15 (photons, n=2)	6-86.4	0.01-2.99	Reset in 2 patients treated with 15 MV photons, outside RT field.	Devices successfully reprogrammed.
2010	Ferrara <sup>92</sup>	8	Various	Photons and electrons (95.6%); Co-60 (4.4%)	6; 18 (59%)	8-79.2	>1 (n=2), <1 (n=6)	No malfunctions.	-
2011	Croshaw <sup>53</sup>	2	Breast	Photons	6	38.5	1.01, 1.68	No malfunctions.	-
2011	Soejima <sup>49</sup>	2	Various	NA	NA	20-74 (range in the whole study)	NA	No malfunctions.	-
2012	Makkar <sup>50</sup>	19	Various	Photons; photons and electrons	6; 16. Both with or without electrons (6-16 MeV)	NA	0.921+/- 0.726	Partial resets in 2 devices after 1.23 Gy and 0.04 Gy 16 MV photons to the ICD, respectively.	RT completed successfully in both cases.
2013	Dell'Oca <sup>99</sup>	1	Mediastinum	Photons	6	64	<5	No malfunctions.	-
2013	Elders <sup>51</sup>	15 (17 RT courses)	Various	Photons; photons and electrons (n=1)	6-18	16-70	<1	6 malfunctions in 5 RT courses at 10 and 18 MV: invalid data retrieval (n=2), reset (n=1), inappropriate tachycardia sensing (n=1), reset and trend data error 9 months after the reset (n=1).	RT completed successfully in all patients.
2013	Gomez <sup>72</sup>	14	Various	Protons	NA	46.8-87.5 (range in the whole study)	0.13-21 (range in the whole study)	Reset after 40 Gy to the tumor (n=1); resets after 32.5 Gy and 47.5 Gy to the tumor (n=1).	Both devices successfully reprogrammed.

Year	Author	Devices, n	Tumor	Beam type	Maximal beam energy	Tumor dose, Gy	Device dose, Gy	Outcome	Clinical consequences
						study)		Distance from device to RT field 5 cm and 8 cm, respectively.	
2013	Zaremba <sup>100</sup> (Study II)	5	Thorax	Photons	6; 18	37	37	Converting to back-up mode at midnight (n=1).	None (animal study; all devices explanted after the irradiations).
2014	Ahmed <sup>46</sup>	1	Lung	Photons	15	69.6	52.4	No malfunctions.	-
2014	Gossman <sup>64</sup>	40 (out of 107 devices)	Various	Presumably photons	Various	NA	<2 in 85%, >2 in 15%, not exceeding 6.5 (the whole study)	Failure at 0.3 Gy (n=1); increase in sensor rate during RT (n=1); irregular heartbeat leading to reprogramming (n=1); twinging in the chest wall resulting in respiratory arrest (n=1) (the whole study).	Not specified in more detail.
2015	Zaremba <sup>97</sup> (Study III)	73 (19 CRT-D)	Various (thorax 36% in the whole study)	MV photons, kV photons, electrons	9 (IQR 6-15) in all interrogated devices in the study	Various	NA	Reset (n=3), reset and increase in pacing threshold (n=1) out of 59 interrogated ICDs.	Devices successfully reprogrammed.

Co = cobalt; CRT-D = cardiac resynchronization therapy defibrillator; Gy = gray; ICD = implantable cardioverter defibrillator; IQR = interquartile range; kV = kilovolt; MeV = megaelectronvolt; MV = megavolt; RT = radiotherapy.

## Predictors of device malfunction

The PM/ICD malfunctions seem to occur unpredictably as far as radiation dose is concerned both *in vivo*<sup>45-49,51,85,97</sup> and *in vitro*.<sup>65-67</sup> Besides the effects of radiation dose, RT beam energy has emerged as an important subject in several studies.<sup>48,51,69</sup> We have clearly demonstrated this in Study I where a nine times higher risk of malfunction was observed in devices exposed to 18 MV photons compared with 6 MV. This was also proven to be the case in Study III, as beam energy  $\geq$ 15 MV was the strongest predictor of PM/ICD malfunctions.

In Study III, there was also a trend of increased risk of malfunctions in ICDs compared to PMs, which is in line with previous reports that ICDs might be more sensitive to ionizing radiation than PMs.<sup>39,101</sup> Unexpectedly, we observed that device malfunctions were associated with location of the RT fields below the diaphragm. This might be because beam energies used below the diaphragm were higher than those used above. Moreover, after adjustment for beam energy, the anatomical location was not significantly associated with malfunction.

The correlation between damaging effects of RT on PMs/ICDs and dose rate was suggested by Mouton et al.<sup>66</sup> Although not addressed directly in later studies, this concern was discussed by Hurkmans et al., who underlined that the dose rate to a PM/ICD located outside a direct beam with a dose rate of 1-10 Gy/min at isocenter is expected to be  $<1$  Gy/min, hence not posing a considerable threat.<sup>41</sup> In addition, the effects of dose rate on the device seem to be transient and reversible.<sup>62</sup> The results of Study II appear in line with these considerations, as no ICD dysfunction was observed during the irradiation procedures.

Contemporary PMs/ICDs are relatively well protected against the effects of electromagnetic interference (EMI).<sup>13</sup> Besides ionizing radiation, LINACs include several potential sources of EMI, such as couch drive motors, x-ray transformers, waveguides, power

supplies, klystrons, or magnetrons,<sup>13</sup> potentially leading to pacing inhibition, fixed rate pacing, or reprogramming.<sup>41</sup> However, these effects are usually transient and are observed when the machine is turned on or off.<sup>102</sup> In addition, modern LINACs are sufficiently shielded, and EMI typically does not pose any threats to the function of PMs/ICDs.<sup>41,103</sup> This is supported by the findings of this project, especially Study III, where no events of symptomatic inhibition or rapid pacing were observed during RT.

It is unclear whether the different PM/ICD brands differ in risk of RT-induced malfunctions. The spectrum of the PM/ICD manufacturers represented in this project was relatively broad, and some were represented by few devices. Therefore, we were unable to draw any conclusions on this aspect.

## Possible mechanisms

According to the literature, the manifestations of PM/ICD malfunctions can in broad terms be divided into three groups: 1) transient effects due to interference, occurring during the irradiation only; 2) reverting to backup settings (reset), recoverable after reprogramming the device; and 3) permanent damage to the device.<sup>77</sup> Hashii et al. grouped PM/ICD errors during RT into hard errors and soft errors, with hard errors representing damage to the hardware of the device, while soft errors consist of software alterations.<sup>69</sup> The latter group can be divided into severe reset requiring reprogramming of the device, moderate reset not requiring correction by the programmer, and minor error not detectable at interrogation and only recorded in the data log of the device.

The majority of all observed malfunctions in Studies I-III were electrical resets requiring reprogramming. Electrical reset, power-on-reset or fallback to back-up mode, is a relatively well known phenomenon in PM/ICD technology.<sup>76,86</sup> It is a rather adequate behavior of the device, indicating that an error has been detected.<sup>88,104</sup> In such case, the device switches from a software-

controlled mode to a basic hardware-driven condition.<sup>76</sup> The purpose of this reversal is to ascertain basic pacing and an effective shock therapy in case of ICD.<sup>87,88,104</sup> Electrical reset seems to pose no harm to the device itself, as the errors occur at software level without physical damage to the components, and the device can then usually be reprogrammed at interrogation.<sup>48,49,71,88,105</sup> As demonstrated in Study II and Study III, in some of these cases, assistance by a manufacturer technician might be necessary.<sup>48</sup>

Production of high energy photons in a LINAC is accompanied by a generation of secondary neutrons which have a notoriously high capability of ionizing components of CMOS devices and inducing errors and resets.<sup>75,76,106</sup> This underpins our findings of PMs/ICDs failing at high beam energies. Moreover, neutron production during electron RT is 5% and 20% of that during photon RT at 15 MV and 25 MV nominal energies, respectively.<sup>107,108</sup> Thus, these aspects are in line with the fact that no device malfunctions were observed during electron or kV photon RT in Study III.

In contrast to soft errors, hard errors consist of structural damage to the device.<sup>73</sup> There have been only few published cases describing other device damages than electrical resets in relation to RT.<sup>51,82,84</sup> In Study III, an increase in pacing threshold occurred in 2 (0.4%) out of 453. Both patients were treated with 6 MV photons to the thorax and abdomen/pelvis, respectively. Still, as the negative effect of RT on the devices cannot be ruled out, a control group without exposure to RT would probably be needed to draw any reliable conclusions on this aspect. No deviations in lead parameters were observed in Study II. Hard errors seem to be observed more often in vitro where the devices typically are exposed to supratherapeutic doses of radiation.<sup>65-67</sup> Hence, every device does have a threshold in radiation dose ultimately damaging the circuit.<sup>109</sup> On the contrary, soft errors appear to be more stochastic and less predictable.<sup>48</sup>

## Safety measures

One of the dilemmas in treating PM/ICD patients undergoing RT arises when a decision has to be made whether to surgically remove the device from the vicinity of the RT field in order to avoid malfunction of the PM/ICD.<sup>45,46,110</sup> Although the removal permits reduction of the radiation dose to the device, every surgical intervention to the PM/ICD exposes the patient to the risk of complications and is associated with increased healthcare costs.<sup>42-44,59,111,112</sup> In terms of device relocation prior to RT in practice, it has not been performed in the majority of the published patient series.<sup>48,49,51,64,72,92</sup> In other series, the rate of device relocation varies between 7 and 31% of RT courses,<sup>50,68</sup> which is higher than 3.5% observed in Study III. However, this comparison may be limited due to small sample sizes in the previous studies. The rate of relocation-related complications in Study III was in line with general practice.<sup>59,112</sup>

It is generally recommended to estimate the radiation dose to the PM/ICD in a patient scheduled for RT.<sup>36,38,41,103,113</sup> While the devices in Studies I and II received the entire administered radiation dose, the device doses were in most cases not available in Study III due to the retrospective design of the study. PM/ICD dose calculations are reported in most prospective studies,<sup>48-51,53,71,92,94</sup> but only in less than a half of the retrospective case series.<sup>64,68,72,90,96</sup> Some authors recommend to supplement the estimated values from treatment simulation with in vivo dosimetry,<sup>113</sup> while others suggest that dose estimation only has to be accurate

enough to determine in which of the three risk categories of device malfunction the patient will fall, especially as in most cases the dose to the device will be low (<2 Gy).<sup>41</sup> In most larger prospective case series, the patients were evaluated in the PM/ICD clinic according to a predefined protocol, including controls before, during, and after RT. However, as demonstrated in Study III, only about a third of these patients are referred systematically to cardiologists in routine clinical practice.

Inactivation of antitachycardia therapies either by reprogramming or application of a magnet to ICDs is recommended before RT in several publications.<sup>39,41,113</sup> The purpose of magnet ( $\geq 90$  gauss) application to ICDs is to prevent oversense from EMI, while bradycardia pacing mode is generally unaffected, contrary to PMs which switch to asynchronous pacing mode under application of a magnet.<sup>114</sup> Although this precaution is supported by some in vitro data indicating oversense in ICDs during RT,<sup>67</sup> to the best of our knowledge, there have been no published data of inappropriate shocks during RT in ICD patients. In Study III, application of a magnet was documented in 10.8% of RT courses with ICDs, and no inappropriate shocks were described in relation to RT in the patients with ICDs. Gossman et al. reported that in 3.6% of cases a magnet was applied during RT, not specifying device types in these patients.<sup>64</sup>

## Current safety recommendations

Several sets of practical recommendations on safe RT in PM/ICD patients have been issued since the publication of AAPM TG-34 guidelines (Table 10).<sup>36,38-41,56,103,113,115</sup>

After identifying that the patient scheduled for RT has a PM/ICD, it is usually recommended to estimate the cumulative radiation dose to the device, where it is essential to take the maximal dose to any part of the generator into account.<sup>40,56,113</sup> Also, all shielding of the device should originate from the LINAC rather than additional shielding with lead.<sup>40,103,113</sup> Although it has previously been recommended to shield the device with a lead alloy during RT in order to keep the dose as low as possible,<sup>14,56,103,116,117</sup> this would have only a limited effect due to scatter within the patient if the device is outside the RT field.<sup>103</sup> In case of direct radiation, a simple lead apron would be also ineffective, as 90% attenuation of 6 MV photon beam can only be achieved by >5 cm thick lead shield.<sup>48</sup> In addition, some authors recommend keeping the PM/ICD at least 3-5 cm from the RT field.<sup>36,40,103</sup>

The Heart Rhythm Society/American Society of Anesthesiologists Expert Consensus Statement mentions that usage of high-energy photon beams might lead to device malfunctions.<sup>115</sup> Similarly, recent multidisciplinary Dutch guidelines warn of using >10 MV photons in PM/ICD patients due to high risk of device malfunctions.<sup>41</sup> On the other hand, the radiation dose still seems to be underscored as the main factor in some recently published review articles, leaving out the significance of beam energy.<sup>118,119</sup>

The device evaluations performed in Study III seem to be in discrepancy with recent recommendations, as just 34.4% RT courses with  $\geq 15$  MV photons led to a subsequent device evaluation in opposite to the advice to control all devices after the RT course.<sup>41,56,113</sup> This proportion was higher (42.0%) when  $\geq 15$  MV photon beams were given above the diaphragm.

There have been propositions to refine the prediction of the clinical consequences of device malfunction during RT by classifying the risk to the patient into low, medium, and high (Table

10).<sup>41,56</sup> In low risk patients, it is suggested that audiovisual assessment of the patient during RT fractions should be sufficient, and that the completed RT course should be followed by a device check, whereas ICDs should be interrogated weekly.<sup>41</sup> Medium and high-risk patients should only be treated at institutions where trained staff with cardiology expertise and access to external pacing are available.<sup>41,56</sup> Weekly and daily device controls are suggested in medium and high-risk patients, respectively.<sup>41</sup> Heart rhythm monitoring during every RT fraction is advocated, especially in high risk patients.<sup>40,41,103,113</sup> There seems to be no solid evidence on late device malfunctions in case a PM/ICD has exhibited normal function during and immediately after the RT; however, performing additional checks during a period of up to six months after RT has been recommended.<sup>41</sup>

In terms of safety measures in ICD patients, most authors recommend to inactivate all antitachycardia functions of the ICDs at least during the first RT fractions, either by reprogramming the device or by application of a magnet to the ICD.<sup>39-41,113</sup> However, this rate was relatively low in Study III (application of a magnet in 10.8%). In addition, heart rhythm monitoring is recommended as soon as antitachycardia therapies are switched off.<sup>40</sup> Provided there is no oversense recorded at interrogations, the therapies

might be considered to be reactivated during the remaining fractions.<sup>41</sup> Taking the radiation dose and PM-dependency into account, surgical relocation of the device might be considered in high risk patients, e.g. in case of ipsilateral breast or lung tumor in a PM-dependent patient.<sup>39,41,56</sup> Usage of a backup temporary PM during RT has been advocated,<sup>56</sup> although not used in Study III and seldom described in the literature.<sup>120</sup>

So far, no universal PM/ICD manufacturer-specific guidelines on safe RT have been published, besides a couple of references to recommendations from the major manufacturers.<sup>39,40</sup> One of the reasons is the fact that not all technical data are available in the public domain, and also, the devices are continuously improved.<sup>121</sup> According to Medtronic, their devices should be able to tolerate cumulative doses of 1-5 Gy depending on the model (Table 11).<sup>122</sup> Boston Scientific and St. Jude Medical cannot rule out that their devices might fail even at scatter radiation, permitting no dose limit to be regarded as safe.<sup>123,124</sup> In addition, Biotronik and Medtronic highlight the importance of beam energy due to the damaging effects of secondary neutrons with a recommendation to limit photon energy to 10 MV.<sup>122,125</sup> Despite some variations, all major manufacturers caution against the PM/ICD being located in the RT field.<sup>122-125</sup>

**Table 10: Brief overview of recommendations on safe radiotherapy in patients with pacemakers and implantable cardioverter defibrillators published during the last decade.**

Recommendation	Sundar 2005 <sup>56</sup> (PM only)	Tondato 2009 <sup>103</sup>	Hudson 2010 <sup>40</sup>	Hurkmans 2012 <sup>41</sup>	Langer 2012 <sup>113</sup> (ICD only)
<b>Method of device dose estimation</b>	RT planning calculations or dosimetry	Not specified	RT planning calculations and dosimetry	RT planning calculations; dosimetry can be considered	RT planning calculations and dosimetry
<b>Maximal PM dose</b>	Low risk if not PM-dependent and <2 Gy; medium risk if PM-dependent and <2 Gy; high risk if PM-dependent and >2 Gy or if the PM in the RT field	2 Gy	<2 Gy (never >5 Gy)	Low risk if not PM-dependent and <2 Gy; medium risk if PM-dependent and <2 Gy or if 2-10 Gy; high-risk if >10 Gy	-
<b>Maximal ICD dose</b>	-	2 Gy	<1 Gy	Low risk if not PM-dependent and <2 Gy; medium risk if PM-dependent and <2 Gy or if 2-10 Gy; high-risk if >10 Gy	As recommended by the ICD producer
<b>Limit beam energy</b>	No	No	No	≤10 MV	No
<b>Device checks</b>					
<b>Before the RT course</b>	Yes	Yes	Not specified	Yes if not evaluated within the past 3 months	Yes
<b>During the RT course</b>	Weekly in high-risk patients	Routinely during RT (after every fraction in PM-dependent patients)	After every fraction (at least weekly in non-PM dependent patients)	Low risk: weekly in case of ICD Medium risk: weekly High risk: within 24 hours after each fraction	Yes (not specified in detail)
<b>After the RT course</b>	Yes	Not specified	Not specified	Yes, including controls at 1, 3, and 6 months after the completed RT	Yes
<b>Inactivation of antitachycardia therapies</b>	-	No	Yes	Yes	Yes
<b>Lead shielding of the device</b>	Yes (if near the RT field)	Can be considered	No	No	No
<b>Heart rhythm monitoring during RT</b>	In high-risk patients	Should be available in high-risk patients	In PM-dependent and ICD patients	In high-risk patients	Yes

Gy = gray; ICD = implantable cardioverter defibrillator; MV = megavolt; PM = pacemaker; RT = radiotherapy.

**Table 11: Summary of recommendations from the major PM/ICD manufacturers regarding safe radiotherapy in PM/ICD patients.**

Recommendation	Biotronik <sup>125</sup>	Boston Scientific <sup>123</sup>	Medtronic <sup>122</sup>	St. Jude Medical <sup>124</sup>
<b>Maximal PM dose</b>	2 Gy	No safe dose (2 Gy as a reference)	5 Gy	No safe dose
<b>Maximal ICD dose</b>	2 Gy	No safe dose (2 Gy as a reference)	1-5 Gy depending on the model	No safe dose
<b>Maximal beam energy</b>	<10 MV	Not stated	≤10 MV	Not stated
<b>Device checks</b>				
<b>Before the RT course</b>	Yes	Specific to each patient	Not stated	Not stated
<b>During the RT course</b>	Not stated	Specific to each patient	Yes (after each fraction if the recommended safe dose is exceeded)	Yes (a detailed evaluation once or twice during the RT course in PM-dependent patients)
<b>After the RT course</b>	Yes, including a supplementary follow-up shortly after the RT	Yes, including subsequent close monitoring of the device function	Yes (intensified follow-up schedule)	Yes
<b>Inactivation of antitachycardia therapies</b>	Yes	Yes	Yes	Yes
<b>Lead shielding of the device</b>	Yes	All available shielding options, including both internal shielding within the LINAC and external shielding of the patient	No (ineffective against neutrons)	Not stated (reduction in the device dose is recommended)
<b>Heart rhythm monitoring during RT</b>	Yes	As determined most appropriate by the physician team	Not stated	Yes

Gy = gray; ICD = implantable cardioverter defibrillator; LINAC = linear accelerator; MV = megavolt; PM = pacemaker; RT = radiotherapy.

## Main findings

The overall aim of this study was to enlighten epidemiological aspects of PM/ICD patients undergoing RT and to evaluate the rate of malfunctions in modern cardiac rhythm devices. We also sought to elucidate the risk factors of these malfunctions in order to be able to tailor safety measures according to the risk. A short summary of the main findings is presented in the following:

### Study I

- Risk of malfunction of modern electronic cardiac rhythm devices during RT correlates with photon beam energy.
- During irradiations with low-energy photons, no PM/ICD malfunctions occurred at therapeutic doses of radiation.

### Study II

- The porcine model is feasible for investigating RT effects on implanted cardiac rhythm devices.
- No oversense was recorded during direct irradiations of modern ICDs *in vivo*.
- The devices were fully functional despite radiation doses considerably higher than the often recommended safe limit of 1 Gy.

### Study III

- The rate of PM/ICD patients undergoing RT in Western Denmark increased from 1.45 to 4.33 per 100,000 person-years from 2003 to 2012.
- Discrepancies were shown between current recommendations and safety measures used in clinical practice, especially in terms of device evaluations after the RT.
- PM/ICD malfunctions were observed in 3.1% RT courses and consisted predominantly of resets.
- Beam energy seems to be the most important risk factor of PM/ICD malfunction during RT.
- Radiation dose plays a lower role in inducing PM/ICD malfunctions than previously anticipated.
- No PM/ICD malfunctions were observed during RT with electrons or kilovolt photons.

## Strengths and limitations

**Study I** was the first study to compare head to head two photon beam energies as a predictor of malfunctions in directly irradiated modern PMs/ICDs. The study sample consisted of 12 devices, limited to 2 or 4 from each manufacturer. This limits the statistical power of the analyses. The unbalanced design of the study limited our ability to compare the incidence rate of all events. When collapsing non-overlapping intervals to achieve a balanced design, some information will be lost. No direct telemetry or monitoring of device output was performed during the irradiations. Minor software errors, possibly not reported at interrogations by the clinical programmer, could have been missed. Measurements of neutron doses were not performed in this study.

In **Study II**, an animal model was applied for the first time in the field of RT and PMs/ICDs, bringing the experimental setup close to a clinical setting. However, the study size is limited. Also, the animals had to be euthanized immediately after the study, as a radiation dose of 37 Gy delivered during 1 day would be expected to result in a severe radiation injury. In this study, the radiation fractions were delivered faster than in the clinical situation. The reason

for this was that a prolonged *in vivo* experiment lasting several weeks with daily irradiations and general anesthesia would be poorly tolerated by the animals.

To our knowledge, **Study III** is the largest study on RT in PM/ICD patients published to date. The population-based design allowed us to include all PM/ICD patients receiving RT in Western Denmark over a period of 10 years. In this study, the proportion of RT patients having a PM/ICD was not examined. Neither the radiation dose to the device nor the distance from the RT field to the PM/ICD were generally available in the medical records. Therefore, the anatomical location of the RT was used as a surrogate marker in the statistical model. Due to high mortality among the patients, approximately one third of devices were never controlled after RT. Being outside the scope of the study, the causes of death were not analyzed. Thus, we may have underestimated both the occurrence and degree of severity of RT-induced device malfunctions. On the other hand, as beam energy and proportion of ICDs were lower in RT courses without control, we find it unlikely that we underestimated the occurrence to a major extent.

In this study, we can neither rule out transient asymptomatic effects of radiation on the devices that were not detectable at subsequent PM/ICD controls. The subject of RT in PM/ICD patients was approached from a clinical and epidemiological perspective, and no dosimetric measurements were performed. Neither were we able to assess in technical detail the changes of the CMOS in the devices that exhibited malfunctions during the irradiations. Finally, the studied beam types were limited to photons and electrons.

## Future perspectives

Reduction in neutron contamination during high-energy photon beaming and circuit hardening against errors might be among the possible approaches to prevent software-based PM/ICD malfunctions. We could also suggest manufacturers to consider a built-in “safe-RT” approach in the devices. This could be a hardware-based pacing mode running on especially radiation-resistant components of the circuit. It could be activated prior to RT, similarly to MRI-dedicated mode in some models. This could possibly avoid erratic and unexpected switching of the device to fall-back settings.

In our opinion, further studies on RT and PMs/ICDs should also include close imitations of real clinical scenarios, and the animal model seems to allow experiments that can serve as an interim step between *in vitro* and clinical studies. Furthermore, as device malfunctions during RT are relatively rare, large study populations may be needed, and therefore, a web-based registry of PM/ICD patients undergoing RT could be an option.<sup>40,118</sup>


In the waiting time for an update on international guidelines on safe RT in this patient group, the following suggestions could be advocated by the authors (Figure 9):

- 1) device relocation before RT is usually unnecessary as PM/ICD malfunctions only weakly correlate with radiation dose;
- 2) inactivation of antitachycardia therapies during RT as well as heart rhythm monitoring might be redundant;
- 3) patients undergoing RT with electrons or kV photons normally do not need supplementary device evaluations in PM/ICD clinic;
- 4) as the malfunctions mainly consist of resets, some safety measures such as repeated visits at PM/ICD clinic might

be abandoned in selected cases in favor of the comfort of the patient;

5) photon beam energy should be limited to  $\leq 10$  MV when possible.

**Figure 9: Flowchart of suggested safety measures during radiotherapy in patients with pacemakers and implantable cardioverter defibrillators.**



<sup>a</sup> If available, remote monitoring with daily evaluations could be considered instead of controls in the clinic. <sup>b</sup> In case the last evaluation is more than three months old. <sup>c</sup> Device evaluations after every fraction in case of a reset. CRT-D = cardiac resynchronization therapy defibrillator; CRT-P = cardiac resynchronization therapy pacemaker; ICD = implantable cardioverter defibrillator; MV = megavolt; RT = radiotherapy.

## Conclusions

The focus of this project is RT in cancer patients with implanted PM/ICD. With previous data in the field mainly originating from in vitro experiments and smaller-scale patient series, the current work approaches the subject from experimental, clinical, and epidemiological standpoints. In accordance with theoretical expectations, we highlight a substantial increase in the size of this patient group over the last decade. At times difficult to predict, PM/ICD malfunctions occur at a rate that is sufficient to pose a clinical challenge in tailoring optimal safety measures during RT in PM/ICD patients, and we also demonstrate that variations exist in applying these measures in clinic. We show that device malfunctions occurring in relation to RT mainly consist of software impairments eventually leading to a reset of the PM/ICD, wiping out the individually tailored parameters of the device. No fatal events or device replacements due to the malfunctions were observed in Study III, suggesting that the clinical consequences of these malfunctions tend to be relatively mild.

We found that radiation dose plays a smaller role in inducing PM/ICD malfunctions than generally anticipated. As one of the

explanations for this, we confirm the previous observations that beam energy is the essential factor in inducing the damaging effects on PMs/ICDs. In line with this finding, a considerable number of device malfunctions occurred during RT to remote anatomical areas, such as abdomen and pelvis, where photon beams of higher energy are typically used. In addition, we observed no device malfunctions during RT with kV photons or electrons. The animal model in Study II proved to be feasible in investigating the effects of RT on heart rhythm devices and permitted us to bring the experiments close to a clinical setting where we demonstrate that modern devices might tolerate relatively high radiation doses. Based on all three studies, RT may be delivered safely in carefully selected patients without the need to remove the PM/ICD from the vicinity of the RT field. On the other hand, devices can still malfunction even at exposure to only minimal radiation doses during high-energy photon RT.

# Summary

## Background and objectives

The number of individuals with pacemakers (PMs) and implantable cardioverter defibrillators (ICDs) increases. Approximately 6,000 of these devices were implanted in Denmark in 2013. Due to the ageing of the world population and with age being a risk factor for both cardiovascular diseases and cancer, there is a concern that increasing numbers of PM/ICD patients will develop malignancies and receive radiotherapy (RT) treatment. Modern PMs/ICDs have been shown to be susceptible to malfunctions during RT, possibly posing a threat to the safety of the patient. These malfunction events may be difficult to predict as they seem to only weakly correlate with radiation dose. Some malfunctions have been reported during only scatter radiation, while other devices appear to be able to resist high doses of direct irradiation. Little is also known about the size of this patient group, and in what proportion of RT courses PMs/ICDs eventually fail. Data on safety measures applied while treating PM/ICD patients with RT are scarce as well.

Following hypotheses were addressed by this thesis:

- Modern PMs/ICDs can resist higher doses of ionizing radiation than generally anticipated (Study I and II).
- Animal models are feasible for studying the effects of RT on ICDs *in vivo* (Study II).
- The rate of RT in PM/ICD patients in the general population is increasing (Study III).
- The usage of safety measures varies during RT in PM/ICD patients in clinical practice (Study III).
- PM/ICD malfunctions can be predicted based on parameters of RT and/or type of the device (Study I and III).

## Methods

In Study I, 12 modern devices (10 PMs, 2 ICDs) from five different manufacturers were irradiated *in vitro* with either 6 megavolt (MV) photons or 18 MV photons. Reaching cumulative radiation doses to the device of 150 grays (Gy), the function of the devices was assessed regularly during the experiment. The malfunction rate was then compared between the 6 MV and 18 MV groups.

In Study II, five pigs were implanted with a functional ICD (one from each major manufacturer), and a clinical scenario with an ICD located in the radiation field was simulated. Every device was irradiated with increasing doses of 6 MV and 18 MV photons with a

cumulative device dose of 37 Gy. The functionality of the ICDs was controlled during and after the irradiations.

In Study III, after conducting a search at the Danish National Patient Registry, an observational study of 560 patients undergoing 678 RT courses during 2003-2012 in Western Denmark was performed. Medical records of both PM and ICD patients undergoing all types of external beam RT were reviewed in order to identify applied safety measures and recorded device malfunctions.

## Results

In the *in vitro* study (Study I), the devices failed earlier and at a higher rate during irradiations with 18 MV photons compared with irradiations with 6 MV photons. The first malfunction during exposure to 6 MV photons occurred at radiation doses considerably higher than used during cancer RT. The animal model (Study II) was feasible for investigating RT effects on implanted cardiac devices. All devices were fully functional during the *in vivo* irradiations, although one recoverable malfunction was observed in one of the ICDs. The rate of RT courses in PM/ICD patients increased 3-fold from 2003 to 2012 (Study III). Applied safety measures during RT in this patient group varies, with only at most a third of patients being seen at PM/ICD clinics due to the RT. PM/ICD malfunctions were observed in 3.1% of RT courses with beam energy being the strongest predictor for malfunction. There was a trend of device malfunctions to be more frequent in ICDs compared with PMs, while radiation dose was not associated with device malfunctions.

## Conclusion

With an increasing number of PM/ICD patients undergoing RT, there is a considerable risk that these electronic devices may malfunction during exposure to ionizing radiation. As the beam energy emerges as the most important predictor of PM/ICD malfunctions, the radiation dose plays a smaller role in inducing these effects than might be expected. The majority of PM/ICD malfunctions are transient and do not harm the overall function of the device. Variations exist in applying safety measures in these patients, and a tailored approach should be taken based primarily on the risk of PM/ICD malfunction during RT.

# Dansk resumé

## Baggrund og formål

Antallet patienter med pacemakere (PM) og implanterbare kardioverterdefibrillatorer (ICD) er stigende. Alene i 2013 blev der indopereret omkring 6.000 af disse elektroniske apparater i Danmark. Da både risikoen for hjertesygdomme og cancer stiger med alderen og på grund af den stigende levealder, må det forventes at andelen af patienter med PM/ICD, som udvikler cancer med efterfølgende behov for stråleterapi, vil stige over tid. Moderne PM/ICD kan imidlertid beskadiges af de ioniserende stråler, som anvendes under denne behandling, hvilket potentielt kan bringe patienten i fare. Disse svigt af PM/ICD kan dog være vanskelige at forudsige, idet de ikke synes at hænge sammen med stråledosis. Nogle apparater påvirkes efter blot at have været utsat for spredt stråling, mens andre ikke viser tegn på svigt selv efter høj stråledosis. Desuden ved man kun lidt om, hvor stort et problem disse svigt reelt udgør, og om, hvilke sikkerhedsforanstaltninger, der bliver anvendt hos PM/ICD patienter under stråleterapi.

Formål med denne thesis var at teste følgende hypoteser:

- Moderne PM/ICD kan tåle højere stråledosis end hidtil antaget (Studie I og II).
- Dyremodel kan anvendes til at undersøge virkningen af stråleterapi på ICD in vivo (Studie II).
- Antallet af stråleterapi behandlinger hos PM/ICD patienter i Danmark er stigende (Studie III).
- Der er forskelle i anvendelse af sikkerhedsforanstaltninger under stråleterapi hos PM/ICD patienter (Studie III).
- Risikoen for PM/ICD svigt under stråleterapi kan forudsiges med baggrund i parametre af stråleterapien og/eller typen af PM/ICD (Studie I og III).

## Metode

I Studie I blev 12 moderne PM/ICD (10 PM, 2 ICD) fra fem forskellige producenter bestrålet in vitro med enten 6 megavolt (MV) fotoner eller 18 MV fotoner. Funktionen af PM/ICD blev kontrolleret under strålingerne. Samtlige devices modtog en kumuleret stråledosis på 150 gray (Gy), hvorefter antallet af svigt blev sammenlignet mellem 6 MV og 18 MV grupperne.

I Studie II fik fem grise indopereret en fungerende ICD enhed (én fra hver af de store producenter), hvorefter en klinisk situation med en ICD beliggende i strålefeltet blev simuleret. Hver ICD blev bestrålet med stigende doser af både 6 MV og 18 MV fotoner med

en kumuleret dosis på 37 Gy til hver ICD. Funktionen af ICD enhederne blev kontrolleret under og efter strålingerne.

I Studie III indgik 560 patienter som gennemgik 678 stråleterapi behandlinger i Jylland i løbet af 2003-2012. Patienterne blev identificeret i Landspatientregistret. Deres patientjournaler blev gennemgået for at identificere dels anvendte sikkerhedsforanstaltninger, forekomst af svigt af PM/ICD og mulige prædiktorer for disse svigt.

## Resultater

In vitro studiet (Studie I) viste, at PM/ICD svigtede tidligere og oftere under bestrålinger med 18 MV fotoner sammenlignet med 6 MV fotoner. Under sidstnævnte opstod der først svigt ved en kumuleret stråledosis, som var væsentligt højere end den, man anvender ved behandling af cancer i klinikken. Studie II viste, at undersøgelse af funktionen af PM/ICD under stråleterapi er gennemførlig i en dyremodel. Samtlige ICDer var fuldt fungerende under in vivo bestrålingerne, selvom der efterfølgende opstod en imidlertid omprogrammering til nominel standard af en ICD. Studie III viste, at antallet stråleterapi behandlinger hos PM/ICD patienter er steget markant med en tredobling af antallet fra 2003 til 2012. Der er betydelige forskelle i håndteringen af PM/ICD patienter under RT, og kun højst en tredjedel af patienterne bliver kontrolleret i PM/ICD-ambulatorier i forbindelse med stråleterapi. Svigt af PM/ICD som hovedsageligt bestod af omprogrammeringer til nominel standard opstod ved 3,1% af stråleterapi behandlingerne, og den stærkeste prædiktor for disse episoder var anvendt stråleenergi. ICDer udviste tilbøjelighed til at svigte hyppigere end PM, og der var ingen sammenhæng mellem stråledosis og PM/ICD svigt.

## Konklusion

Set i lyset af stigende antal PM/ICD patienter, som gennemgår stråleterapi for cancer, findes der en risiko for, at disse apparater sviger under strålebehandling. Mens stråleenergi spiller en afgørende rolle i forekomst af svigt af PM/ICD, synes stråledosis at have mindre betydning end hidtil antaget. Hovedparten af de svigt, der opstår, er forbigående og medfører som regel ingen varig beskadigelse af PM/ICD. Der findes forskelle i praktisk håndtering af disse patienter, og behandlingen bør tilpasses ud fra forventede risiko for PM/ICD svigt under stråleterapi.

# Appendices

## Appendix A

### Diagnosis and procedure codes used during the search in the Danish National Patient Registry in Study III

<b>ICD-10</b>	
Presence of cardiac pacemaker, excluding loop recorder	Z95.0, excluding Z95.0L
<b>SKS</b>	
Pacemaker- or ICD-related intervention, excluding application of loop recorder	BFC, excluding BFCAS
<b>NOMESCO Classification of Surgical Procedures</b>	
Implantation or replacement of permanent transvenous cardiac pacemaker	FPE
Implantation or replacement of permanent epicardial pacemaker	FPF
Implantation of permanent cardioverter-defibrillator	FPG
Removal of permanent cardiac pacemaker or cardioverter-defibrillator	FPH
Revision of pacemaker pulse generator or electrode	FPJ
Other operations for arrhythmias and disturbances of impulse propagation	FPW
Temporary use of transvenous or epicardial pacemaker	TFP00
<b>Codes used historically before NOMESCO Classification of Surgical Procedures</b>	
Implantation of pacemaker	30930
Transvenous implantation of cardiac electrode	32100
Surgical implantation of cardiac electrode	32105
Transvenous implantation of atrial electrode	32110
Surgical implantation of atrial electrode	32115
Implantation of generator (accelerator cordis)	32120
Replacement of generator	32121
Implantatio convertatoris cordis (defibrillator)	32122
Replacement of transvenous electrode	32130
Revision of electrode	32131
Implantation, revision, and removal of defibrillator	32159
Other pacemaker/defibrillator operation	32199
External pacemaker	32490
Implantationes electrodum	32600
Implantatio acceleratoris cordis	32610
Acceleratio cordis endocardialis (venous electrode)	32620
Implantationes electrodum in myocardium	32640
Excisio corporis alieni cordis (pacemaker, electrode etc.)	32660
Replacement of pacemaker	32680
<b>SKS</b>	
Planning of radiotherapy	BWGA
External radiotherapy	BWGC

ICD = implantable cardioverter defibrillator; ICD-10 = International Classification of Diseases, 10<sup>th</sup> edition; NOMESCO = Nordic Medico-Statistical Committee; SKS = Danish Hospital Sector Classification System.

## Appendix B

### Characteristics of radiotherapy (RT) courses in Study III resulting in device malfunction versus courses without malfunction by device manufacturer, including courses with complete data on both the device and the RT (n=658)

Manufacturer	All RT courses (n=658)	PM/ICD controlled after RT [n=453 (68.8%)]	No PM/ICD malfunction [n=439 (96.9%)] <sup>*</sup>	PM/ICD malfunction [n=14 (3.1%)] <sup>*</sup>
Biotronik, n (%)	15	12 (80.0)	11 (91.7)	1 (8.3)
Boston Scientific, n (%)	11	8 (72.7)	8 (100.0)	0 (0.0)
CPI, n (%)	1	1 (100.0)	1 (100.0)	0 (0.0)
Ela, n (%)	29	18 (62.1)	16 (88.9)	2 (11.1)
Guidant, n (%)	47	31 (66.0)	31 (100.0)	0 (0.0)
Medtronic, n (%)	267	191 (71.5)	188 (98.4)	3 (1.6)
Pacesetter, n (%)	48	32 (66.7)	32 (100.0)	0 (0.0)
Siemens, n (%)	4	2 (50.0)	2 (100.0)	0 (0.0)
Sorin, n (%)	1	0 (0.0)	0 (-)	0 (-)
St. Jude Medical, n (%)	199	134 (67.3)	126 (94.0)	8 (6.0)
Telectronics, n (%)	1	0 (0.0)	0 (-)	0 (-)
Ventritex, n (%)	1	1 (100.0)	1 (100.0)	0 (0.0)
Vitatron, n (%)	34	23 (67.6)	23 (100.0)	0 (0.0)

ICD = implantable cardioverter defibrillator; PM = pacemaker; RT = radiotherapy. <sup>\*</sup>Percentages are row percentages to allow for comparison of RT courses without and with malfunctions.

## Appendix C

### Recorded device malfunctions during radiotherapy in Study III

Device type	Manufacturer	Model	Anatomic area irradiated	Cumulative tumor dose / number of fractions	Beam type	Maximal beam energy, MV	Description
AAI	St. Jude Medical	Victory SR 5610	Thorax	48 Gy/24	Photons and electrons (6 MeV)	18	Reset (and diaphragmatic pacing) after 10 Gy to the tumor. Device was relocated before RT.
DDD	Ela	Symphony DR 2550	Abdomen and pelvis	70 Gy/35	Photons	15	Spontaneous reprogramming to unipolar pacing, output 3.5 V/0.35 ms at both leads, basic rate 60 bpm. Dose at the event uncertain.
DDD	Ela	Symphony DR 2550	Head and neck	20 Gy/4	Photons	6	Program error preventing readout of historical data at a control 14 days after RT, otherwise normal function.
DDD	Medtronic	Adapta ADDR01	Thorax	30 Gy/10	Photons	6	Permanent increase in atrial pacing threshold from 1.25 V (before RT) to 2.75 V (after RT), other lead parameters unchanged.
DDD	St. Jude Medical	Affinity DR 5330R	Thorax	48 Gy/24	Photons	18	Intermittent warnings about invalid data since 14 Gy to the tumor, otherwise normal function.
DDD	St. Jude Medical	Identity ADx XL DR 5386	Thorax	20 Gy/4	Photons	18	Reset requiring reprogramming by the manufacturer after 20 Gy to the tumor.
DDD	St. Jude Medical	Victory XL DR 5816	Spine	8 Gy/1	Photons	18	Reset.
DDD	St. Jude Medical	Victory XL DR 5816	Abdomen and pelvis	78 Gy/39	Photons	15	Reset after 68 Gy to the tumor.
DDD	St. Jude Medical	Zephyr XL DR 5826	Lower extremity	8 Gy/1	Photons	15	Reset; battery parameters not available ("Data not read"); Fast Path summary: "Diagnostics cleared because they were invalid".
DDD	St. Jude Medical	Zephyr XL DR 5826	Abdomen and pelvis	45 Gy/25	Photons	15	Reset, dose at the event uncertain.
VVI-ICD	Biotronik	Lumax 540 VR-T	Abdomen and pelvis	78 Gy/39	Photons	6	Reset and permanent increase in pacing threshold to 4.4 V after 4 Gy to the tumor, other lead parameters unchanged
DDD-ICD	St. Jude Medical	Atlas + DR V-243	Abdomen and pelvis	60 Gy/30	Photons	18	Reset requiring reprogramming by the manufacturer after 40 Gy.
CRT-D	Medtronic	InSyncMaximo7304	Abdomen and pelvis	78 Gy/39	Photons	18	Reset after 78 Gy to tumor.
CRT-D	Medtronic	InSync III Marquis 7279	Abdomen and pelvis	25 Gy/5	Photons	18	Reset during the period of RT, dose at the event uncertain.

AAI = single chamber atrial pacemaker; bpm = beats per minute; CRT-D = cardiac resynchronization therapy implantable cardioverter defibrillator; DDD = dual chamber pacemaker; DDD-ICD = dual chamber implantable cardioverter defibrillator; Gy = gray; MeV = megaelectronvolt; ms = miliseconds; MV = megavolt; V = volt; VVI-ICD = single chamber implantable cardioverter defibrillator.

## References

1. Larsson B, Elmquist H, Ryden L, Schuller H. Lessons from the first patient with an implanted pacemaker: 1958-2001. *Pacing Clin Electrophysiol* 2003; 26:114-124.
2. Zipes DP, Camm AJ, Borggrefe M, Buxton AE, Chaitman B, Fromer M, Gregoratos G, et al. ACC/AHA/ESC 2006 guidelines for management of patients with ventricular arrhythmias and the prevention of sudden cardiac death--executive summary: A report of the American college of Cardiology/American heart association task force and the European society of cardiology committee for practice guidelines (writing committee to develop guidelines for management of patients with ventricular arrhythmias and the prevention of sudden cardiac death) developed in collaboration with the European heart rhythm association and the heart rhythm society. *Eur Heart J* 2006; 27:2099-2140.
3. Brignole M, Auricchio A, Baron-Esquivias G, Bordachar P, Boriani G, Breithardt OA, Cleland J, et al. 2013 ESC guidelines on cardiac pacing and cardiac resynchronization therapy: The task force on cardiac pacing and resynchronization therapy of the European society of cardiology (ESC). Developed in collaboration with the European heart rhythm association (EHRA). *Eur Heart J* 2013; 34:2281-2329.
4. Goldenberg I, Moss AJ. Implantable device therapy. *Prog Cardiovasc Dis* 2008; 50:449-474.
5. Stone KR, McPherson CA. Assessment and management of patients with pacemakers and implantable cardioverter defibrillators. *Crit Care Med* 2004; 32:S155-S165.
6. Bristow MR, Saxon LA, Boehmer J, Krueger S, Kass DA, De Marco T, Carson P, et al. Cardiac-resynchronization therapy with or without an implantable defibrillator in advanced chronic heart failure. *N Engl J Med* 2004; 350:2140-2150.
7. Cleland JG, Daubert JC, Erdmann E, Freemantle N, Gras D, Kappenberger L, Tavazzi L, et al. The effect of cardiac resynchronization on morbidity and mortality in heart failure. *N Engl J Med* 2005; 352:1539-1549.
8. Moss AJ, Zareba W, Hall WJ, Klein H, Wilber DJ, Cannom DS, Daubert JP, et al. Prophylactic implantation of a defibrillator in patients with myocardial infarction and reduced ejection fraction. *N Engl J Med* 2002; 346:877-883.
9. Bardy GH, Lee KL, Mark DB, Poole JE, Packer DL, Boineau R, Domanski M, et al. Amiodarone or an implantable cardioverter-defibrillator for congestive heart failure. *N Engl J Med* 2005; 352:225-237.
10. Mond HG, Proclemer A. The 11th world survey of cardiac pacing and implantable cardioverter-defibrillators: Calendar year 2009--a world society of arrhythmia's project. *Pacing Clin Electrophysiol* 2011; 34:1013-1027.
11. Kuck KH, Hindricks G, Padeletti L, Raatikainen P, Arnar DO. The EHRA white book 2014: The current status of cardiac electrophysiology in ESC member countries. Les Templiers: European Heart Rhythm Association; 2014.
12. Pinski SL, Trohman RG. Interference in implanted cardiac devices, part I. *Pacing Clin Electrophysiol* 2002; 25:1367-1381.
13. Beinart R, Nazarian S. Effects of external electrical and magnetic fields on pacemakers and defibrillators: From engineering principles to clinical practice. *Circulation* 2013; 128:2799-2809.
14. Pinski SL, Trohman RG. Interference in implanted cardiac devices, part II. *Pacing Clin Electrophysiol* 2002; 25:1496-1509.
15. Vaupel JW. Biodemography of human ageing. *Nature* 2010; 464:536-542.
16. Torre LA, Bray F, Siegel RL, Ferlay J, Lortet-Tieulent J, Jemal A. Global cancer statistics, 2012. *CA Cancer J Clin* 2015; 65:87-108.
17. Fernandez-Ballesteros R, Robine JM, Walker A, Kalache A. Active ageing: A global goal. *Curr Gerontol Geriatr Res* 2013; 2013:298012.
18. Alwan A. Global status report on noncommunicable diseases 2010. World Health Organization; 2011.
19. Sanderson JE, Mayosi B, Yusuf S, Reddy S, Hu S, Chen Z, Timmis A. Global burden of cardiovascular disease. *Heart* 2007; 93:1175.
20. Labarthe DR, Dunbar SB. Global cardiovascular health promotion and disease prevention: 2011 and beyond. *Circulation* 2012; 125:2667-2676.
21. Mendis S, Puska P, Norrving B. Global atlas on cardiovascular disease prevention and control. World Health Organization; 2011.
22. Mathers CD, Loncar D. Projections of global mortality and burden of disease from 2002 to 2030. *PLoS Med* 2006; 3:e442.
23. Delaney G, Jacob S, Featherstone C, Barton M. The role of radiotherapy in cancer treatment: Estimating optimal utilization from a review of evidence-based clinical guidelines. *Cancer* 2005; 104:1129-1137.
24. Kapa S, Fong L, Blackwell CR, Herman MG, Schomberg PJ, Hayes DL. Effects of scatter radiation on ICD and CRT function. *Pacing Clin Electrophysiol* 2008; 31:727-732.
25. Gelblum DY, Amols H. Implanted cardiac defibrillator care in radiation oncology patient population. *Int J Radiat Oncol Biol Phys* 2009; 73:1525-1531.
26. Ferrara T, Baiotto B, Malinvernini G, Caria N, Garibaldi E, Barboni G, Stasi M, et al. Irradiation of pacemakers and cardio-defibrillators in patients submitted to radiotherapy: A clinical experience. *Tumori* 2010; 96:76-83.
27. Overgaard M, Christensen JJ. Postoperative radiotherapy in DBCG during 30 years. Techniques, indications and clinical radiobiological experience. *Acta Oncol* 2008; 47:639-653.
28. Mauguen A, Le Pechoux C, Saunders MI, Schild SE, Turrisi AT, Baumann M, Sause WT, et al. Hyperfractionated or accelerated radiotherapy in lung cancer: An individual patient data meta-analysis. *J Clin Oncol* 2012; 30:2788-2797.
29. Horning SJ, Weller E, Kim K, Earle JD, O'Connell MJ, Habermann TM, Glick JH. Chemotherapy with or without radiotherapy in limited-stage diffuse aggressive non-Hodgkin's lymphoma: Eastern cooperative oncology group study 1484. *J Clin Oncol* 2004; 22:3032-3038.
30. Engert A, Plutschow A, Eich HT, Lohri A, Dorken B, Borchmann P, Berger B, et al. Reduced treatment intensity in patients with early-stage Hodgkin's lymphoma. *N Engl J Med* 2010; 363:640-652.
31. Lutz S, Berk L, Chang E, Chow E, Hahn C, Hoskin P, Howell D, et al. Palliative radiotherapy for bone metastases: An ASTRO evidence-based guideline. *Int J Radiat Oncol Biol Phys* 2011; 79:965-976.
32. Lutz S, Chow E. A review of recently published radiotherapy treatment guidelines for bone metastases: Contrasts or convergence? *Journal of Bone Oncology* 2012; 1:18-23.
33. Hildner FJ, Linhart JW, Poole DO. Irradiation of pulmonary tumor with overlying artificial cardiac pacemaker. *Radiology* 1969; 92:148-149.
34. Eipper HH, Laufenberg E. Performance of implantable pacemakers during betatron high energy irradiation. *Strahlentherapie* 1972; 143:644-654.
35. Walz BJ, Reder RE, Pastore JO, Littman P, Johnson R. Cardiac pacemakers. Does radiation therapy affect performance? *JAMA* 1975; 234:72-73.
36. Last A. Radiotherapy in patients with cardiac pacemakers. *Br J Radiol* 1998; 71:4-10.
37. Marbach JR, Meoz-Mendez RT, Huffman JK, Hudgins PT, Almond PR. The effects of cardiac pacemakers of ionizing radiation and electromagnetic interference from radiotherapy machines. *Int J Radiat Oncol Biol Phys* 1978; 4:1055-1058.
38. Marbach JR, Sontag MR, Van Dyk J, Wolbarst AB. Management of radiation oncology patients with implanted cardiac pacemakers: Report of AAPM task group no. 34. American association of physicists in medicine. *Med Phys* 1994; 21:85-90.

39. Solan AN, Solan MJ, Bednarz G, Goodkin MB. Treatment of patients with cardiac pacemakers and implantable cardioverter-defibrillators during radiotherapy. *Int J Radiat Oncol Biol Phys* 2004; 59:897-904.

40. Hudson F, Coulshed D, D'Souza E, Baker C. Effect of radiation therapy on the latest generation of pacemakers and implantable cardioverter defibrillators: A systematic review. *J Med Imaging Radiat Oncol* 2010; 54:53-61.

41. Hurkmans CW, Kneegjens JL, Oei BS, Maas AJ, Uiterwaal GJ, Borden AJ, Ploegmakers MM, et al. Management of radiation oncology patients with a pacemaker or ICD: A new comprehensive practical guideline in the Netherlands. *Radiat Oncol* 2012; 7:198.

42. Harcombe AA, Newell SA, Ludman PF, Wistow TE, Sharples LD, Schofield PM, Stone DL, et al. Late complications following permanent pacemaker implantation or elective unit replacement. *Heart* 1998; 80:240-244.

43. Catanchin A, Murdock CJ, Athan E. Pacemaker infections: A 10-year experience. *Heart Lung Circ* 2007; 16:434-439.

44. Gould PA, Gula LJ, Champagne J, Healey JS, Cameron D, Simpson C, Thibault B, et al. Outcome of advisory implantable cardioverter-defibrillator replacement: One-year follow-up. *Heart Rhythm* 2008; 5:1675-1681.

45. Kesek M, Nyholm T, Asklund T. Radiotherapy and pacemaker: 80 gy to target close to the device may be feasible. *Europace* 2012; 14:1595.

46. Ahmed I, Zou W, Jabbour SK. High dose radiotherapy to automated implantable cardioverter-defibrillator: A case report and review of the literature. *Case Rep Oncol Med* 2014; 2014:989857.

47. Lau DH, Wilson L, Stiles MK, John B, Shashidhar, Dimitri H, Brooks AG, et al. Defibrillator reset by radiotherapy. *Int J Cardiol* 2008; 130:e37-8.

48. Gelblum DY, Amols H. Implanted cardiac defibrillator care in radiation oncology patient population. *Int J Radiat Oncol Biol Phys* 2009; 73:1525-1531.

49. Soejima T, Yoden E, Nishimura Y, Ono S, Yoshida A, Fukuda H, Fukuhara N, et al. Radiation therapy in patients with implanted cardiac pacemakers and implantable cardioverter defibrillators: A prospective survey in Japan. *J Radiat Res* 2011; 52:516-521.

50. Makkar A, Prisciandaro J, Agarwal S, Lusk M, Horwood L, Moran J, Fox C, et al. Effect of radiation therapy on permanent pacemaker and implantable cardioverter-defibrillator function. *Heart Rhythm* 2012; 9:1964-1968.

51. Elders J, Kunze-Busch M, Jan Smeenk R, Smeets JL. High incidence of implantable cardioverter defibrillator malfunctions during radiation therapy: Neutrons as a probable cause of soft errors. *Europace* 2013; 15:60-65.

52. Ghilezan M, Martinez AA. Brachytherapy. In: Gunderson LL, Tepper JE, editors. *Clinical radiation oncology*. 3rd ed. Philadelphia: Elsevier Saunders; 2012. p. 259-86.

53. Croshaw R, Kim Y, Lappinen E, Julian T, Trombetta M. Avoiding mastectomy: Accelerated partial breast irradiation for breast cancer patients with pacemakers or defibrillators. *Ann Surg Oncol* 2011; 18:3500-3505.

54. Keshtgar MR, Eaton DJ, Reynolds C, Pigott K, Davidson T, Gauthier Fleckenstein B, Wenz F. Pacemaker and radiotherapy in breast cancer: Is targeted intraoperative radiotherapy the answer in this setting? *Radiat Oncol* 2012; 7:128.

55. Kim Y, Arshoun Y, Trombetta MG. Pacemaker/implantable cardioverter-defibrillator dose in balloon high-dose-rate brachytherapy for breast cancer treatment. *Brachytherapy* 2012; 11:380-386.

56. Sundar S, Symonds RP, Deehan C. Radiotherapy to patients with artificial cardiac pacemakers. *Cancer Treat Rev* 2005; 31:474-486.

57. McCollough CH, Zhang J, Primak AN, Clement WJ, Buysman JR. Effects of CT irradiation on implantable cardiac rhythm management devices. *Radiology* 2007; 243:766-774.

58. Lyng E, Sandegaard JL, Rebolj M. The Danish national patient register. *Scand J Public Health* 2011; 39:30-33.

59. Kirkfeldt RE, Johansen JB, Nohr EA, Jorgensen OD, Nielsen JC. Complications after cardiac implantable electronic device implantations: An analysis of a complete, nationwide cohort in Denmark. *Eur Heart J* 2014; 35:1186-1194.

60. Hu FB, Goldberg J, Hedeker D, Flay BR, Pentz MA. Comparison of population-averaged and subject-specific approaches for analyzing repeated binary outcomes. *Am J Epidemiol* 1998; 147:694-703.

61. Liang K, Zeger SL. Longitudinal data analysis using generalized linear models. *Biometrika* 1986; 73:13-22.

62. Rodriguez F, Filimonov A, Henning A, Coughlin C, Greenberg M. Radiation-induced effects in multiprogrammable pacemakers and implantable defibrillators. *Pacing Clin Electrophysiol* 1991; 14:2143-2153.

63. Tsekos A, Momm F, Brunner M, Guttenberger R. The cardiac pacemaker patient--might the pacer be directly irradiated? *Acta Oncol* 2000; 39:881-883.

64. Gossman MS, Wilkinson JD, Mallick A. Treatment approach, delivery, and follow-up evaluation for cardiac rhythm disease management patients receiving radiation therapy: Retrospective physician surveys including chart reviews at numerous centers. *Med Dosim* 2014; 39:320-324.

65. Hurkmans CW, Scheepers E, Springorum BG, Uiterwaal H. Influence of radiotherapy on the latest generation of pacemakers. *Radiother Oncol* 2005; 76:93-98.

66. Mouton J, Haug R, Bridier A, Dodinot B, Eschwege F. Influence of high-energy photon beam irradiation on pacemaker operation. *Phys Med Biol* 2002; 47:2879-2893.

67. Hurkmans CW, Scheepers E, Springorum BG, Uiterwaal H. Influence of radiotherapy on the latest generation of implantable cardioverter-defibrillators. *Int J Radiat Oncol Biol Phys* 2005; 63:282-289.

68. Kapa S, Fong L, Blackwell CR, Herman MG, Schomberg PJ, Hayes DL. Effects of scatter radiation on ICD and CRT function. *Pacing Clin Electrophysiol* 2008; 31:727-732.

69. Hashii H, Hashimoto T, Okawa A, Shida K, Isobe T, Hanamura M, Nishimura T, et al. Comparison of the effects of high-energy photon beam irradiation (10 and 18 MV) on 2 types of implantable cardioverter-defibrillators. *Int J Radiat Oncol Biol Phys* 2013; 85:840-845.

70. Bradley P, Normand E. Single event upsets in implantable cardioverter defibrillators. *IEEE Trans Nucl Sci* 1998; 45:2929-2940.

71. Oshiro Y, Sugahara S, Noma M, Sato M, Sakakibara Y, Sakae T, Hayashi Y, et al. Proton beam therapy interference with implanted cardiac pacemakers. *Int J Radiat Oncol Biol Phys* 2008; 72:723-727.

72. Gomez DR, Poenisch F, Pinnix CC, Sheu T, Chang JY, Memon N, Mohan R, et al. Malfunctions of implantable cardiac devices in patients receiving proton beam therapy: Incidence and predictors. *Int J Radiat Oncol Biol Phys* 2013; 87:570-575.

73. Hashimoto T, Isobe T, Hashii H, Kumada H, Tada H, Okumura T, Tsuboi K, et al. Influence of secondary neutrons induced by proton radiotherapy for cancer patients with implantable cardioverter defibrillators. *Radiat Oncol* 2012; 7:10.

74. Raitt MH, Stelzer KJ, Laramore GE, Bardy GH, Dolack GL, Poole JE, Kudenchuk PJ. Runaway pacemaker during high-energy neutron radiation therapy. *Chest* 1994; 106:955-957.

75. Koivunoro H, Seren T, Hyvonen H, Kotiluoto P, Iivonen P, Auterinen I, Seppala T, et al. Epithermal neutron beam interference with cardiac pacemakers. *Appl Radiat Isot* 2011; 69:1904-1906.

76. Trigano A, Hubert G, Marfaing J, Castellani K. Experimental study of neutron-induced soft errors in modern cardiac pacemakers. *J Interv Card Electrophysiol* 2012; 33:19-25.

77. Souliman SK, Christie J. Pacemaker failure induced by radiotherapy. *Pacing Clin Electrophysiol* 1994; 17:270-273.

78. Mouton J, Trochet R, Vicrey J, Sauvage M, Chauvenet B, Ostrovski A, et al. In: *Electromagnetic and radiation environment effects on pacemakers*. Radiation and its effects on components and systems, 1999. RADECS 99. 1999 fifth European conference on; IEEE; 1999. p. 34-8.

79. Zaremba T, Jakobsen AR, Thøgersen AM, Oddershede L, Riahia S. The effect of radiotherapy beam energy on modern cardiac devices: An in vitro study. *Europace* 2014; 16:612-616.

80. Hoecht S, Rosenthal P, Sancar D, Behrens S, Hinkelbein W, Hoeller U. Implantable cardiac defibrillators may be damaged by radiation therapy. *J Clin Oncol* 2002; 20:2212-2213.

81. Mollerup M, Naslund L, Lipinski M, Meyer A, Libey B, Dornfeld K. Radiation tolerance of contemporary implantable cardioverter-defibrillators. *J Interv Card Electrophysiol* 2014; 39:171-175.

82. John J, Kaye GC. Shock coil failure secondary to external irradiation in a patient with implantable cardioverter defibrillator. *Pacing Clin Electrophysiol* 2004; 27:690-691.

83. Nemec J. Runaway implantable defibrillator--a rare complication of radiation therapy. *Pacing Clin Electrophysiol* 2007; 30:716-718.

84. Zweng A, Schuster R, Hawlicek R, Weber HS. Life-threatening pacemaker dysfunction associated with therapeutic radiation: A case report. *Angiology* 2009; 60:509-512.

85. Thomas D, Becker R, Katus HA, Schoels W, Karle CA. Radiation therapy-induced electrical reset of an implantable cardioverter defibrillator device located outside the irradiation field. *J Electrocardiol* 2004; 37:73-74.

86. Ferrick AM, Bernstein N, Aizer A, Chinitz L. Cosmic radiation induced software electrical resets in ICDs during air travel. *Heart Rhythm* 2008; 5:1201-1203.

87. Bhakta D, Foreman LD. Cosmic radiation: Not science fiction, but clinical reality. *Heart Rhythm* 2008; 5:1204-1205.

88. Ferrick AM. To the editor—response. *Heart Rhythm* 2008; 5:e1-e2.

89. Nibhanupudy JR, de Jesus MA, Fujita M, Goldson AL. Radiation dose monitoring in a breast cancer patient with a pacemaker: A case report. *J Natl Med Assoc* 2001; 93:278-281.

90. Ampil FL, Caldito G. Radiotherapy for palliation of lung cancer in patients with compromised hearts [2]. *J Palliat Med* 2006; 9:241-242.

91. Mitra D, Ghosh K, Gupta P, Jayanti J, Dev A, Sur P. Radiation dose monitoring in a lung cancer patient with a pacemaker - a case report. *Indian Journal of Radiology and Imaging* 2006; 16:875-877.

92. Ferrara T, Baiotto B, Malinvernini G, Caria N, Garibaldi E, Barboni G, Stasi M, et al. Irradiation of pacemakers and cardio-defibrillators in patients submitted to radiotherapy: A clinical experience. *Tumori* 2010; 96:76-83.

93. Dasgupta T, Barani IJ, Roach M, 3rd. Successful radiation treatment of anaplastic thyroid carcinoma metastatic to the right cardiac atrium and ventricle in a pacemaker-dependent patient. *Radiat Oncol* 2011; 6:16.

94. Wadasadawala T, Pandey A, Agarwal JP, Jalali R, Laskar SG, Chowdhary S, Budrukka A, et al. Radiation therapy with implanted cardiac pacemaker devices: A clinical and dosimetric analysis of patients and proposed precautions. *Clin Oncol (R Coll Radiol)* 2011; 23:79-85.

95. Kirova YM, Menard J, Chargari C, Mazal A, Kirov K. Case study thoracic radiotherapy in an elderly patient with pacemaker: The issue of pacing leads. *Med Dosim* 2012; 37:192-194.

96. Ampil FL, Nathan C-, Ghali G, Kim D. Postoperative radiotherapy for advanced head and neck cancer in patients with cardiac pacemakers. *Journal of Radiotherapy in Practice* 2014; 13:115-118.

97. Zaremba T, Jakobsen AR, Søgaard M, Thøgersen AM, Johansen MB, Madsen LB, Riahi S. Risk of device malfunction in cancer patients with implantable cardiac device undergoing radiotherapy: A population-based cohort study. *Pacing Clin Electrophysiol* 2015; 38:343-356.

98. Sepe S, Schaffer P, Krimmel K, Schaffer M. Irradiation treatment of laryngeal cancer in a patient with an implantable cardioverter-defibrillator (ICD). *Onkologie* 2007; 30:378-380.

99. Dell'Oca I, Tsachiris D, Oppizzi M, Della Bella P, Gulletta S. Radiotherapy and implanted cardioverter defibrillators: Novel techniques make it feasible. *J Cardiovasc Med (Hagerstown)* 2013.

100. Zaremba T, Jakobsen AR, Thøgersen AM, Riahi S, Kjaergaard B. Effects of high-dose radiotherapy on implantable cardioverter defibrillators: An in vivo porcine study. *Pacing Clin Electrophysiol* 2013; 36:1558-1563.

101. Frizzell B. Radiation therapy in oncology patients who have a pacemaker or implantable cardioverter-defibrillator. *Community Oncology* 2009; 6:469-471.

102. Venselaar JL. The effects of ionizing radiation on eight cardiac pacemakers and the influence of electromagnetic interference from two linear accelerators. *Radiother Oncol* 1985; 3:81-87.

103. Tondato F, Ng DW, Srivathsan K, Altemose GT, Halyard MY, Scott LR. Radiotherapy-induced pacemaker and implantable cardioverter defibrillator malfunction. *Expert Rev Med Devices* 2009; 6:243-249.

104. Levine E, Ferrick et al do us all a service by calling attention to the issue of "soft resets" in cardiac rhythm management (CRM) devices. *Heart Rhythm* 2008; 5:e1; author reply e1-2.

105. Franchi F, Malavasi V, Parmiggiani M, Bertoni F, Bocchedi M, Nikolskaya N, et al. In: *Influence of radiotherapy on cardiac implantable electronic devices: A single center experience*. European heart journal; ; 2012. p. 541.

106. Kry SF, Johnson JL, White RA, Howell RM, Kudchaker RJ, Gillin MT. Neutron-induced electronic failures around a high-energy linear accelerator. *Med Phys* 2011; 38:34-39.

107. Nath R, Meigooni AS, King CR, Smolen S, d'Errico F. Superheated drop detector for determination of neutron dose equivalent to patients undergoing high-energy x-ray and electron radiotherapy. *Med Phys* 1993; 20:781-787.

108. Lin J, Chu T, Lin S, Liu M. The measurement of photoneutrons in the vicinity of a Siemens Primus linear accelerator. *Applied Radiation and Isotopes* 2001; 55:315-321.

109. Calfee RV. Therapeutic radiation and pacemakers. *Pacing Clin Electrophysiol* 1982; 5:160-161.

110. Bowers RW, Scott PA, Roberts PR. Use of external defibrillator jacket to facilitate safe delivery of radiotherapy for lung cancer - a report of two cases. *Indian Heart J* 2014; 66:111-114.

111. Borleffs CJ, Thijssen J, de Bie MK, van Rees JB, van Welsema GH, van Erven L, Bax JJ, et al. Recurrent implantable cardioverter-defibrillator replacement is associated with an increasing risk of pocket-related complications. *Pacing Clin Electrophysiol* 2010; 33:1013-1019.

112. de Bie MK, van Rees JB, Thijssen J, Borleffs CJ, Trines SA, Cannegieter SC, Schalij MJ, et al. Cardiac device infections are associated with a significant mortality risk. *Heart Rhythm* 2012; 9:494-498.

113. Langer M, Orlandi E, Carrara M, Previtali P, Haeusler EA. Management of patients with implantable cardioverter defibrillator needing radiation therapy for cancer. *Br J Anaesth* 2012; 108:881-882.

114. Jacob S, Panaich SS, Maheshwari R, Haddad JW, Padanilam BJ, John SK. Clinical applications of magnets on cardiac rhythm management devices. *Europace* 2011; 13:1222-1230.

115. Crossley GH, Poole JE, Rozner MA, Asirvatham SJ, Cheng A, Chung MK, Ferguson TB, Jr., et al. The heart rhythm society (HRS)/American society of anesthesiologists (ASA) expert consensus statement on the perioperative management of patients with implantable defibrillators, pacemakers and arrhythmia monitors: Facilities and patient management: Executive summary this document was developed as a joint project with the American society of anesthesiologists (ASA), and in collaboration with the American heart association (AHA), and the society of thoracic surgeons (STS). *Heart Rhythm* 2011; 8:E1-E18.

116. Muller-Runkel R, Orsolini G, Kalokhe UP. Monitoring the radiation dose to a multiprogrammable pacemaker during radical radiation therapy: A case report. *Pacing Clin Electrophysiol* 1990; 13:1466-1470.

117. Hayes D, Vlietstra R. Pacemaker malfunction. *Ann Intern Med* 1993; 119:828-835.

118. Munshi A, Agarwal JP, Pandey KC. Cancer patients with cardiac pacemakers needing radiation treatment: A systematic review. *J Cancer Res Ther* 2013; 9:193-198.

119. Santoro F, Tarantino N, Pellegrino PL, Caivano M, Lopizzo A, Di Biase M, Brunetti ND. Cardiovascular sequelae of radiation therapy. *Clin Res Cardiol* 2014; 103:955-967.

120. Teskey RJ, Whelan I, Akyurekli Y, Eapen L, Green MS. Therapeutic irradiation over a permanent cardiac pacemaker. *Pacing Clin Electrophysiol* 1991; 14:143-145.

121. Munshi A, Wadasadawala T, Sharma PK, Sharma D, Budrukka A, Jalali R, Dinshaw KA. Radiation therapy planning of a breast cancer patient with in situ pacemaker--challenges and lessons. *Acta Oncol* 2008; 47:255-260.

122. Medtronic. Therapeutic radiation revision 1.0, 21-JAN-2013. Mounds View, MN: CRDM Technical Services; 2013.
123. Boston Scientific. Therapeutic radiation and implantable device systems revision 002-1675, rev. B. US. Marlborough, MA: Technical Services; 2012.
124. St Jude Medical. Effects of therapeutic radiation on St. Jude Medical implantable cardiac rhythm devices revision 10/13. Sylmar, CA: Technical Services; 2013.
125. Biotronik. Radiation therapy and BIOTRONIK CRM devices - pacemakers (IPG), defibrillators (ICD) and CRT-devices. Berlin, Germany: Global Technical Service CRM; 2011.



## The effect of radiotherapy beam energy on modern cardiac devices: an *in vitro* study

**Tomas Zaremba<sup>1</sup>\*, Annette Ross Jakobsen<sup>2</sup>, Anna Margrethe Thøgersen<sup>1</sup>,  
Lars Oddershede<sup>3</sup>, and Sam Riahi<sup>1</sup>**

<sup>1</sup>Department of Cardiology, Center for Cardiovascular Research, Aalborg University Hospital, Hobrovej 18-22, 9100 Aalborg, Denmark; <sup>2</sup>Department of Medical Physics, Oncology Department, Aalborg University Hospital, Aalborg, Denmark; and <sup>3</sup>Department for University Hospital Affairs, Aalborg University Hospital, Aalborg, Denmark

Received 1 June 2013; accepted after revision 15 July 2013

### Aims

Radiotherapy (RT) for malignancies can harm pacemakers (PMs) and implantable cardioverter-defibrillators (ICDs). There is some evidence that, besides cumulative dose, the damaging radiation effects increase with beam energy. The aim of this study was to determine whether modern PMs and ICDs are more sensitive to high-energy than to low-energy photon beams.

### Methods and results

Two groups of unused PMs and explanted ICDs (five PMs and one ICD in each) were subjected to irradiations in a phantom with 6 and 18 megavolt (MV) photons, respectively. The devices were exposed to radiation at doses of 2 gray (Gy) daily to simulate two clinical scenarios with the PM/ICD in the RT field. A cumulative dose of 150 Gy was given to each device, corresponding to approximately twice the therapeutic dose. In the 6 MV group, one episode of PM malfunction was detected after reaching 150 Gy. In the 18 MV group, a total of 14 episodes of malfunction were detected starting at 30 Gy in all five PMs. No episodes appeared in the ICD, at the respective treatment groups. This corresponded to a hazard ratio of 9.11 [~95% confidence interval (CI): 1.04–79.69] by Cox regression analysis between the two groups. In a repeated measures logistic regression model comparing the incidence rate of malfunctions, the odds ratio was 18.29 (~95% CI: 1.52–219.41).

### Conclusion

Photon beam energy plays a considerable role in inducing implantable cardiac device malfunctions. Low-energy RT may be safer in PM/ICD patients despite relatively high radiation dose to the device.

### Keywords

Pacemaker • ICD • Radiotherapy • Ionizing radiation • Device malfunction

\* Corresponding author. Tel: +45 99324739; fax: +45 99322361, Email: tz@rn.dk

Published on behalf of the European Society of Cardiology. All rights reserved. © The Author 2013. For permissions please email: journals.permissions@oup.com.

## Effects of High-Dose Radiotherapy on Implantable Cardioverter Defibrillators: An *In Vivo* Porcine Study

TOMAS ZAREMBA, M.D.,\* ANNETTE R. JAKOBSEN, M.Sc.,† ANNA M. THØGERSEN, M.D.,\* SAM RIAHI, M.D.,\* and BENEDICT KJÆRGAARD, M.D.‡

From the \*Department of Cardiology, Center for Cardiovascular Research; †Department of Medical Physics, Oncology Department; and ‡Department of Heart and Lung Surgery, Center for Cardiovascular Research, Aalborg University Hospital, Aalborg, Denmark

**Introduction:** Although the effects of radiotherapy (RT) on implantable cardioverter defibrillators (ICDs) have been studied *in vitro*, and some information has been gathered from case reports and series, no *in vivo* experiments have previously been performed.

**Methods:** *In vivo* effects of photon RT applied directly to five modern ICD generators from different manufacturers implanted in pigs were studied. The devices were interrogated between and after increasing doses of ionizing radiation. Afterwards, the shock function was tested.

**Results:** All ICDs withstood fractionated irradiations with a cumulative dose of 18.5 gray (Gy) of 6 megavolt (MV) photons and 18.5 Gy of 18 MV photons and were still fully functional. Especially, no oversense was recorded. Induced ventricular fibrillation was detected and treated properly by shock therapy in all cases. However, one of the ICDs converted to back-up mode later the same night.

**Conclusion:** The animal model is feasible for investigating RT effects on implanted cardiac devices. During irradiations with 37 Gy, one recoverable malfunction was present in the tested devices. Additional animal studies could provide supplementary evidence for treating ICD patients, including recommendations for reprogramming of the ICD during RT and avoidance of relocating the device. (PACE 2013; 36:1558–1563)

**radiotherapy, implantable cardioverter defibrillator, animal model, in vivo experimentation**

---

The study has been supported by the Heinrich Kopp Foundation.

Address for reprints: Tomas Zaremba, M.D., Department of Cardiology and Center for Cardiovascular Research, Aalborg University Hospital, Hobrovej 18-22, 9100 Aalborg, Denmark. Fax: 45 99322361; e-mail: tz@rn.dk

Received March 22, 2013; revised June 24, 2013; accepted July 4, 2013.

doi: 10.1111/pace.12249

©2013, The Authors. Journal compilation ©2013 Wiley Periodicals, Inc.

1558

December 2013

PACE, Vol. 36

## Risk of Device Malfunction in Cancer Patients with Implantable Cardiac Device Undergoing Radiotherapy: A Population-Based Cohort Study

TOMAS ZAREMBA, M.D.,\* ANNETTE ROSS JAKOBSEN, M.Sc.,† METTE SØGAARD, D.V.M., Ph.D.,‡ ANNA MARGRETHE THØGERSEN, M.D., Ph.D.,\* MARTIN BERG JOHANSEN, M.Sc.,§ LÆRKE BRUUN MADSEN, M.S.,\* and SAM RIAHI, M.D., Ph.D.,\*,§

From the \*Department of Cardiology, Center for Cardiovascular Research, Aalborg University Hospital, Aalborg, Denmark; †Department of Medical Physics, Oncology Department, Aalborg University Hospital, Aalborg, Denmark;

‡Department of Clinical Epidemiology, Institute of Clinical Medicine, Aarhus University Hospital, Aarhus, Denmark; and §Department of Clinical Medicine, Aalborg University and Aalborg University Hospital, Aalborg, Denmark

**Background:** Pacemakers (PMs) and implantable cardioverter defibrillators (ICDs) may develop malfunction during external beam radiotherapy (RT). We aimed to describe clinical practice in PM/ICD patients undergoing RT and to assess the rate and predictors of device malfunctions.

**Methods:** We reviewed medical records of all PM/ICD patients undergoing RT at four centers in Western Denmark during 2003–2012. Logistic regression was applied to identify predictors of PM/ICD malfunctions.

**Results:** Five hundred sixty patients were included. The annual rate of RT courses in PM/ICD patients increased by 199% from 1.45 treatments per 100,000 person-years in 2003 to 4.33 in 2012. Safety measures included supplementary evaluations of PM/ICD (38.3%), reprogramming (1.5%), relocation of the device (3.5%), and application of a magnet to the ICD during RT (10.8%). At device evaluations after the RT ( $n = 453$ ), malfunctions were detected in 10 (2.5%) PMs and four (6.8%) ICDs. Electrical resets constituted 11 (78.6%) of the malfunctions, and no failures were life-threatening or warranted PM/ICD removal. Factors associated with device malfunctions in logistic regression analysis were beam energy  $\geq 15$  MV (odds ratio [OR] 5.73; 95% confidence interval [CI], 1.58–20.76) and location of tumor below the diaphragm (OR 4.31; 95% CI, 1.42–13.12). However, the effect of tumor location declined (OR 2.27; 95% CI, 0.65–7.95) after adjustment for beam energy.

**Conclusions:** Although the rate of RT in PM/ICD patients is increasing, the damaging effects of RT on the devices seem to be usually transient. Our data suggest that high beam energy plays the pivotal role in inducing impairments in these devices. (PACE 2015; 38:343–356)

**pacemaker, implantable cardioverter defibrillator, radiotherapy, ionizing radiation, device malfunctions, beam energy**

---

Sources of financial support: Dr. Tomas Zaremba has received funding for this study from the Danish Heart Foundation (grant number 14-R97-A5215-22870).  
Conflicts of Interest: None.

Address for reprints: Tomas Zaremba, M.D., Department of Cardiology, Center for Cardiovascular Research, Aalborg University Hospital, Hobrovej 18-22, 9000 Aalborg, Denmark.  
Fax: 45-9766-4480; e-mail: tz@rn.dk.

Received October 24, 2014; revised November 30, 2014;  
accepted December 14, 2014.

doi: 10.1111/pace.12572


©2015 Wiley Periodicals, Inc.

PACE, Vol. 38

March 2015

343





ISSN (online): 2246-1302  
ISBN (online): 978-87-7112-270-1

AALBORG UNIVERSITY PRESS